

LB130448


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

44422



A= 0.0000e+000

B= 2.2352e-004

C= 3.2114e-002

Rho= 0.9997760

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.074	0.074	188	0.000	188	0				0/0)
0.2	0.200	0.210	0.010	796	0.0 %	796					5
2.5	2.500	2.350	-0.150	10370	0.0 %	10370					-6
5.0	5.000	5.024	0.024	22334	0.0 %	22334					0
7.5	7.500	7.580	0.080	33769	0.0 %	33769					1
10.0	10.000	9.961	-0.039	44422	0.0 %	44422					0

LB130448 INNNNNNNNSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Apr 2024 12:30:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	188	-	0.0000	7470A	PPB	23 Apr 2024 12:30:38	S	Std
0.2 - 1	SO2	796	-	0.2000	7470A	PPB	23 Apr 2024 12:32:55	S	Std
2.5 - 1	SO2	10370	-	2.5000	7470A	PPB	23 Apr 2024 12:35:13	S	Std
5.0 - 1	SO2	22334	-	5.0000	7470A	PPB	23 Apr 2024 12:37:30	S	Std
7.5 - 1	SO2	33769	-	7.5000	7470A	PPB	23 Apr 2024 12:39:47	S	Std
10.0 - 1	SO2	44422	-	10.0000	7470A	PPB	23 Apr 2024 12:42:04	S	Std
ICV08 - 1	ICV08	17162	3.8682	-	7470A	PPB	23 Apr 2024 12:44:58	U	SMPL
ICB08 - 1	ICB08	286	0.0960	-	7470A	PPB	23 Apr 2024 12:47:14	U	SMPL
CCV56 - 1	CCV56	21611	4.8626	-	7470A	PPB	23 Apr 2024 12:49:30	U	SMPL
CCB56 - 1	CCB56	91	0.0525	-	7470A	PPB	23 Apr 2024 12:51:45	U	SMPL
CRA - 1	CRA	730	0.1953	-	7470A	PPB	23 Apr 2024 12:54:03	U	SMPL
HighStd - 1	HighStd	44061	9.8807	-	7470A	PPB	23 Apr 2024 12:56:18	U	SMPL
ChkStd - 1	ChkStd	31227	7.0120	-	7470A	PPB	23 Apr 2024 12:58:33	U	SMPL
PB160444BL - 1	PBW	-194	-0.0112	-	7470A	PPB	23 Apr 2024 13:00:51	U	SMPL
PB160444BS - 1	LCSW	16133	3.6382	-	7470A	PPB	23 Apr 2024 13:03:10	U	SMPL
P2189-04 - 1	SILICA	25	0.0377	-	7470A	PPB	23 Apr 2024 13:05:26	U	SMPL
P2208-01 - 1	LNG-11	976	0.2503	-	7470A	PPB	23 Apr 2024 13:15:04	U	SMPL
P2208-01DUP - 1	LNG-11DUP	958	0.2462	-	7470A	PPB	23 Apr 2024 13:20:40	U	SMPL
P2208-01MS - 1	LNG-11MS	20985	4.7227	-	7470A	PPB	23 Apr 2024 13:22:57	U	SMPL
P2208-01MSD - 1	LNG-11MSD	19974	4.4967	-	7470A	PPB	23 Apr 2024 13:25:12	U	SMPL
CCV57 - 1	CCV57	23268	5.2330	-	7470A	PPB	23 Apr 2024 13:27:29	U	SMPL
CCB57 - 1	CCB57	40	0.0411	-	7470A	PPB	23 Apr 2024 13:32:21	U	SMPL
PB160402TB - 1	PB160402TB	51	0.0435	-	7470A	PPB	23 Apr 2024 13:37:11	U	SMPL
P2208-01LX5 - 1		173	0.0708	-	7470A	PPB	23 Apr 2024 13:39:26	U	SMPL
P2208-01A - 1		20419	4.5962	-	7470A	PPB	23 Apr 2024 13:41:42	U	SMPL
CCV58 - 1	CCV58	23367	5.2551	-	7470A	PPB	23 Apr 2024 13:43:58	U	SMPL
CCB58 - 1	CCB58	10	0.0343	-	7470A	PPB	23 Apr 2024 13:48:48	U	SMPL