


LB130453

7471B

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 2.0171e-004

C= 2.0074e-002

Rho= 0.9997733

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.035	0.035	74	0.000	74					2
0.2	0.200	0.205	0.005	919	0.0 %	919					-1
2.5	2.500	2.477	-0.023	12178	0.0 %	12178					-1
5.0	5.000	5.039	0.039	24883	0.0 %	24883					-2
7.5	7.500	7.345	-0.155	36314	0.0 %	36314					1
10.0	10.000	10.099	0.099	49964	0.0 %	49964					

LB130453 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Apr 2024 14:42:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.0000	7471B	PPB	23 Apr 2024 14:46:25	S	Std
0.2 - 1	50.2	919	-	0.2000	7471B	PPB	23 Apr 2024 14:48:42	S	Std
2.5 - 1	52.5	12178	-	2.5000	7471B	PPB	23 Apr 2024 14:51:00	S	Std
5.0 - 1	55	24883	-	5.0000	7471B	PPB	23 Apr 2024 14:53:17	S	Std
7.5 - 1	57.5	36314	-	7.5000	7471B	PPB	23 Apr 2024 14:55:36	S	Std
10.0 - 1	511	49964	-	10.0000	7471B	PPB	23 Apr 2024 14:57:57	S	Std
ICV09 - 1	ICV09	19346	3.9224	-	7471B	PPB	23 Apr 2024 15:01:18	U	SMPL
ICB09 - 1	ICB09	88	0.0378	-	7471B	PPB	23 Apr 2024 15:03:33	U	SMPL
CCV59 - 1	CCV59	25290	5.1214	-	7471B	PPB	23 Apr 2024 15:05:54	U	SMPL
CCB59 - 1	CCB59	310	0.0826	-	7471B	PPB	23 Apr 2024 15:09:30	U	SMPL
CRA - 1	CRA	941	0.2099	-	7471B	PPB	23 Apr 2024 15:11:45	U	SMPL
HighStd - 1	HighStd	51304	10.3688	-	7471B	PPB	23 Apr 2024 15:14:00	U	SMPL
ChkStd - 1	ChkStd	34315	6.9419	-	7471B	PPB	23 Apr 2024 15:16:15	U	SMPL
PB160455BL - 1	PBS	-249	-0.0302	-	7471B	PPB	23 Apr 2024 15:18:43	U	SMPL
PB160455BS - 1	LCSS	18605	3.7730	-	7471B	PPB	23 Apr 2024 15:21:07	U	SMPL
P2199-01 - 1	STOCK-PILE	5490	1.1275	-	7471B	PPB	23 Apr 2024 15:30:03	U	SMPL
P2199-01DUP - 1	STOCK-PILEDUP	5300	1.0892	-	7471B	PPB	23 Apr 2024 15:32:21	U	SMPL
P2199-01MS - 1	STOCK-PILEMS	29603	5.9914	-	7471B	PPB	23 Apr 2024 15:34:40	U	SMPL
P2199-01MSD - 1	STOCK-PILEMSD	26084	5.2816	-	7471B	PPB	23 Apr 2024 15:36:59	U	SMPL
P2225-01 - 1	TR-06-042224	1656	0.3541	-	7471B	PPB	23 Apr 2024 15:39:21	U	SMPL
CCV60 - 1	CCV60	26661	5.3980	-	7471B	PPB	23 Apr 2024 15:41:41	U	SMPL
CCB60 - 1	CCB60	-198	-0.0199	-	7471B	PPB	23 Apr 2024 15:43:57	U	SMPL
P2227-01 - 1	HD-02-04222024	2398	0.5038	-	7471B	PPB	23 Apr 2024 15:46:17	U	SMPL
P2228-01 - 1	EXCAVATION-SOIL	3034	0.6321	-	7471B	PPB	23 Apr 2024 15:48:33	U	SMPL
P2199-01LX5 - 1		1188	0.2597	-	7471B	PPB	23 Apr 2024 15:50:49	U	SMPL
P2199-01A - 1		26682	5.4022	-	7471B	PPB	23 Apr 2024 15:56:15	U	SMPL
CCV61 - 1	CCV61	24184	4.8983	-	7471B	PPB	23 Apr 2024 16:01:55	U	SMPL
CCB61 - 1	CCB61	-164	-0.0130	-	7471B	PPB	23 Apr 2024 16:04:11	U	SMPL