

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130458

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/24/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	04/24/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	04/24/2024	08:15
4	ICV	1	Water	NA	NA	20.3	7.02	04/24/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	04/24/2024	08:25
6	P2245-04	1	Water	NA	NA	20.1	8.88	04/24/2024	08:35
7	P2246-03	1	Water	NA	NA	20.5	11.51	04/24/2024	08:45
8	P2246-03DUP	1	Water	NA	NA	20.6	11.52	04/24/2024	08:46
9	CCV2	1	Water	NA	NA	20.3	12.02	04/24/2024	08:50

WORKLIST(Hardcopy Internal Chain)

VB 130458

WorkList Name : ph p2245 WorkList ID : 179680 Department : Wet-Chemistry Date : 04-24-2024 07:57:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2245-04 E	PIT-01	Water	pH	Cool 4 deg C	PSEG03	I41	04/23/2024	9040C
P2246-03 C	SAW-CUTTING-LIQUID	Water	pH	Cool 4 deg C	PSEG03	I41	04/23/2024	9040C

Date/Time 04/24/24 08:00
Raw Sample Received by: 28160C1
Raw Sample Relinquished by: CD SM

Date/Time 04/24/24 12:00
Raw Sample Received by: CD SM
Raw Sample Relinquished by: 78160C1