

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130460

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	04/24/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/24/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	04/24/2024	08:20
4	ICV	1	Water	NA	NA	20.3	7.00	04/24/2024	08:22
5	CCV1	1	Water	NA	NA	20.3	2.01	04/24/2024	08:25
6	P2240-02	1	Solid	20.02	20	22.0	8.11	04/24/2024	08:45
7	P2240-02DUP	1	Solid	20.03	20	22.1	8.12	04/24/2024	08:46
8	P2246-01	1	Solid	20.02	20	21.8	11.12	04/24/2024	09:00
9	CCV2	1	Water	NA	NA	20.3	12.02	04/24/2024	09:05