

LB130461


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42149



A= 0.0000e+000

B= 2.3683e-004 slope

C= -1.2133e-003 y-intercept

Rho= 0.9997066

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/b)
0.00	0.000	0.058	0.058	248	0.000	248					
0.05	0.050					442					
0.20	0.200	0.217	0.017	920	0.0 %	920	0				
2.50	2.500	2.468	-0.032	10428	0.0 %	10428					
5.00	5.000	4.842	-0.158	20450	0.0 %	20450					
7.50	7.500	7.634	0.134	32240	0.0 %	32240					
10.0	10.000	9.981	-0.019	42149	0.0 %	42149					

g/b)

248

442

920

10428

20450

32240

42149

LB130461

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	248	-		0 SFAM01.1	PPB	4/24/2024 10:26	Std
	0.2 S01	920	-		0.2 SFAM01.1	PPB	4/24/2024 10:29	Std
	2.5 S02	10428	-		2.5 SFAM01.1	PPB	4/24/2024 10:31	Std
	5 S03	20450	-		5 SFAM01.1	PPB	4/24/2024 10:33	Std
	7.5 S04	32240	-		7.5 SFAM01.1	PPB	4/24/2024 10:35	Std
	10 S05	42149	-		10 SFAM01.1	PPB	4/24/2024 10:40	Std
ICV010	ICV010	18062	4.2765 -		SFAM01.1	PPB	4/24/2024 10:46	SMPL
ICB010	ICB010	9	0.0009 -		SFAM01.1	PPB	4/24/2024 10:49	SMPL
CCV062	CCV062	22273	5.2738 -		SFAM01.1	PPB	4/24/2024 10:51	SMPL
CCB062	CCB062	9	0.0009 -		SFAM01.1	PPB	4/24/2024 10:53	SMPL
PB160480BL	PBS480	224	0.0518 -		SFAM01.1	PPB	4/24/2024 10:55	SMPL
P2236-01	E0807	10641	2.5189 -		SFAM01.1	PPB	4/24/2024 10:58	SMPL
P2236-02	E0808	3343	0.7905 -		SFAM01.1	PPB	4/24/2024 11:00	SMPL
P2236-03	E0809	1274	0.3005 -		SFAM01.1	PPB	4/24/2024 11:07	SMPL
P2236-04	E0809D	1104	0.2602 -		SFAM01.1	PPB	4/24/2024 11:09	SMPL
P2236-05	E0809S	13870	3.2837 -		SFAM01.1	PPB	4/24/2024 11:12	SMPL
CCV063	CCV063	22777	5.3931 -		SFAM01.1	PPB	4/24/2024 11:14	SMPL
CCB063	CCB063	62	0.0135 -		SFAM01.1	PPB	4/24/2024 11:16	SMPL