

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130462

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/24/2024	12:00
2	P2183-06		1	12	100.00	>212.0	>212.0	04/24/2024	12:30
3	P2183-06DUP		1	13	100.00	>212.0	>212.0	04/24/2024	13:00
4	P2226-01		1	14	100.00	>212.0	>212.0	04/24/2024	13:30
5	P2245-04		1	12	100.00	>212.0	>212.0	04/24/2024	14:00
6	P2246-03		1	13	100.00	>212.0	>212.0	04/24/2024	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130462

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/24/24 flashpoint WorkList ID : 179689 Department : Wet-Chemistry Date : 04-24-2024 08:03:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2183-06	ANTIFREEZE-DRUM-1103	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/18/2024	1010B
P2226-01	GROUND-WATER	Water	Flash Point	Cool 4 deg C	TULL02	I41	04/22/2024	1010B
P2245-04	PIT-01	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/23/2024	1010B
P2246-03	SAW-CUTTING-LIQUID	Water	Flash Point	Cool 4 deg C	PSEG03	I41	04/23/2024	1010B

Date/Time 04/24/24 10:15
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: JH WOC

Date/Time 04/24/24 15:10
Raw Sample Received by: JH WOC
Raw Sample Relinquished by: N.P.W.C