

LB130465


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

42497



A= 0.0000e+000

B= 2.3394e-004

C= -6.4119e-002

Rho= 0.9996742

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.007	0.007	302	0.000	302	0				
0.2	0.200	0.211	0.011	1174	0.0 %	1174					
2.5	2.500	2.422	-0.078	10629	0.0 %	10629					
5.0	5.000	5.004	0.004	21663	0.0 %	21663					
7.5	7.500	7.679	0.179	33099	0.0 %	33099					
10.0	10.000	9.878	-0.122	42497	0.0 %	42497					

120697
120697

LB130465 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Apr 2024 12:14:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	302	-	0.000	7470A	PPB	24 Apr 2024 12:24:23	S	Std
0.2 - 1	52.2	1174	-	0.200	7470A	PPB	24 Apr 2024 12:26:40	S	Std
2.5 - 1	52.5	10629	-	2.500	7470A	PPB	24 Apr 2024 12:28:57	S	Std
5.0 - 1	55	21663	-	5.000	7470A	PPB	24 Apr 2024 12:31:14	S	Std
7.5 - 1	57.5	33099	-	7.500	7470A	PPB	24 Apr 2024 12:33:32	S	Std
10.0 - 1	60	42497	-	10.000	7470A	PPB	24 Apr 2024 12:35:50	S	Std
ICV11 - 1	ICV11	17509	4.0320	-	7470A	PPB	24 Apr 2024 12:41:20	U	SMPL
ICB11 - 1	ICB11	-99	-0.0873	-	7470A	PPB	24 Apr 2024 12:43:35	U	SMPL
CCV64 - 1	CCV64	22616	5.2267	-	7470A	PPB	24 Apr 2024 12:45:53	U	SMPL
CCB64 - 1	CCB64	192	-0.0192	-	7470A	PPB	24 Apr 2024 12:50:51	U	SMPL
CRA - 1	CRA	1165	0.2084	-	7470A	PPB	24 Apr 2024 12:53:06	U	SMPL
HighStd - 1	HighStd	43646	10.1465	-	7470A	PPB	24 Apr 2024 12:55:21	U	SMPL
ChkStd - 1	ChkStd	31574	7.3223	-	7470A	PPB	24 Apr 2024 12:57:36	U	SMPL
PB160489BL - 1	PBW	152	-0.0286	-	7470A	PPB	24 Apr 2024 13:02:36	U	SMPL
PB160489BS - 1	LCSW	15811	3.6347	-	7470A	PPB	24 Apr 2024 13:04:51	U	SMPL
P2218-01 - 1	B-131-GW-01	1625	0.3160	-	7470A	PPB	24 Apr 2024 13:16:30	U	SMPL
P2218-01DUP - 1	B-131-GW-01DUP	1473	0.2805	-	7470A	PPB	24 Apr 2024 13:18:46	U	SMPL
P2218-01MS - 1	B-131-GW-01MS	20203	4.6622	-	7470A	PPB	24 Apr 2024 13:21:04	U	SMPL
P2218-01MSD - 1	B-131-GW-01MSD	19279	4.4460	-	7470A	PPB	24 Apr 2024 13:23:21	U	SMPL
P2218-02 - 1	B-131-GW-02	1333	0.2477	-	7470A	PPB	24 Apr 2024 13:25:38	U	SMPL
CCV65 - 1	CCV65	21396	4.9413	-	7470A	PPB	24 Apr 2024 13:39:08	U	SMPL
CCB65 - 1	CCB65	123	-0.0353	-	7470A	PPB	24 Apr 2024 13:41:23	U	SMPL
P2224-01 - 1	CLEANOUT-TOTE	5993	1.3379	-	7470A	PPB	24 Apr 2024 13:43:40	U	SMPL
P2246-03 - 1	SAW-CUTTING-LIQUID	733	0.1074	-	7470A	PPB	24 Apr 2024 13:45:56	U	SMPL
P2218-01LX5 - 1		340	0.0154	-	7470A	PPB	24 Apr 2024 13:48:13	U	SMPL
P2218-01A - 1		20129	4.6449	-	7470A	PPB	24 Apr 2024 13:50:29	U	SMPL
CCV66 - 1	CCV66	20595	4.7539	-	7470A	PPB	24 Apr 2024 14:05:58	U	SMPL
CCB66 - 1	CCB66	210	-0.0150	-	7470A	PPB	24 Apr 2024 14:08:14	U	SMPL