

LB130471


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44571



A= 0.0000e+000

B= 2.2637e-004 stop

C= -4.7955e-002 7=1201 2v<415

Rho= 0.9999625

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9, D
0.00	0.000	0.012	0.012	265	0.000	265					
0.05	0.050					442	0				
0.20	0.200	0.234	0.034	1245	0.0 %	1245					
2.50	2.500	2.460	-0.040	11078	0.0 %	11078					
5.00	5.000	4.990	-0.010	22253	0.0 %	22253					
7.50	7.500	7.463	-0.037	33180	0.0 %	33180					
10.0	10.000	10.042	0.042	44571	0.0 %	44571					

9, D
1
17
-2
0
0
0

LB130471

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	265	-		0 SFAM01.1	PPB	4/24/2024 15:43	Std
	0.2 S01	1245	-		0.2 SFAM01.1	PPB	4/24/2024 15:45	Std
	2.5 S02	11078	-		2.5 SFAM01.1	PPB	4/24/2024 15:50	Std
	5 S03	22253	-		5 SFAM01.1	PPB	4/24/2024 15:52	Std
	7.5 S04	33180	-		7.5 SFAM01.1	PPB	4/24/2024 15:55	Std
	10 S05	44571	-		10 SFAM01.1	PPB	4/24/2024 16:00	Std
ICV012	ICV012	17417	3.8948 -		SFAM01.1	PPB	4/24/2024 16:05	SMPL
ICB012	ICB012	-179	-0.0885 -		SFAM01.1	PPB	4/24/2024 16:07	SMPL
CCV067	CCV067	22514	5.0486 -		SFAM01.1	PPB	4/24/2024 16:10	SMPL
CCB067	CCB067	-218	-0.0973 -		SFAM01.1	PPB	4/24/2024 16:12	SMPL
PB160469BL	PBW469	118	-0.0212 -		SFAM01.1	PPB	4/24/2024 16:14	SMPL
P2021-01	E07W7	219	0.0016 -		SFAM01.1	PPB	4/24/2024 16:16	SMPL
P2021-02	E07W7D	242	0.0068 -		SFAM01.1	PPB	4/24/2024 16:19	SMPL
P2021-03	E07W7S	4133	0.8876 -		SFAM01.1	PPB	4/24/2024 16:21	SMPL
P2064-01	E1CY8	124	-0.0199 -		SFAM01.1	PPB	4/24/2024 16:23	SMPL
P2064-02	E1D00	180	-0.0072 -		SFAM01.1	PPB	4/24/2024 16:25	SMPL
P2064-03	E1CY9	197	-0.0034 -		SFAM01.1	PPB	4/24/2024 16:28	SMPL
P2064-04	E1D01	231	0.0043 -		SFAM01.1	PPB	4/24/2024 16:30	SMPL
CCV068	CCV068	22446	5.0332 -		SFAM01.1	PPB	4/24/2024 16:32	SMPL
CCB068	CCB068	-125	-0.0763 -		SFAM01.1	PPB	4/24/2024 16:35	SMPL