

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130481

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/25/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.01	04/25/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	04/25/2024	08:13
4	ICV	1	Water	NA	NA	20.3	6.98	04/25/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	04/25/2024	08:20
6	P2260-02	1	Solid	20.02	20	22.9	6.58	04/25/2024	08:25
7	P2260-02DUP	1	Solid	20.03	20	22.8	6.59	04/25/2024	08:26
8	P2260-03	1	Solid	20.02	20	23.4	6.60	04/25/2024	08:30
9	CCV2	1	Water	NA	NA	20.2	12.02	04/25/2024	08:32

WORKLIST(Hardcopy Internal Chain)

130481

WorkList Name : ph p2261 solids WorkList ID : 179714 Department : Wet-Chemistry Date : 04-25-2024 07:54:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2260-02	60238	Solid	pH	Cool 4 deg C	PSEG03	I31	04/24/2024	9045D
P2260-03	60239	Solid	pH	Cool 4 deg C	PSEG03	I31	04/24/2024	9045D

Date/Time 04/25/24 08:00
 Raw Sample Received by: JD (cc)
 Raw Sample Relinquished by: CP sm

Date/Time 04/25/24 12:00
 Raw Sample Received by: CP sm
 Raw Sample Relinquished by: JD (cc)