

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130494

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/25/2024	10:30
2	P2257-07		1	12	100.00	>212.0	>212.0	04/25/2024	11:00
3	P2257-07DUP		1	13	100.00	>212.0	>212.0	04/25/2024	11:30
4	P2257-12		1	12	100.00	>212.0	>212.0	04/25/2024	12:00
5	P2257-13		1	14	100.00	>212.0	>212.0	04/25/2024	12:30
6	P2260-04		1	13	100.00	>212.0	>212.0	04/25/2024	13:00
7	P2261-04		1	12	100.00	>212.0	>212.0	04/25/2024	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

L13-130494

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/25/24 flashpoint

WorkList ID : 179755

Department : Wet-Chemistry

Date : 04-25-2024 10:04:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2257-07	41-45	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B
P2257-12	413-416	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B
P2257-13	4524	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B
P2260-04	60240	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B
P2261-04	A069	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B

Date/Time 04/25/2024 10:20

Raw Sample Received by: N.P.W.C.

Raw Sample Relinquished by: [Signature]

Date/Time 04/25/2024 14:15

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: N.P.W.C.