


LB130498

SFAM01.1

INSTRUMENT ID: C111



Linear

A= 0.0000e+000

B= 2.1065e-004 slope

C= -8.3174e-002 y-intercept

Rho= 0.9999239

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b D
0.00	0.000	0.052	0.052	644	0.000	644					
0.05	0.050					442	0				
0.20	0.200	0.219	0.019	1436	0.0 %	1436					
2.50	2.500	2.418	-0.082	11874	0.0 %	11874					
5.00	5.000	4.997	-0.003	24116	0.0 %	24116					
7.50	7.500	7.471	-0.029	35862	0.0 %	35862					
10.0	10.000	10.042	0.042	48066	0.0 %	48066					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	644	-		0 SFAM01.1	PPB	4/25/2024 16:19	Std
	0.2 S01	1436	-		0.2 SFAM01.1	PPB	4/25/2024 16:21	Std
	2.5 S02	11874	-		2.5 SFAM01.1	PPB	4/25/2024 16:24	Std
	5 S03	24116	-		5 SFAM01.1	PPB	4/25/2024 16:26	Std
	7.5 S04	35862	-		7.5 SFAM01.1	PPB	4/25/2024 16:28	Std
	10 S05	48066	-		10 SFAM01.1	PPB	4/25/2024 16:30	Std
ICV015	ICV015	18904	3.899 -		SFAM01.1	PPB	4/25/2024 16:33	SMPL
ICB015	ICB015	323	-0.0151 -		SFAM01.1	PPB	4/25/2024 16:36	SMPL
CCV076	CCV076	24640	5.1073 -		SFAM01.1	PPB	4/25/2024 16:38	SMPL
CCB076	CCB076	304	-0.0191 -		SFAM01.1	PPB	4/25/2024 16:40	SMPL
PB160535BL	PBS635	587	0.0405 -		SFAM01.1	PPB	4/25/2024 16:42	SMPL
P2250-01	E0812	6253	1.234 -		SFAM01.1	PPB	4/25/2024 16:45	SMPL
P2250-02	E0813	1100	0.1485 -		SFAM01.1	PPB	4/25/2024 16:47	SMPL
P2250-03	E0814	1393	0.2103 -		SFAM01.1	PPB	4/25/2024 16:49	SMPL
P2250-04	E08M8	1387	0.209 -		SFAM01.1	PPB	4/25/2024 16:51	SMPL
P2250-05	E08M9	1246	0.1793 -		SFAM01.1	PPB	4/25/2024 16:54	SMPL
P2250-06	E08N0	1220	0.1738 -		SFAM01.1	PPB	4/25/2024 16:56	SMPL
P2250-07	E08N1	1042	0.1363 -		SFAM01.1	PPB	4/25/2024 16:58	SMPL
P2250-08	E08N2	1067	0.1416 -		SFAM01.1	PPB	4/25/2024 17:01	SMPL
P2250-09	E08N2D	884	0.103 -		SFAM01.1	PPB	4/25/2024 17:03	SMPL
P2250-10	E08N2S	14143	2.8961 -		SFAM01.1	PPB	4/25/2024 17:05	SMPL
PB160536BL	PBS636	478	0.0175 -		SFAM01.1	PPB	4/25/2024 17:07	SMPL
P2046-01	GCGZ0	2699	0.4854 -		SFAM01.1	PPB	4/25/2024 17:14	SMPL
P2046-02	GCGZ0D	2616	0.4679 -		SFAM01.1	PPB	4/25/2024 17:16	SMPL
P2046-03	GCGZ0S	14720	3.0176 -		SFAM01.1	PPB	4/25/2024 17:19	SMPL
P2046-04	GCGZ1	3276	0.6069 -		SFAM01.1	PPB	4/25/2024 17:21	SMPL
P2046-05	GCGZ2	3317	0.6156 -		SFAM01.1	PPB	4/25/2024 17:23	SMPL
P2046-06	GCGZ3	3551	0.6648 -		SFAM01.1	PPB	4/25/2024 17:26	SMPL
P2046-07	GCGZ4	1986	0.3352 -		SFAM01.1	PPB	4/25/2024 17:28	SMPL
P2046-08	GCGZ5	876	0.1014 -		SFAM01.1	PPB	4/25/2024 17:30	SMPL
CCV077	CCV077	25822	5.3563 -		SFAM01.1	PPB	4/25/2024 17:32	SMPL
CCB077	CCB077	533	0.0291 -		SFAM01.1	PPB	4/25/2024 17:35	SMPL
P2046-09	GCGZ6	868	0.0997 -		SFAM01.1	PPB	4/25/2024 17:37	SMPL
P2046-10	GCGZ7	1303	0.1913 -		SFAM01.1	PPB	4/25/2024 17:39	SMPL
P2046-11	GCGZ8	2062	0.3512 -		SFAM01.1	PPB	4/25/2024 17:41	SMPL
P2046-12	GCGZ9	2215	0.3834 -		SFAM01.1	PPB	4/25/2024 17:44	SMPL
P2046-13	GCH00	2256	0.3921 -		SFAM01.1	PPB	4/25/2024 17:46	SMPL
P2046-14	GCH01	2335	0.4087 -		SFAM01.1	PPB	4/25/2024 17:48	SMPL
P2046-15	GCH02	2227	0.3859 -		SFAM01.1	PPB	4/25/2024 17:51	SMPL
P2046-16	GCH03	3431	0.6396 -		SFAM01.1	PPB	4/25/2024 17:53	SMPL
P2046-17	GCH04	2396	0.4215 -		SFAM01.1	PPB	4/25/2024 17:55	SMPL
P2046-18	GCH05	1551	0.2435 -		SFAM01.1	PPB	4/25/2024 17:57	SMPL
P2046-19	GCH06	1460	0.2244 -		SFAM01.1	PPB	4/25/2024 18:00	SMPL
P2046-20	GCH07	3096	0.569 -		SFAM01.1	PPB	4/25/2024 18:02	SMPL
P2046-21	GCH08	2549	0.4538 -		SFAM01.1	PPB	4/25/2024 18:04	SMPL
PB160537BL	PBS637	603	0.0438 -		SFAM01.1	PPB	4/25/2024 18:06	SMPL

LB130498

INSTRUMENT ID : CV1

P2047-01	GCH09	4853	0.9391 -	SFAM01.1	PPB	4/25/2024 18:09 SMPL
P2047-02	GCH10	9435	1.9043 -	SFAM01.1	PPB	4/25/2024 18:11 SMPL
P2047-03	GCH10D	10085	2.0412 -	SFAM01.1	PPB	4/25/2024 18:13 SMPL
P2047-04	GCH10S	22778	4.715 -	SFAM01.1	PPB	4/25/2024 18:15 SMPL
P2047-05	GCH11	17980	3.7043 -	SFAM01.1	PPB	4/25/2024 18:18 SMPL
P2047-06	GCH12	1468	0.2261 -	SFAM01.1	PPB	4/25/2024 18:20 SMPL
CCV078	CCV078	25843	5.3607 -	SFAM01.1	PPB	4/25/2024 18:22 SMPL
CCB078	CCB078	557	0.0342 -	SFAM01.1	PPB	4/25/2024 18:25 SMPL
P2047-07	GCH13	11525	2.3446 -	SFAM01.1	PPB	4/25/2024 18:27 SMPL
P2047-08	GCH14	1649	0.2642 -	SFAM01.1	PPB	4/25/2024 18:29 SMPL
P2047-09	GCH15	2191	0.3784 -	SFAM01.1	PPB	4/25/2024 18:31 SMPL
P2047-10	GCH16	8430	1.6926 -	SFAM01.1	PPB	4/25/2024 18:34 SMPL
P2047-11	GCH17	2576	0.4595 -	SFAM01.1	PPB	4/25/2024 18:36 SMPL
P2047-12	GCH18	3804	0.7181 -	SFAM01.1	PPB	4/25/2024 18:38 SMPL
P2047-13	GCH19	3922	0.743 -	SFAM01.1	PPB	4/25/2024 18:41 SMPL
P2047-14	GCH20	8705	1.7505 -	SFAM01.1	PPB	4/25/2024 18:43 SMPL
P2047-15	GCH21	2812	0.5092 -	SFAM01.1	PPB	4/25/2024 18:45 SMPL
P2047-16	GCH22	345034	HIGH -	SFAM01.1	PPB	4/25/2024 18:47
CCV079	CCV079	24496	5.0769 -	SFAM01.1	PPB	4/25/2024 18:55 SMPL
CCB079	CCB079	420	0.0053 -	SFAM01.1	PPB	4/25/2024 18:57 SMPL