


Calibration for 005CALS.d

Batch Folder: D:\Agilent\ICPMH\1\DATA\P7042524-MS3.b\
 Analysis File: P7042524-MS3.batch.bin
 DA Date-Time: 2024-04-25 12:28:23
 Calibration Title:
 Calibration Method: External Calibration
 VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	004CALB.d	S00	2024-04-25 12:16:07
2	006CALS.d	S02	2024-04-25 12:22:39
3	007CALS.d	S03	2024-04-25 12:25:55
4	008CALS.d	S04	2024-04-25 12:28:57
5	009CALS.d	S05	2024-04-25 12:31:47
6	010CALS.d	S06	2024-04-25 12:34:34
7	011CALS.d	S07	2024-04-25 12:37:20
8	012CALS.d	S08	2024-04-25 12:40:08

9 Be [No Gas] ISTD:6 Li [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	91.03	0.0000	P	9.6
2	☐	1.000	1.079	3401.34	0.0013	P	1.3
3	☐	50.000	54.442	169473.64	0.0635	P	1.1
4	☐	125.000	137.527	412152.79	0.1602	P	1.5
5	☐	250.000	266.081	795780.84	0.3100	P	1.2
6	☐	500.000	508.672	1461564.36	0.5926	A	0.7
7	☐	1000.000	989.856	2872464.48	1.1531	A	1.1
8	☐			1920.03	0.0008	P	17.7


$$y = 0.0012 * x + 3.4875E-005$$

$$R = 0.9998$$


$$DL = 0.008644$$

Weight: <None>


Min Conc: 0

12 C [No Gas] No available calibration points

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

12 C [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

23 Na [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	37410.93	0.0741	P	0.4
2	☐	500.000	521.713	545753.85	1.1210	P	4.1
3	☐	5000.000	5345.871	5186767.25	10.8018	A	0.8
4	☐	12500.000	13285.935	12252791.76	26.7354	A	0.7
5	☐	25000.000	26809.494	24038965.48	53.8736	A	0.9
6	☐	50000.000	53486.433	47793262.62	107.4071	A	1.1
7	☐	100000.00	105975.00	95220544.14	212.7377	A	0.8
8	☐	500000.00	498342.75	456861459.8	1,000.115	A	4.4


$$y = 0.0020 * x + 0.0741$$

$$R = 0.9999$$

$$DL = 0.4941$$

Weight: <None>

Min Conc: 0

24 Mg [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1402.30	0.0028	P	6.5
2	☐	500.000	525.213	320122.24	0.6574	P	3.1
3	☐	5000.000	5477.875	3279757.38	6.8302	A	1.2
4	☐	12500.000	13480.170	7701253.70	16.8039	A	0.5
5	☐	25000.000	27131.582	15090486.03	33.8185	A	0.4
6	☐	50000.000	53782.527	29828914.84	67.0353	A	1.2
7	☐	100000.00	106110.80	59197091.89	132.2552	A	0.8
8	☐	500000.00	498263.69	283645660.2	621.0196	A	5.0


$$y = 0.0012 * x + 0.0028$$

$$R = 0.9999$$

$$DL = 0.4363$$

Weight: <None>

Min Conc: 0

27 Al [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	326.67	0.0006	P	6.5
2	☐	20.000	21.630	6951.64	0.0143	P	2.2
3	☐	1000.000	1122.538	339784.68	0.7077	P	0.4
4	☐	2500.000	2767.636	799177.22	1.7438	P	0.1
5	☐	5000.000	5364.653	1507927.30	3.3795	A	1.2
6	☐	10000.000	10859.551	3043831.79	6.8403	A	0.6
7	☐	20000.000	21104.427	5949627.00	13.2929	A	1.7
8	☐	100000.00	99667.010	28678114.29	62.7739	A	4.1

$$y = 6.2983E-004 * x + 6.4709E-004$$


$$R = 0.9999$$

$$DL = 0.2015$$

Weight: <None>

Min Conc: 0

28 Si [No Gas] ISTD : 45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	500934.14	0.0759	P	0.3
2	☐	10.000	6.696	573534.01	0.0875	P	0.5
3	☐	50.000	68.427	1239490.03	0.1945	A	0.5
4	☐	125.000	150.006	2063544.67	0.3360	A	0.5
5	☐	250.000	281.282	3315975.02	0.5637	A	1.4
6	☐	500.000	526.451	5797949.01	0.9889	A	0.2
7	☐	1000.000	974.940	10501936.09	1.7666	A	1.5
8	☐			857043.09	0.1461	P	2.7

$$y = 0.0017 * x + 0.0759$$


$$R = 0.9987$$

$$DL = 0.4395$$

Weight: <None>

Min Conc: 0

31 P [He] ISTD : 45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	-16.498	516.68	0.0010	P	7.5
2	☐	0.000	16.498	1030.04	0.0021	P	5.6
3	☐	1000.000	1029.152	17126.35	0.0357	P	1.1
4	☐	2500.000	2565.253	39670.22	0.0866	P	0.7
5	☐	5000.000	5042.700	75251.37	0.1686	P	0.5
6	☐	10000.000	10197.913	151040.90	0.3394	P	0.9
7	☐	20000.000	19880.754	295530.97	0.6602	P	1.1
8	☐			641.13	0.0014	P	9.8

$$y = 3.3131E-005 * x + 0.0016$$


$$R = 0.9999$$

$$DL = 8.806$$

Weight: <None>

Min Conc: 0

34 S [No Gas] ISTD:45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	-46.145	534257.87	0.0809	P	1.0
2	☐	0.000	46.145	537575.80	0.0820	P	0.3
3	☐	1000.000	1084.068	599677.74	0.0941	P	1.7
4	☐	2500.000	2497.781	679534.63	0.1107	P	1.0
5	☐	5000.000	4799.921	809189.17	0.1376	P	1.2
6	☐	10000.000	10867.486	1222378.22	0.2085	A	1.7
7	☐	20000.000	19612.351	1847071.40	0.3107	A	2.2
8	☐			473157.56	0.0807	P	1.7

$$y = 1.1691\text{E-}005 * x + 0.0815$$


$$R = 0.9985$$

$$DL = 135.9$$


Weight: <None>

Min Conc: 0


35 Cl [No Gas] No available calibration points



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

35 Cl [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

39 K [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	105809.74	0.2096	P	0.1
2	☐	500.000	529.112	421763.18	0.8664	P	4.3
3	☐	2500.000	2670.831	1692577.61	3.5251	A	0.5
4	☐	6250.000	6608.056	3855494.46	8.4126	A	0.8
5	☐	12500.000	13046.105	7319850.44	16.4045	A	1.7
6	☐	25000.000	26343.546	14644384.36	32.9115	A	1.8
7	☐	50000.000	52407.867	29213923.18	65.2667	A	0.3
8	☐	250000.00	249346.04	141476729.0	309.7383	A	4.8


$$y = 0.0012 * x + 0.2096$$

$$R = 0.9999$$

$$DL = 0.7471$$

Weight: <None>

Min Conc: 0

43 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	629.38	0.0001	P	1.3
2	☐	500.000	501.396	21715.99	0.0033	P	1.1
3	☐	5000.000	5535.660	226899.56	0.0356	P	0.4
4	☐	12500.000	13865.934	546902.36	0.0891	P	0.6
5	☐	25000.000	27661.729	1044525.18	0.1776	P	0.3
6	☐	50000.000	52836.467	1988119.64	0.3391	A	1.9
7	☐	100000.00	103565.27	3949813.09	0.6645	A	3.4
8	☐	500000.00	498830.70	18771652.88	3.2004	A	1.1


$$y = 6.4157E-006 * x + 9.5323E-005$$

$$R = 1.0000$$

$$DL = 0.5986$$

Weight: <None>

Min Conc: 0

44 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	24334.28	0.0037	P	2.3
2	☐	500.000	509.142	367306.26	0.0560	P	0.8
3	☐	5000.000	5195.307	3426141.55	0.5377	A	0.6
4	☐	12500.000	13006.251	8233462.20	1.3406	A	0.9
5	☐	25000.000	26028.093	15760357.72	2.6792	A	1.1
6	☐	50000.000	51882.117	31291242.31	5.3368	A	0.9
7	☐	100000.00	102074.29	62393266.20	10.4963	A	2.1
8	☐	500000.00	499330.90	301084815.1	51.3317	A	0.4

$$y = 1.0279E-004 * x + 0.0037$$


$$R = 1.0000$$

$$DL = 2.454$$

Weight: <None>

Min Conc: 0

47 Ti [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	134.45	0.0000	P	15.5
2	☐	5.000	4.931	9817.71	0.0015	P	1.8
3	☐	500.000	512.725	978773.13	0.1536	P	0.5
4	☐	1250.000	1272.505	2340969.17	0.3812	A	0.7
5	☐	2500.000	2552.541	4498037.71	0.7646	A	1.0
6	☐	5000.000	5026.666	8828449.17	1.5057	A	0.8
7	☐	10000.000	9970.082	17753136.54	2.9865	A	2.0
8	☐			6810.56	0.0012	P	26.4

$$y = 2.9954E-004 * x + 2.0374E-005$$


$$R = 1.0000$$

$$DL = 0.03172$$

Weight: <None>

Min Conc: 0

51 V [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	10.00	0.0000	P	33.3
2	☐	5.000	4.957	22569.18	0.0463	P	2.1
3	☐	50.000	51.166	229557.06	0.4781	P	0.3
4	☐	125.000	127.278	545021.44	1.1892	P	0.4
5	☐	250.000	255.241	1064144.25	2.3848	P	0.7
6	☐	500.000	505.655	2102366.62	4.7245	A	0.5
7	☐	1000.000	995.520	4163101.99	9.3015	A	1.7
8	☐			1897.92	0.0042	P	5.5


$$y = 0.0093 * x + 1.9809E-005$$

$$R = 1.0000$$

$$DL = 0.002118$$

Weight: <None>

Min Conc: 0

52 Cr [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	976.71	0.0019	P	4.6
2	☐	2.000	2.043	11666.82	0.0240	P	4.2
3	☐	50.000	52.319	271801.64	0.5661	P	0.6
4	☐	125.000	129.173	639193.62	1.3947	P	0.5
5	☐	250.000	257.671	1240547.55	2.7802	P	1.2
6	☐	500.000	508.263	2439494.32	5.4822	A	1.2
7	☐	1000.000	993.313	4794610.49	10.7121	A	1.1
8	☐			97448.57	0.2133	P	3.6


$$y = 0.0108 * x + 0.0019$$

$$R = 0.9999$$

$$DL = 0.02497$$

Weight: <None>

Min Conc: 0

54 Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2913.65	0.0058	P	2.2
2	☐	50.000	59.530	16645.94	0.0342	P	5.6
3	☐	2500.000	3021.521	695629.93	1.4487	P	0.3
4	☐	6250.000	7249.939	1589412.42	3.4680	A	1.3
5	☐	12500.000	14318.125	3053512.73	6.8435	A	2.0
6	☐	25000.000	28443.301	6046929.08	13.5891	A	0.9
7	☐	50000.000	55745.711	11918300.24	26.6276	A	1.3
8	☐	250000.00	248385.40	54193070.30	118.6243	A	4.1


$$y = 4.7756E-004 * x + 0.0058$$

$$R = 0.9997$$

$$DL = 0.7803$$

Weight: <None>

Min Conc: 0



	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	960.04	0.0019	P	11.0
2	☐	1.000	0.965	4711.89	0.0097	P	6.3
3	☐	500.000	512.927	1986764.97	4.1376	A	0.3
4	☐	1250.000	1271.901	4700911.67	10.2573	A	1.1
5	☐	2500.000	2535.489	9122907.16	20.4456	A	1.2
6	☐	5000.000	5072.630	18200915.70	40.9026	A	0.7
7	☐	10000.000	9951.429	35915098.08	80.2404	A	0.9
8	☐			29549.03	0.0647	P	4.3


$$y = 0.0081 * x + 0.0019$$

$$R = 1.0000$$

$$DL = 0.07781$$

Weight: <None>

Min Conc: 0



	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	24488.90	0.0485	P	0.7
2	☐	50.000	57.899	266861.97	0.5482	P	4.0
3	☐	2500.000	2855.529	11855481.77	24.6910	A	0.6
4	☐	6250.000	7079.769	28022771.25	61.1451	A	0.2
5	☐	12500.000	14063.734	54177204.75	121.4150	A	0.6
6	☐	25000.000	28146.498	108108516.1	242.9457	A	0.5
7	☐	50000.000	55202.808	213252105.7	476.4348	A	0.6
8	☐	250000.00	248542.30	979793105.4	2,144.905	A	4.5


$$y = 0.0086 * x + 0.0485$$

$$R = 0.9997$$

$$DL = 0.1104$$

Weight: <None>

Min Conc: 0

57 Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1150.06	0.0023	P	2.8
2	☐	50.000	59.600	7380.76	0.0152	P	1.8
3	☐	2500.000	2993.787	311547.42	0.6489	P	0.7
4	☐	6250.000	7416.302	735111.78	1.6040	P	0.3
5	☐	12500.000	14353.175	1384201.71	3.1022	A	1.2
6	☐	25000.000	28294.993	2720261.80	6.1132	A	0.7
7	☐	50000.000	54587.253	5277978.08	11.7916	A	0.5
8	☐	250000.00	248626.29	24527377.41	53.6987	A	4.8


$$y = 2.1597E-004 * x + 0.0023$$

$$R = 0.9998$$

$$DL = 0.8759$$

Weight: <None>

Min Conc: 0

59 Co [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	144.45	0.0003	P	6.8
2	☐	1.000	1.058	8317.93	0.0171	P	2.7
3	☐	50.000	52.467	400096.84	0.8332	P	0.3
4	☐	125.000	129.947	945605.08	2.0633	P	0.2
5	☐	250.000	258.340	1830192.73	4.1016	A	0.7
6	☐	500.000	508.161	3590106.72	8.0677	A	0.4
7	☐	1000.000	993.093	7056925.10	15.7663	A	0.9
8	☐			20935.80	0.0459	P	6.5


$$y = 0.0159 * x + 2.8617E-004$$

$$R = 0.9999$$

$$DL = 0.003689$$

Weight: <None>

Min Conc: 0

60 Ni [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1377.86	0.0027	P	4.9
2	☐	1.000	0.932	3171.47	0.0065	P	2.8
3	☐	50.000	52.278	103186.11	0.2149	P	0.3
4	☐	125.000	128.706	240647.85	0.5251	P	0.3
5	☐	250.000	255.438	463804.03	1.0394	P	1.1
6	☐	500.000	501.503	906916.62	2.0381	P	0.8
7	☐	1000.000	997.312	1812927.32	4.0504	A	1.0
8	☐			26028.20	0.0570	P	3.4


$$y = 0.0041 * x + 0.0027$$

$$R = 1.0000$$

$$DL = 0.09851$$

Weight: <None>

Min Conc: 0

63 Cu [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1232.67	0.0024	P	9.3
2	☐	2.000	2.071	12001.12	0.0247	P	4.1
3	☐	500.000	539.651	2780345.10	5.7903	A	1.5
4	☐	1250.000	1308.500	6432870.70	14.0363	A	0.9
5	☐	2500.000	2589.056	12391579.41	27.7705	A	0.7
6	☐	5000.000	5113.198	24402871.60	54.8423	A	1.4
7	☐	10000.000	9911.842	47582454.66	106.3085	A	1.3
8	☐			16315.54	0.0357	P	3.9


$$y = 0.0107 * x + 0.0024$$

$$R = 0.9999$$

$$DL = 0.0634$$

Weight: <None>

Min Conc: 0

66 Zn [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1412.30	0.0028	P	5.9
2	☐	5.000	5.000	5655.57	0.0116	P	6.1
3	☐	500.000	535.774	455333.23	0.9483	P	0.5
4	☐	1250.000	1326.305	1073969.83	2.3434	P	0.4
5	☐	2500.000	2635.039	2076137.51	4.6529	A	1.5
6	☐	5000.000	5165.076	4057230.74	9.1178	A	1.0
7	☐	10000.000	9872.375	7799674.19	17.4249	A	0.5
8	☐			8157.85	0.0179	P	4.3


$$y = 0.0018 * x + 0.0028$$

$$R = 0.9997$$

$$DL = 0.2816$$

Weight: <None>

Min Conc: 0

75 As [He] ISTD:89 Y [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	4.44	0.0000	P	114.
2	☐	1.000	1.160	618.91	0.0002	P	3.9
3	☐	50.000	51.656	27757.12	0.0079	P	2.3
4	☐	125.000	128.016	65355.19	0.0196	P	0.9
5	☐	250.000	255.053	127990.49	0.0390	P	1.0
6	☐	500.000	501.650	254359.58	0.0767	P	0.4
7	☐	1000.000	997.452	501902.00	0.1526	P	1.0
8	☐			400.01	0.0001	P	16.9

$$y = 1.5298E-004 * x + 1.2227E-006$$


$$R = 1.0000$$

$$DL = 0.02746$$

Weight: <None>

Min Conc: 0

77 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1.11	0.0000	P	173.
2	☐	5.000	5.957	100.00	0.0000	P	11.0
3	☐	50.000	51.446	870.04	0.0002	P	10.6
4	☐	125.000	125.248	2011.27	0.0006	P	3.1
5	☐	250.000	262.237	4138.39	0.0013	P	0.3
6	☐	500.000	507.916	8097.85	0.0024	P	1.0
7	☐	1000.000	992.874	15707.31	0.0048	P	2.5
8	☐			5.56	0.0000	P	91.8

$$y = 4.8097\text{E-}006 * x + 3.0461\text{E-}007$$


$$R = 0.9999$$

$$DL = 0.3291$$

Weight: <None>

Min Conc: 0

78 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	244.45	0.0001	P	14.9
2	☐	5.000	4.783	495.57	0.0001	P	3.5
3	☐	50.000	53.691	3232.61	0.0009	P	5.8
4	☐	125.000	126.223	6919.45	0.0021	P	3.1
5	☐	250.000	257.558	13647.51	0.0042	P	0.7
6	☐	500.000	502.588	26696.30	0.0081	P	1.6
7	☐	1000.000	996.480	52313.13	0.0159	P	0.6
8	☐			216.67	0.0001	P	5.9


$$y = 1.5893\text{E-}005 * x + 6.7208\text{E-}005$$

$$R = 1.0000$$


$$DL = 1.888$$

Weight: <None>


Min Conc: 0

81 Br [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			24203.29		P	0.3
2	☐			23418.73		P	1.7
3	☐			22510.60		P	0.7
4	☐			21016.12		P	0.1
5	☐			20082.71		P	1.5
6	☐			20008.16		P	1.7
7	☐			20311.88		P	1.8
8	☐			20594.47		P	1.5

81 Br [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			133.34		P	4.3
2	☐			64.44		P	29.9
3	☐			55.55		P	6.9
4	☐			43.34		P	48.0
5	☐			56.67		P	15.6
6	☐			62.22		P	12.4
7	☐			66.66		P	8.7
8	☐			108.89		P	30.7

82 Se [No Gas] ISTD:89 Y [No Gas]

	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	336.14	0.0000	P	19.2
2	☐	5.000	4.965	1351.32	0.0001	P	3.1
3	☐	50.000	53.568	11076.81	0.0007	P	0.4
4	☐	125.000	133.602	26273.24	0.0017	P	0.4
5	☐	250.000	261.650	50295.08	0.0033	P	0.2
6	☐	500.000	515.828	98224.64	0.0065	P	0.2
7	☐	1000.000	987.920	190752.74	0.0124	P	1.8
8	☐			471.09	0.0000	P	6.6


$$y = 1.2515\text{E-}005 * x + 2.0311\text{E-}005$$

$$R = 0.9997$$

$$DL = 0.9372$$


Weight: <None>

Min Conc: 0

83 Kr [No Gas] No available calibration points

	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐			217.78		P	11.7
2	☐			183.34		P	13.1
3	☐			204.45		P	15.1
4	☐			184.45		P	16.8
5	☐			186.67		P	9.4
6	☐			181.12		P	13.3
7	☐			211.11		P	10.3
8	☐			210.00		P	6.9

86 Sr [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	426.68	0.0000	P	5.0
2	☐	1.000	0.953	4590.77	0.0003	P	3.8
3	☐	1250.000	1272.234	5447496.59	0.3397	A	0.4
4	☐	3125.000	3174.337	13156978.97	0.8475	A	0.1
5	☐	6250.000	6292.047	25641761.56	1.6798	A	0.2
6	☐	12500.000	12561.863	50868933.69	3.3536	A	0.4
7	☐	25000.000	24951.278	102603334.0	6.6612	A	1.6
8	☐			36844.06	0.0025	P	25.8

$$y = 2.6697E-004 * x + 2.5764E-005$$


$$R = 1.0000$$

$$DL = 0.01437$$

Weight: <None>

Min Conc: 0

88 Sr [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1283.40	0.0001	P	6.9
2	☐	1.000	0.939	37131.32	0.0023	P	2.3
3	☐	1250.000	1258.767	47053860.97	2.9340	A	0.6
4	☐	3125.000	3165.185	114535507.1	7.3775	A	0.2
5	☐	6250.000	6311.221	224542072.2	14.7103	A	0.4
6	☐	12500.000	12638.923	446856215.5	29.4590	A	0.9
7	☐	25000.000	24909.772	894297782.2	58.0600	A	2.1
8	☐			315636.83	0.0214	P	25.9

$$y = 0.0023 * x + 7.7519E-005$$


$$R = 1.0000$$

$$DL = 0.006894$$

Weight: <None>

Min Conc: 0

90 Zr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	715.58	0.0000	P	4.8
2	☐	1.000	0.917	22170.18	0.0014	P	3.8
3	☐	50.000	50.944	1168025.25	0.0728	P	0.6
4	☐	125.000	128.306	2846709.75	0.1834	A	0.4
5	☐	250.000	252.182	5500706.72	0.3603	A	0.4
6	☐	500.000	504.698	10937900.95	0.7211	A	0.3
7	☐	1000.000	996.645	21934032.73	1.4240	A	1.6
8	☐			15149.12	0.0010	P	12.1

$$y = 0.0014 * x + 4.3218E-005$$


$$R = 1.0000$$

$$DL = 0.004364$$

Weight: <None>

Min Conc: 0

91 Zr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	153.34	0.0000	P	6.0
2	☐	1.000	0.934	5056.48	0.0003	P	2.2
3	☐	50.000	50.690	260650.89	0.0163	P	0.8
4	☐	125.000	127.412	633985.19	0.0408	P	0.5
5	☐	250.000	253.300	1239118.53	0.0812	P	0.3
6	☐	500.000	507.484	2466828.82	0.1626	A	0.9
7	☐	1000.000	995.097	4911635.00	0.3189	A	1.3
8	☐			3464.89	0.0002	P	13.5

$$y = 3.2043E-004 * x + 9.2575E-006$$


$$R = 1.0000$$

$$DL = 0.005183$$

Weight: <None>

Min Conc: 0

94 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	306.67	0.0000	P	10.0
2	☐	5.000	5.639	26145.62	0.0016	P	2.5
3	☐	500.000	503.411	2258296.60	0.1408	A	1.3
4	☐	1250.000	1262.445	5481544.71	0.3531	A	0.4
5	☐	2500.000	2507.968	10706862.48	0.7014	A	0.3
6	☐	5000.000	5069.556	21506456.48	1.4178	A	0.9
7	☐	10000.000	9961.503	42911452.97	2.7859	A	1.8
8	☐			27806.85	0.0019	P	13.0

$$y = 2.7967E-004 * x + 1.8524E-005$$


$$R = 1.0000$$

$$DL = 0.0198$$

Weight: <None>

Min Conc: 0

95 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	122.22	0.0000	P	22.1
2	☐	5.000	4.752	31402.56	0.0019	P	2.6
3	☐	500.000	512.500	3301694.43	0.2059	A	0.3
4	☐	1250.000	1277.512	7967195.23	0.5132	A	0.4
5	☐	2500.000	2517.251	15436129.08	1.0112	A	0.7
6	☐	5000.000	5064.287	30858194.26	2.0344	A	0.5
7	☐	10000.000	9959.480	61624385.20	4.0008	A	1.6
8	☐			39045.16	0.0027	P	14.0

$$y = 4.0171E-004 * x + 7.3796E-006$$


$$R = 1.0000$$

$$DL = 0.0122$$

Weight: <None>

Min Conc: 0

96 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	113.33	0.0000	P	12.8
2	☐	5.000	4.832	34745.66	0.0021	P	1.3
3	☐	500.000	509.099	3571021.41	0.2227	A	0.0
4	☐	1250.000	1291.633	8770485.15	0.5649	A	0.4
5	☐	2500.000	2519.147	16818820.31	1.1018	A	0.8
6	☐	5000.000	5074.099	33662915.05	2.2193	A	0.5
7	☐	10000.000	9952.505	67048116.22	4.3529	A	2.0
8	☐			48441.01	0.0033	P	12.3

$$y = 4.3737\text{E-}004 * x + 6.8440\text{E-}006$$


$$R = 1.0000$$

$$DL = 0.006031$$

Weight: <None>

Min Conc: 0

97 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	60.00	0.0000	P	14.6
2	☐	5.000	4.698	19247.20	0.0012	P	1.6
3	☐	500.000	507.840	2029969.41	0.1266	A	0.6
4	☐	1250.000	1276.463	4939525.03	0.3182	A	0.4
5	☐	2500.000	2513.250	9562386.04	0.6264	A	1.1
6	☐	5000.000	5058.658	19125592.49	1.2609	A	0.6
7	☐	10000.000	9963.659	38252673.87	2.4835	A	2.0
8	☐			24100.29	0.0016	P	14.9

$$y = 2.4925\text{E-}004 * x + 3.6230\text{E-}006$$


$$R = 1.0000$$

$$DL = 0.006377$$

Weight: <None>

Min Conc: 0

98 Mo [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	126.67	0.0000	P	11.9
2	☐	5.000	4.701	49771.08	0.0030	P	1.6
3	☐	500.000	505.691	5226866.52	0.3259	A	0.7
4	☐	1250.000	1268.322	12690635.92	0.8174	A	0.6
5	☐	2500.000	2524.972	24840291.57	1.6273	A	0.0
6	☐	5000.000	5069.526	49559825.93	3.2672	A	0.7
7	☐	10000.000	9956.419	98837802.97	6.4168	A	1.8
8	☐			61905.14	0.0042	P	14.1

$$y = 6.4449\text{E-}004 * x + 7.6516\text{E-}006$$


$$R = 1.0000$$

$$DL = 0.004223$$

Weight: <None>

Min Conc: 0

106 Cd [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	7756.59	0.0006	P	1.9
2	☐	1.000	0.931	8022.28	0.0006	P	2.7
3	☐	50.000	55.496	26502.00	0.0021	P	0.5
4	☐	125.000	132.982	52687.33	0.0042	P	1.4
5	☐	250.000	259.773	94172.54	0.0077	P	0.9
6	☐	500.000	511.130	177998.12	0.0145	P	1.4
7	☐	1000.000	990.720	343714.40	0.0275	P	1.6
8	☐			7147.38	0.0006	P	1.8

$$y = 2.7171\text{E-}005 * x + 6.1128\text{E-}004$$


$$R = 0.9998$$

$$DL = 1.302$$

Weight: <None>

Min Conc: 0

107 Ag [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	116.67	0.0000	P	17.6
2	☐	1.000	1.001	15120.19	0.0012	P	3.2
3	☐	50.000	55.662	828056.33	0.0662	P	1.2
4	☐	125.000	134.990	2002574.71	0.1606	A	1.3
5	☐	250.000	265.304	3874359.84	0.3156	A	2.1
6	☐	500.000	514.207	7508254.96	0.6116	A	1.4
7	☐	1000.000	987.539	14663929.23	1.1745	A	1.1
8	☐			5581.17	0.0005	P	23.1

$$y = 0.0012 * x + 9.1970E-006$$


$$R = 0.9997$$

$$DL = 0.004085$$

Weight: <None>

Min Conc: 0

108 Cd [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	27.78	0.0000	P	56.9
2	☐	1.000	0.933	254.45	0.0000	P	0.7
3	☐	50.000	55.443	13400.72	0.0011	P	3.8
4	☐	125.000	134.177	32298.13	0.0026	P	0.6
5	☐	250.000	261.630	61980.46	0.0050	P	0.5
6	☐	500.000	515.016	121968.47	0.0099	P	1.0
7	☐	1000.000	988.165	237941.72	0.0191	P	0.6
8	☐			198.89	0.0000	P	8.1


$$y = 1.9285E-005 * x + 2.1913E-006$$

$$R = 0.9997$$

$$DL = 0.1939$$

Weight: <None>

Min Conc: 0

109 Ag [No Gas] ISTD :159 Tb [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	36.67	0.0000	P	27.6
2	☐	1.000	0.986	14406.13	0.0011	P	2.2
3	☐	50.000	54.508	788495.02	0.0630	P	0.6
4	☐	125.000	133.982	1932718.44	0.1550	A	0.6
5	☐	250.000	263.531	3742729.70	0.3048	A	1.2
6	☐	500.000	513.629	7294197.53	0.5941	A	0.9
7	☐	1000.000	988.454	14273656.18	1.1433	A	0.7
8	☐			5295.50	0.0005	P	21.9


$$y = 0.0012 * x + 2.8915E-006$$

$$R = 0.9997$$

$$DL = 0.002067$$

Weight: <None>

Min Conc: 0

111 Cd [No Gas] ISTD :159 Tb [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	5441.28	0.0004	P	2.0
2	☐	1.000	1.067	8715.66	0.0007	P	2.0
3	☐	50.000	54.292	172649.26	0.0138	P	0.9
4	☐	125.000	131.842	410449.88	0.0329	P	1.2
5	☐	250.000	261.116	795183.98	0.0648	P	1.4
6	☐	500.000	513.438	1558399.45	0.1269	P	0.5
7	☐	1000.000	989.432	3048781.45	0.2442	A	1.2
8	☐			6001.78	0.0005	P	4.0

$$y = 2.4638E-004 * x + 4.2882E-004$$


$$R = 0.9998$$

$$DL = 0.105$$

Weight: <None>

Min Conc: 0

118 Sn [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	2487.06	0.0002	P	12.6
2	☐	5.000	4.927	53870.14	0.0043	P	2.0
3	☐	50.000	52.367	544654.23	0.0436	P	0.9
4	☐	125.000	127.801	1321990.48	0.1060	P	0.8
5	☐	250.000	256.342	2608142.72	0.2124	A	1.9
6	☐	500.000	509.067	5177172.49	0.4217	A	0.3
7	☐	1000.000	993.413	10270580.33	0.8227	A	1.1
8	☐			13405.24	0.0011	P	5.6

$$y = 8.2791\text{E-}004 * x + 1.9605\text{E-}004$$


$$R = 0.9999$$

$$DL = 0.08956$$

Weight: <None>

Min Conc: 0

121 Sb [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	54.44	0.0000	P	33.7
2	☐	2.000	1.974	23449.19	0.0019	P	2.9
3	☐	50.000	51.805	609328.12	0.0487	P	0.7
4	☐	125.000	127.276	1492784.56	0.1197	P	0.6
5	☐	250.000	259.816	3000266.24	0.2443	A	1.1
6	☐	500.000	509.366	5880865.54	0.4790	A	1.3
7	☐	1000.000	992.488	11652805.52	0.9334	A	1.5
8	☐			10288.20	0.0009	P	3.9


$$y = 9.4042\text{E-}004 * x + 4.2909\text{E-}006$$

$$R = 0.9999$$

$$DL = 0.004617$$

Weight: <None>

Min Conc: 0

135 Ba [No Gas] ISTD : 159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	118.89	0.0000	P	18.5
2	☐	10.000	9.322	31087.30	0.0025	P	0.9
3	☐	250.000	248.417	819229.65	0.0655	P	0.7
4	☐	625.000	636.467	2092819.40	0.1678	A	1.2
5	☐	1250.000	1267.689	4104144.10	0.3342	A	1.6
6	☐	2500.000	2527.408	8180767.44	0.6664	A	1.3
7	☐	5000.000	4980.521	16394423.09	1.3132	A	1.0
8	☐			9963.62	0.0009	P	14.5


$$y = 2.6366E-004 * x + 9.3676E-006$$

$$R = 1.0000$$

$$DL = 0.0197$$

Weight: <None>

Min Conc: 0

137 Ba [No Gas] ISTD : 159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	234.45	0.0000	P	9.9
2	☐	10.000	9.331	53352.91	0.0042	P	1.5
3	☐	250.000	251.641	1422076.13	0.1137	P	1.4
4	☐	625.000	640.964	3611846.51	0.2896	A	1.0
5	☐	1250.000	1266.957	7028148.58	0.5724	A	1.8
6	☐	2500.000	2529.685	14032375.49	1.1429	A	0.4
7	☐	5000.000	4978.842	28082813.75	2.2494	A	0.8
8	☐			17073.90	0.0015	P	14.1


$$y = 4.5179E-004 * x + 1.8475E-005$$

$$R = 1.0000$$


$$DL = 0.01218$$

Weight: <None>


Min Conc: 0

150 Nd [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			4.44		P	43.4
2	☐			13.33		P	25.0
3	☐			107.78		P	18.6
4	☐			213.34		P	5.4
5	☐			478.90		P	10.1
6	☐			854.48		P	7.3
7	☐			1754.57		P	2.5
8	☐			271.12		P	6.3

150 Nd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			2.22		P	173.
2	☐			4.44		P	114.
3	☐			15.55		P	44.6
4	☐			37.78		P	25.5
5	☐			86.67		P	21.4
6	☐			125.56		P	16.2
7	☐			252.23		P	8.6
8	☐			144.45		P	16.2

156 Dy [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			11.11		P	45.8
2	☐			12.22		P	31.5
3	☐			26.67		P	33.1
4	☐			43.33		P	20.3
5	☐			120.00		P	5.6
6	☐			177.78		P	33.6
7	☐			351.12		P	14.5
8	☐			441.12		P	5.4

156 Dy [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			6.67		P	50.0
2	☐			8.89		P	57.3
3	☐			38.89		P	19.8
4	☐			87.78		P	12.2
5	☐			151.12		P	4.6
6	☐			332.23		P	5.8
7	☐			698.91		P	4.5
8	☐			317.78		P	2.4

160 Gd [No Gas] Noavailable calibration poin...


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			75.56		P	22.6
2	☐			80.00		P	29.2
3	☐			72.22		P	5.3
4	☐			77.78		P	19.8
5	☐			96.66		P	6.0
6	☐			142.22		P	9.5
7	☐			198.89		P	12.6
8	☐			425.57		P	12.7

160 Gd [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			421.12		P	12.3
2	☐			350.01		P	13.2
3	☐			375.57		P	15.3
4	☐			346.68		P	11.7
5	☐			354.45		P	7.1
6	☐			368.90		P	12.5
7	☐			430.01		P	4.3
8	☐			595.57		P	5.1


164 Er [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			82.22		P	8.4
2	☐			82.22		P	6.2
3	☐			95.56		P	8.8
4	☐			93.33		P	40.3
5	☐			83.33		P	16.0
6	☐			144.45		P	20.9
7	☐			206.67		P	20.6
8	☐			463.35		P	12.3

164 Er [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			55.56		P	15.1
2	☐			47.78		P	28.2
3	☐			47.78		P	35.8
4	☐			63.33		P	5.3
5	☐			63.33		P	9.1
6	☐			94.44		P	10.2
7	☐			130.00		P	5.1
8	☐			305.56		P	12.2

203 Tl [No Gas] ISTD :209 Bi [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	224.45	0.0000	P	23.0
2	☐	1.000	0.849	11175.76	0.0016	P	1.3
3	☐	50.000	46.253	594881.70	0.0838	P	0.7
4	☐	125.000	114.919	1473974.98	0.2081	P	0.5
5	☐	250.000	250.410	3195900.13	0.4533	A	1.3
6	☐	500.000	499.902	6316462.61	0.9050	A	1.1
7	☐	1000.000	1001.394	12404885.51	1.8128	A	1.1
8	☐			5390.04	0.0009	P	18.0

$$y = 0.0018 * x + 3.0693E-005$$


$$R = 0.9999$$

$$DL = 0.0117$$

Weight: <None>

Min Conc: 0

205 Tl [No Gas] ISTD :209 Bi [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	505.58	0.0001	P	24.8
2	☐	1.000	0.838	26051.69	0.0037	P	2.3
3	☐	50.000	45.863	1393914.67	0.1963	P	0.7
4	☐	125.000	123.594	3745873.87	0.5288	A	0.7
5	☐	250.000	250.027	7540978.01	1.0696	A	1.4
6	☐	500.000	498.724	14890964.92	2.1334	A	0.5
7	☐	1000.000	1001.014	29301662.34	4.2821	A	2.1
8	☐			12987.61	0.0022	P	19.9

$$y = 0.0043 * x + 6.9137E-005$$


$$R = 1.0000$$

$$DL = 0.01202$$

Weight: <None>

Min Conc: 0

206 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1270.08	0.0002	P	9.4
2	☐	1.000	0.862	9952.62	0.0014	P	1.4
3	☐	250.000	249.758	2513965.61	0.3540	A	1.1
4	☐	625.000	621.010	6232560.39	0.8798	A	1.5
5	☐	1250.000	1251.126	12495987.87	1.7724	A	0.8
6	☐	2500.000	2496.176	24682371.85	3.5359	A	0.7
7	☐	5000.000	5002.141	48487218.72	7.0856	A	2.7
8	☐			34955.33	0.0059	P	9.4

$$y = 0.0014 * x + 1.7368E-004$$


$$R = 1.0000$$

$$DL = 0.03446$$

Weight: <None>

Min Conc: 0

207 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1110.06	0.0002	P	8.9
2	☐	1.000	0.817	8481.60	0.0012	P	1.5
3	☐	250.000	248.904	2245260.18	0.3161	A	0.4
4	☐	625.000	620.525	5581772.42	0.7879	A	0.4
5	☐	1250.000	1250.358	11191990.67	1.5875	A	0.5
6	☐	2500.000	2492.092	22084671.48	3.1638	A	0.5
7	☐	5000.000	5004.479	43475011.30	6.3532	A	2.8
8	☐			30886.60	0.0052	P	10.2

$$y = 0.0013 * x + 1.5180E-004$$


$$R = 1.0000$$

$$DL = 0.03192$$

Weight: <None>

Min Conc: 0

208 Pb [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	4985.98	0.0007	P	5.2
2	☐	1.000	0.850	39462.83	0.0055	P	1.1
3	☐	250.000	249.306	10115736.26	1.4242	A	0.9
4	☐	625.000	624.673	25274208.58	3.5676	A	0.8
5	☐	1250.000	1252.268	50419513.14	7.1512	A	0.5
6	☐	2500.000	2498.554	99587920.88	14.2676	A	0.3
7	☐	5000.000	5000.231	195383931.8	28.5524	A	2.3
8	☐			139835.09	0.0236	P	9.9

$$y = 0.0057 * x + 6.8184E-004$$


$$R = 1.0000$$

$$DL = 0.01873$$

Weight: <None>

Min Conc: 0

238 U [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	43.34	0.0000	P	106.
2	☐	1.000	0.799	34653.86	0.0049	P	1.7
3	☐	50.000	48.237	2081008.53	0.2930	A	0.7
4	☐	125.000	122.559	5273918.01	0.7444	A	1.0
5	☐	250.000	244.165	10456465.05	1.4831	A	0.9
6	☐	500.000	498.197	21122014.68	3.0261	A	0.3
7	☐	1000.000	1002.754	41680512.71	6.0908	A	1.5
8	☐			13361.59	0.0022	P	25.1


$$y = 0.0061 * x + 5.9255E-006$$

$$R = 1.0000$$


$$DL = 0.003119$$

Weight: <None>


Min Conc: 0

6 Li (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2612458.88		A	1.3
2	☐	1.000		2632104.79		A	0.7
3	☐	1.000		2670582.34		A	2.0
4	☐	1.000		2572638.49		A	2.2
5	☐	1.000		2567320.76		A	0.9
6	☐	1.000		2466309.20		A	2.0
7	☐	1.000		2491079.95		A	0.5
8	☐	1.000		2293535.24		A	0.6

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6603097.61		A	0.8
2	☐	1.000		6556370.46		A	0.4
3	☐	1.000		6371787.40		A	1.2
4	☐	1.000		6141457.89		A	1.1
5	☐	1.000		5882544.64		A	0.9
6	☐	1.000		5863308.04		A	1.0
7	☐	1.000		5945135.82		A	1.0
8	☐	1.000		5865522.20		A	0.4

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		504802.78		P	0.2
2	☐	1.000		487425.98		P	4.6
3	☐	1.000		480165.05		P	0.7
4	☐	1.000		458299.60		P	0.1
5	☐	1.000		446226.82		P	0.7
6	☐	1.000		444997.71		P	0.8
7	☐	1.000		447616.77		P	0.9
8	☐	1.000		457414.84		P	4.5

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		16560234.06		A	0.5
2	☐	1.000		16387933.09		A	0.5
3	☐	1.000		16037184.76		A	1.0
4	☐	1.000		15524943.24		A	0.4
5	☐	1.000		15264611.30		A	0.8
6	☐	1.000		15167978.39		A	1.0
7	☐	1.000		15403324.91		A	0.2
8	☐	1.000		14683319.92		A	1.6

89 Y (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3636459.91		A	0.2
2	☐	1.000		3460615.71		A	4.1
3	☐	1.000		3512591.86		A	1.3
4	☐	1.000		3337106.30		A	0.8
5	☐	1.000		3280239.43		A	0.6
6	☐	1.000		3314408.56		A	0.5
7	☐	1.000		3289270.75		A	0.6
8	☐	1.000		3290281.48		A	3.6

103 Rh (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		8939206.47		A	0.1
2	☐	1.000		8836047.85		A	0.8
3	☐	1.000		8540268.35		A	1.3
4	☐	1.000		8340395.85		A	0.6
5	☐	1.000		7979150.16		A	1.6
6	☐	1.000		7822496.55		A	1.9
7	☐	1.000		7757691.27		A	0.4
8	☐	1.000		7038153.23		A	1.4

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4965866.52		A	0.6
2	☐	1.000		4729552.92		A	4.7
3	☐	1.000		4751168.44		A	1.1
4	☐	1.000		4482846.88		A	0.6
5	☐	1.000		4427755.53		A	0.6
6	☐	1.000		4353271.19		A	0.5
7	☐	1.000		4295548.13		A	0.1
8	☐	1.000		3965444.80		A	4.5

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		9664796.72		A	0.2
2	☐	1.000		9585171.34		A	0.6
3	☐	1.000		9301513.68		A	0.9
4	☐	1.000		9163778.43		A	1.3
5	☐	1.000		8874418.83		A	0.5
6	☐	1.000		8742499.83		A	1.0
7	☐	1.000		8701423.23		A	0.6
8	☐	1.000		8455796.32		A	2.7

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2965612.15		A	0.3
2	☐	1.000		2843921.29		A	4.7
3	☐	1.000		2830252.53		A	1.8
4	☐	1.000		2725839.24		A	1.1
5	☐	1.000		2687777.14		A	0.1
6	☐	1.000		2641467.19		A	0.5
7	☐	1.000		2604516.42		A	0.3
8	☐	1.000		2656422.48		A	3.2

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		12689058.70		A	0.3
2	☐	1.000		12601038.01		A	0.6
3	☐	1.000		12505588.01		A	0.8
4	☐	1.000		12471877.31		A	1.2
5	☐	1.000		12279156.62		A	0.9
6	☐	1.000		12278511.90		A	2.1
7	☐	1.000		12484420.37		A	0.3
8	☐	1.000		11651791.77		A	1.5

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7932531.27		A	0.6
2	☐	1.000		7602056.00		A	5.0
3	☐	1.000		7735861.00		A	2.1
4	☐	1.000		7621953.50		A	0.5
5	☐	1.000		7495108.64		A	0.7
6	☐	1.000		7583698.43		A	0.3
7	☐	1.000		7515656.90		A	1.2
8	☐	1.000		7339009.06		A	4.9

165 Ho (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		12316146.48		A	1.0
2	☐	1.000		12198877.46		A	0.7
3	☐	1.000		12051069.13		A	1.9
4	☐	1.000		12088165.79		A	1.0
5	☐	1.000		11935712.60		A	1.7
6	☐	1.000		11917768.02		A	1.6
7	☐	1.000		12108683.71		A	0.5
8	☐	1.000		11328035.39		A	1.9

165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7840700.72		A	0.7
2	☐	1.000		7522005.51		A	5.6
3	☐	1.000		7695351.76		A	0.4
4	☐	1.000		7526002.32		A	1.1
5	☐	1.000		7478859.68		A	0.8
6	☐	1.000		7498774.20		A	0.6
7	☐	1.000		7473394.82		A	0.6
8	☐	1.000		7251418.92		A	4.5

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7312476.35		A	0.0
2	☐	1.000		7132730.17		A	0.5
3	☐	1.000		7102143.51		A	1.3
4	☐	1.000		7084619.69		A	1.3
5	☐	1.000		7050696.22		A	1.7
6	☐	1.000		6980101.84		A	1.6
7	☐	1.000		6843401.77		A	1.3
8	☐	1.000		5926640.12		A	1.8

209 Bi (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5294232.66		A	1.2
2	☐	1.000		4990773.05		A	4.9
3	☐	1.000		5177782.70		A	0.4
4	☐	1.000		5016448.08		A	0.9
5	☐	1.000		4977314.61		A	0.8
6	☐	1.000		4895058.68		A	0.8
7	☐	1.000		4717587.78		A	1.6
8	☐	1.000		4312835.98		A	4.7

US EPA Tune Check Report

Operator Name Jaswal

Acq/Data Batch D:\Agilent\ICPMH\1\DATA\IP7042524-MS3.b

Acq. Date-Time 2024-04-25 11:22:19

Report Comment ---

Instrument Name G8403A JP14410463

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		17374	173742.24			0.956	5.000
24		111919	1119189.96			0.715	5.000
25		15578	155779.42			0.489	5.000
26		18437	184373.16			0.110	5.000
59		134053	1340533.73			0.501	5.000
113		17042	170416.73			0.472	5.000
115		207406	2074055.05			0.660	5.000
206		44634	446341.15			0.370	5.000
207		40859	408594.78			0.317	5.000
208		104011	1040114.32			0.784	5.000
220		0	4.20			41.752	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	


Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	17598	17485	17345	17230	17213
24	110935	111305	112782	112607	111965
25	15456	15666	15597	15591	15579
26	18459	18441	18452	18426	18408
59	135151	134104	133714	133926	133371
113	17160	17069	17045	16962	16972
115	208899	207551	208479	206496	205602
206	44715	44537	44870	44442	44608
207	40931	40879	41027	40721	40739
208	104872	104301	104408	103730	102746

US EPA Tune Check Report

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	0	1	0	0	0

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	29715.76	9.00	8.90 - 9.10	
24	186308.95	24.00	23.90 - 24.10	
25	25835.93	25.00	24.90 - 25.10	
26	30143.34	25.95	25.90 - 26.10	
59	224612.80	58.95	58.90 - 59.10	
113	31113.30	113.00	112.90 - 113.10	
115	376396.02	115.00	114.90 - 115.10	
206	84854.97	206.00	205.90 - 206.10	
207	76621.95	207.00	206.90 - 207.10	
208	191407.27	208.00	207.90 - 208.10	
220	0.60	220.15	-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.61	0.737	0.900	
24	0.63	0.817	0.900	
25	0.63	0.742	0.900	
26	0.64	0.784	0.900	
59	0.62	0.770	0.900	
113	0.56	0.732	0.900	
115	0.56	0.767	0.900	
206	0.54	0.780	0.900	
207	0.54	0.743	0.900	
208	0.56	0.813	0.900	
220	0.26	0.442		

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode --- Nebulizer Gas 0.82 L/min Dilution Gas 0.40 L/min

US EPA Tune Check Report

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	7.5 V	Deflect	13.8 V
Extract 2	-135.0 V	Cell Entrance	-30 V	Plate Bias	-35 V
Omega Bias	-75 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	190 V		

QP Parameters

Mass Gain	133	Axis Gain	0.9976	QP Bias	-3.0 V
Mass Offset	130	Axis Offset	0.13		

Hardware Settings

Torch

Torch H	-0.7 mm	Torch V	0.9 mm
---------	---------	---------	--------

EM

Discriminator	4.7 mV	Analog HV	2302 V	Pulse HV	1348 V
---------------	--------	-----------	--------	----------	--------

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		40447	404473.95			0.749	
89		37724	377241.73			0.299	
205		45690	456899.96			0.527	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	40824	40714	40313	40134	40252
89	37787	37859	37710	37555	37710
205	45922	45841	45809	45351	45527

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	---	Nebulizer Gas	0.82 L/min	Dilution Gas	0.40 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	8.2 V	Deflect	2.4 V
-----------	-------	------------	-------	---------	-------

US EPA Tune Check Report

Extract 2	-140.0 V	Cell Entrance	-40 V	Plate Bias	-60 V
Omega Bias	-75 V	Cell Exit	-60 V		

Cell Parameters

Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	3.9 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	200 V		

QP Parameters

Mass Gain	133	Axis Gain	0.9976	QP Bias	-13.0 V
Mass Offset	130	Axis Offset	0.13		

Hardware Settings

Torch

Torch H	-0.7 mm	Torch V	0.9 mm
---------	---------	---------	--------

EM

Discriminator	4.7 mV	Analog HV	2302 V	Pulse HV	1348 V
---------------	--------	-----------	--------	----------	--------