

66130511

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : Rm Instrument ID : Konelab

4/26/2024 12:39

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.011	0.0	0.116	
ICB1	0.004	0.0	0.009	
CCV1	1.005	0.0	0.116	
CCB1	0.005	0.0	0.009	
PB160540BL	0.001	0.0	0.008	
PB160540BS	1.008	0.0	0.116	
P2199-01	0.002	0.0	0.009	
P2199-01DUP	0.000	0.0	0.008	
P2199-01MS	1.012	0.0	0.116	
P2199-01MSD	1.015	0.0	0.117	
CCV2	1.005	0.0	0.116	
CCB2	0.005	0.0	0.009	

N	12
Mean	0.506
SD	0.5256
CV%	103.87

Aquakem v. 7.2AQ1

Results from time period:

Fri Apr 26 12:23:04 2024

Fri Apr 26 12:33:17 2024

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0012	mg/l	4/26/2024 8:27	
0.05PPM	A	Phenolics-	P	0.0523	mg/l	4/26/2024 8:27	
0.1PPM	A	Phenolics-	P	0.1044	mg/l	4/26/2024 8:27	
0.25PPM	A	Phenolics-	P	0.2532	mg/l	4/26/2024 8:27	
0.50PPM	A	Phenolics-	P	0.4957	mg/l	4/26/2024 8:27	
1.0PPM	A	Phenolics-	P	0.9906	mg/l	4/26/2024 8:27	
2.0PPM	A	Phenolics-	P	2.0051	mg/l	4/26/2024 8:27	
ICV1	S	Phenolics-	P	1.0105	mg/l	4/26/2024 12:23	
ICB1	S	Phenolics-	P	0.0042	mg/l	4/26/2024 12:23	
CCV1	S	Phenolics-	P	1.0052	mg/l	4/26/2024 12:23	
CCB1	S	Phenolics-	P	0.0047	mg/l	4/26/2024 12:23	
PB160540BL	S	Phenolics-	P	0.0009	mg/l	4/26/2024 12:23	
PB160540BS	S	Phenolics-	P	1.0084	mg/l	4/26/2024 12:23	
P2199-01	S	Phenolics-	P	0.002	mg/l	4/26/2024 12:33	
P2199-01DUP	S	Phenolics-	P	0.0002	mg/l	4/26/2024 12:33	
P2199-01MS	S	Phenolics-	P	1.012	mg/l	4/26/2024 12:33	
P2199-01MSD	S	Phenolics-	P	1.0149	mg/l	4/26/2024 12:33	
CCV2	S	Phenolics-	P	1.0049	mg/l	4/26/2024 12:33	
CCB2	S	Phenolics-	P	0.005	mg/l	4/26/2024 12:33	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

4/26/2024 9:24


Test Phenolics-

Accepted 4/26/2024 9:24

Factor 9.371
Bias 0.008

Coeff. of det. 0.999947

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	-0.0012	0.0000	-
2	Phenol-2	0.014	0.0523	0.0500	4.6
3	Phenol-2	0.020	0.1044	0.1000	4.4
4	Phenol-2	0.035	0.2532	0.2500	1.3
5	Phenol-2	0.061	0.4957	0.5000	-0.9
6	Phenol-2	0.114	0.9906	1.0000	-0.9
7	Phenol-2	0.222	2.0051	2.0000	0.3

04/26/2024
RM