


LB130513

SFAM01.1

INSTRUMENT ID: CV-1



Linear

A= 0.0000e+000

B= 2.2242e-004 slope

C= -9.4627e-002 intercept

Rho= 0.9998770

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9.12
0.00	0.000	0.043	0.043	618	0.000	618					
0.05	0.050					442					
0.20	0.200	0.210	0.010	1370	0.0 %	1370	0				
2.50	2.500	2.387	-0.113	11156	0.0 %	11156					
5.00	5.000	5.072	0.072	23231	0.0 %	23231					
7.50	7.500	7.493	-0.007	34113	0.0 %	34113					
10.0	10.000	9.995	-0.005	45366	0.0 %	45366					

11-5-911

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	618	-		0 SFAM01.1	PPB	4/26/2024 12:08	Std
	0.2 S01	1370	-		0.2 SFAM01.1	PPB	4/26/2024 12:10	Std
	2.5 S02	11156	-		2.5 SFAM01.1	PPB	4/26/2024 12:12	Std
	5 S03	23231	-		5 SFAM01.1	PPB	4/26/2024 12:16	Std
	7.5 S04	34113	-		7.5 SFAM01.1	PPB	4/26/2024 12:19	Std
	10 S05	45366	-		10 SFAM01.1	PPB	4/26/2024 12:21	Std
ICV016	ICV016	18461	4.0114 -		SFAM01.1	PPB	4/26/2024 12:24	SMPL
ICB016	ICB016	467	0.0092 -		SFAM01.1	PPB	4/26/2024 12:26	SMPL
CCV080	CCV080	24257	5.3005 -		SFAM01.1	PPB	4/26/2024 12:28	SMPL
CCB080	CCB080	413	-0.0028 -		SFAM01.1	PPB	4/26/2024 12:31	SMPL
PB160558BL	PBS558	652	0.0504 -		SFAM01.1	PPB	4/26/2024 12:33	SMPL
P2102-01	MCOAA0	9052	1.9187 -		SFAM01.1	PPB	4/26/2024 12:35	SMPL
P2102-02	MCOAA3	1132	0.1571 -		SFAM01.1	PPB	4/26/2024 12:37	SMPL
P2102-03	MCOAA5	2465	0.4536 -		SFAM01.1	PPB	4/26/2024 12:40	SMPL
P2102-04	MCOAA6	3179	0.6124 -		SFAM01.1	PPB	4/26/2024 12:42	SMPL
P2102-05	MCOAA7	649	0.0497 -		SFAM01.1	PPB	4/26/2024 12:44	SMPL
P2102-06	MCOAA7D	564	0.0308 -		SFAM01.1	PPB	4/26/2024 12:46	SMPL
P2102-07	MCOAA7S	11467	2.4558 -		SFAM01.1	PPB	4/26/2024 12:49	SMPL
P2102-08	MCOAA8	3972	0.7888 -		SFAM01.1	PPB	4/26/2024 12:51	SMPL
P2102-09	MCOAA9	2913	0.5533 -		SFAM01.1	PPB	4/26/2024 12:53	SMPL
P2102-10	MCOAB0	1020	0.1322 -		SFAM01.1	PPB	4/26/2024 12:56	SMPL
P2102-11	MCOAB1	535	0.0244 -		SFAM01.1	PPB	4/26/2024 12:58	SMPL
P2102-12	MCOAC2	1325	0.2001 -		SFAM01.1	PPB	4/26/2024 13:00	SMPL
P2102-13	MCOAC3	915	0.1089 -		SFAM01.1	PPB	4/26/2024 13:02	SMPL
P2102-14	MCOAC5	1036	0.1358 -		SFAM01.1	PPB	4/26/2024 13:05	SMPL
P2102-15	MCOAC6	1467	0.2317 -		SFAM01.1	PPB	4/26/2024 13:07	SMPL
P2102-16	MCOAC8	2215	0.398 -		SFAM01.1	PPB	4/26/2024 13:09	SMPL
P2102-17	MCOAC9	1356	0.207 -		SFAM01.1	PPB	4/26/2024 13:11	SMPL
P2102-18	MCOAD0	1263	0.1863 -		SFAM01.1	PPB	4/26/2024 13:14	SMPL
P2102-19	MCOAD1	995	0.1267 -		SFAM01.1	PPB	4/26/2024 13:16	SMPL
CCV081	CCV081	25236	5.5183 -		SFAM01.1	PPB	4/26/2024 13:18	SMPL
CCB081	CCB081	566	0.0313 -		SFAM01.1	PPB	4/26/2024 13:21	SMPL
P2102-20	MCOAL7	3286	0.6362 -		SFAM01.1	PPB	4/26/2024 13:23	SMPL
P2102-21	MCOAL8	4620	0.9329 -		SFAM01.1	PPB	4/26/2024 13:25	SMPL
PB160559BL	PBS559	596	0.0379 -		SFAM01.1	PPB	4/26/2024 13:27	SMPL
P2103-01	MCOAC7	563	0.0306 -		SFAM01.1	PPB	4/26/2024 13:30	SMPL
P2103-02	MCOAC7D	530	0.0233 -		SFAM01.1	PPB	4/26/2024 13:32	SMPL
P2103-03	MCOAC7S	12043	2.5839 -		SFAM01.1	PPB	4/26/2024 13:34	SMPL
P2103-04	MCOAB2	396	-0.0065 -		SFAM01.1	PPB	4/26/2024 13:36	SMPL
P2103-05	MCOAB3	1125	0.1556 -		SFAM01.1	PPB	4/26/2024 13:39	SMPL
P2103-06	MCOAB4	21985	4.7952 -		SFAM01.1	PPB	4/26/2024 13:41	SMPL
P2103-07	MCOAD2	378	-0.0106 -		SFAM01.1	PPB	4/26/2024 13:43	SMPL
P2103-08	MCOAD3	992	0.126 -		SFAM01.1	PPB	4/26/2024 13:46	SMPL
P2103-09	MCOAD4	1010	0.13 -		SFAM01.1	PPB	4/26/2024 13:48	SMPL
CCV082	CCV082	25314	5.5356 -		SFAM01.1	PPB	4/26/2024 13:56	SMPL
CCB082	CCB082	555	0.0288 -		SFAM01.1	PPB	4/26/2024 13:58	SMPL

LB130513

INSTRUMENT ID : CV1

P2047-16DLX25	GCH22	20270	4.4137 -	SFAM01.1	PPB	4/26/2024 14:07	SMPL
CCV083	CCV083	23146	5.0534 -	SFAM01.1	PPB	4/26/2024 14:09	SMPL
CCB083	CCB083	484	0.013 -	SFAM01.1	PPB	4/26/2024 14:12	SMPL