

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130517

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/26/2024	10:40
2	P2263-01		1	4	7.00	44.6	43.6	04/26/2024	11:10
3	P2263-02		1	12	100.00	>212.0	>212.0	04/26/2024	11:40
4	P2263-02DUP		1	13	100.00	>212.0	>212.0	04/26/2024	12:10
5	P2263-03		1	14	100.00	>212.0	>212.0	04/26/2024	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130517

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/26/24 flashpoint WorkList ID : 179793 Department : Wet-Chemistry Date : 04-26-2024 08:04:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2263-01	60105	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B
P2263-02	60156	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B
P2263-03	60157	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/24/2024	1010B

Date/Time 04/26/2024 10:30
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: J.D.Garcia

Date/Time 04/26/2024 13:25
Raw Sample Received by: J.D.Garcia
Raw Sample Relinquished by: N.P.W.C