

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130520

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/26/2024	13:40
2	CAL2	1	Water	NA	NA	20.3	7.00	04/26/2024	13:41
3	CAL3	1	Water	NA	NA	20.3	10.03	04/26/2024	13:42
4	ICV	1	Water	NA	NA	20.3	7.01	04/26/2024	13:45
5	CCV1	1	Water	NA	NA	20.2	2.01	04/26/2024	13:50
6	P2283-05	1	Water	NA	NA	20.7	7.90	04/26/2024	13:55
7	P2283-05DUP	1	Water	NA	NA	20.8	7.91	04/26/2024	13:56
8	CCV2	1	Water	NA	NA	20.5	12.02	04/26/2024	14:00

UP 130520

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p2283 WorkList ID : 179795 Department : Wet-Chemistry Date : 04-26-2024 13:19:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2283-05 H	CONTAIN-WATER	Water	pH	Cool 4 deg C	PSEG03	I31	04/26/2024	9040C

Date/Time 04/26/24 13:30
Raw Sample Received by: JQ(COC)
Raw Sample Relinquished by: JTCM

Date/Time 04/26/24 16:00
Raw Sample Received by: JTCM
Raw Sample Relinquished by: JQ(COC)