

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130522

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	04/26/2024	13:30
2	P2283-05		1	12	100.00	>212.0	>212.0	04/26/2024	14:00
3	P2283-05DUP		1	13	100.00	>212.0	>212.0	04/26/2024	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-1305-22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/26/24 flashpoint w

WorkList ID : 179797

Department : Wet-Chemistry

Date : 04-26-2024 13:00:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2283-05	CONTAIN-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	I31	04/26/2024	1010B

Date/Time 04/26/2024 13:22:00
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: J.A. Celis

Date/Time 04/26/2024 15:15
Raw Sample Received by: J.P. Celis
Raw Sample Relinquished by: N.P.W.C