

LB180123


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4096E



A= 0.0000e+000

B= 2.4357e-004 slope

C= -4.2543e-002 y-intercept

Rho= 0.9998933

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.00	0.000	0.003	0.003	189	0.000	189					1
0.05	0.050					442	0				3
0.20	0.200	0.205	0.005	1018	0.0 %	1018					1
2.50	2.500	2.471	-0.029	10319	0.0 %	10319					1
5.00	5.000	4.975	-0.025	20600	0.0 %	20600					1
7.50	7.500	7.609	0.109	31413	0.0 %	31413					1
10.0	10.000	9.936	-0.064	40968	0.0 %	40968					1

LB130523

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	189	-		0 SFAM01.1	PPB	4/26/2024 15:58	Std
	0.2 S01	1018	-		0.2 SFAM01.1	PPB	4/26/2024 16:00	Std
	2.5 S02	10319	-		2.5 SFAM01.1	PPB	4/26/2024 16:03	Std
	5 S03	20600	-		5 SFAM01.1	PPB	4/26/2024 16:05	Std
	7.5 S04	31413	-		7.5 SFAM01.1	PPB	4/26/2024 16:07	Std
	10 S05	40968	-		10 SFAM01.1	PPB	4/26/2024 16:09	Std
ICV017	ICV017	16611	4.0035 -		SFAM01.1	PPB	4/26/2024 16:14	SMPL
ICB017	ICB017	-67	-0.0589 -		SFAM01.1	PPB	4/26/2024 16:16	SMPL
CCV084	CCV084	22111	5.3431 -		SFAM01.1	PPB	4/26/2024 16:19	SMPL
CCB084	CCB084	-126	-0.0732 -		SFAM01.1	PPB	4/26/2024 16:21	SMPL
PB160571BL	PBW571	194	0.0047 -		SFAM01.1	PPB	4/26/2024 16:23	SMPL
P2106-01	MCOAF6	178	0.0008 -		SFAM01.1	PPB	4/26/2024 16:25	SMPL
P2106-02	MCOAF7	149	-0.0063 -		SFAM01.1	PPB	4/26/2024 16:28	SMPL
P2106-03	MCOAF7D	203	0.0069 -		SFAM01.1	PPB	4/26/2024 16:30	SMPL
P2106-04	MCOAF7S	5114	1.2031 -		SFAM01.1	PPB	4/26/2024 16:35	SMPL
P2106-05	MCOAF8	140	-0.0084 -		SFAM01.1	PPB	4/26/2024 16:37	SMPL
P2106-06	MCOAF9	186	0.0028 -		SFAM01.1	PPB	4/26/2024 16:40	SMPL
P2106-07	MCOAG0	186	0.0028 -		SFAM01.1	PPB	4/26/2024 16:42	SMPL
P2106-08	MCOAG1	213	0.0093 -		SFAM01.1	PPB	4/26/2024 16:44	SMPL
P2106-09	MCOAG2	190	0.0037 -		SFAM01.1	PPB	4/26/2024 16:46	SMPL
P2106-10	MCOAK5	163	-0.0028 -		SFAM01.1	PPB	4/26/2024 16:49	SMPL
P2106-11	MCOAL3	141	-0.0082 -		SFAM01.1	PPB	4/26/2024 16:51	SMPL
P2106-12	MCOAL4	149	-0.0063 -		SFAM01.1	PPB	4/26/2024 16:53	SMPL
P2106-13	MCOAL9	175	0.0001 -		SFAM01.1	PPB	4/26/2024 16:55	SMPL
P2281-01	E08P3	165	-0.0024 -		SFAM01.1	PPB	4/26/2024 16:58	SMPL
CCV085	CCV085	21335	5.1541 -		SFAM01.1	PPB	4/26/2024 17:00	SMPL
CCB085	CCB085	-78	-0.0615 -		SFAM01.1	PPB	4/26/2024 17:02	SMPL