

66130525

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

4/26/2024 15:27

Test: CNEPA-NEW

Sample Id	Result	Dil. 1 +	Response	Errors
ICV001 ICV001	92.389	0.0	0.085	
ICB001 ICB001	1.400	0.0	0.002	
CCV001 CCV001	243.575	0.0	0.222	
CCB001 CCB001	1.351	0.0	0.002	
PB160560BL PBW560	1.538	0.0	0.002	
P2281-02 E08P4	1.190	0.0	0.002	
CCV002 CCV002	244.205	0.0	0.222	
CCB002 CCB002	1.015	0.0	0.002	

N	8
Mean	73.333
SD	109.8619
CV%	149.81

04/26/2024
RM

Aquakem v. 7.2AQ1

Results from time period:

Fri Apr 26 14:07:28 2024

Fri Apr 26 15:22:38 2024

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
S0.0	A	CNEPA-NEW	P	1.1411	µg/l	4/26/2024 14:07:28
S5.0	A	CNEPA-NEW	P	4.5098	µg/l	4/26/2024 14:07:29
S10.0	A	CNEPA-NEW	P	8.363	µg/l	4/26/2024 14:07:30
S100.0	A	CNEPA-NEW	P	93.9437	µg/l	4/26/2024 14:07:31
S250.0	A	CNEPA-NEW	P	261.587	µg/l	4/26/2024 14:07:32
S500.0	A	CNEPA-NEW	P	495.4554	µg/l	4/26/2024 14:07:33
ICV001 ICV001	S	CNEPA-NEW	P	92.389	µg/l	4/26/2024 14:49:30
ICB001 ICB001	S	CNEPA-NEW	P	1.3997	µg/l	4/26/2024 14:49:31
CCV001 CCV001	S	CNEPA-NEW	P	243.5751	µg/l	4/26/2024 14:49:33
CCB001 CCB001	S	CNEPA-NEW	P	1.3511	µg/l	4/26/2024 14:49:36
PB160560BL PBW560	S	CNEPA-NEW	P	1.5382	µg/l	4/26/2024 14:49:37
P2281-02 E08P4	S	CNEPA-NEW	P	1.1902	µg/l	4/26/2024 14:49:39
CCV002 CCV002	S	CNEPA-NEW	P	244.2047	µg/l	4/26/2024 14:57:06
CCB002 CCB002	S	CNEPA-NEW	P	1.0151	µg/l	4/26/2024 14:57:09

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

4/26/2024 14:09

Test CNEPA-NEW

Accepted

4/26/2024 14:09

Factor slope

1103 0.000906

Bias y-intercept


0.001

04/29/2024
RM

Coeff. of det.

0.999012

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCNS0.0	0.002	1.1411	0.0000	-
2	5.0PPBCNS5.0	0.005	4.5098	5.0000	-4.8
3	10PPBCNS10.0	0.008	8.3630	10.0000	-6.4
4	100PPBCNS100.0	0.086	93.9437	100.0000	-6.1
5	250PPBCNS250.0	0.238	261.5870	250.0000	4.6
6	500PPBCNS500.0	0.450	495.4554	500.0000	-0.9

04/26/2024
RM