

LB130532


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46494



A= 0.0000e+000

B= 2.1311e-004 *slope*C= -6.0427e-002 *-10 intercept*

Rho= 0.9995849

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>6/8/11</i>
0.00	0.000	-0.018	-0.018	197	0.000	197					<i>=</i>
0.05	0.050					442	0				<i>7</i>
0.20	0.200	0.213	0.013	1283	0.0 %	1283					<i>-4</i>
2.50	2.500	2.409	-0.091	11586	0.0 %	11586					<i>2</i>
5.00	5.000	5.075	0.075	24098	0.0 %	24098					<i>2</i>
7.50	7.500	7.674	0.174	36292	0.0 %	36292					<i>2</i>
10.0	10.000	9.848	-0.152	46494	0.0 %	46494					<i>-2</i>

LB130532

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	197	-		0 SFAM01.1	PPB	4/29/2024 10:00	Std
	0.2 S01	1283	-		0.2 SFAM01.1	PPB	4/29/2024 10:02	Std
	2.5 S02	11586	-		2.5 SFAM01.1	PPB	4/29/2024 10:05	Std
	5 S03	24098	-		5 SFAM01.1	PPB	4/29/2024 10:07	Std
	7.5 S04	36292	-		7.5 SFAM01.1	PPB	4/29/2024 10:09	Std
	10 S05	46494	-		10 SFAM01.1	PPB	4/29/2024 10:12	Std
ICV018	ICV018	17301	3.6266 -		SFAM01.1	PPB	4/29/2024 10:14	SMPL
ICB018	ICB018	-123	-0.0866 -		SFAM01.1	PPB	4/29/2024 10:17	SMPL
CCV086	CCV086	24355	5.1299 -		SFAM01.1	PPB	4/29/2024 10:19	SMPL
CCB086	CCB086	-181	-0.099 -		SFAM01.1	PPB	4/29/2024 10:21	SMPL
PB160470BL	PBW470	271	-0.0027 -		SFAM01.1	PPB	4/29/2024 10:24	SMPL
PB160243TB	LEB243	192	-0.0195 -		SFAM01.1	PPB	4/29/2024 10:26	SMPL
P2065-01	E1CY1	215	-0.0146 -		SFAM01.1	PPB	4/29/2024 10:28	SMPL
P2065-02	E1CY1D	212	-0.0152 -		SFAM01.1	PPB	4/29/2024 10:30	SMPL
P2065-03	E1CY1S	5280	1.0648 -		SFAM01.1	PPB	4/29/2024 10:33	SMPL
P2065-04	E1CX7	212	-0.0152 -		SFAM01.1	PPB	4/29/2024 10:35	SMPL
CCV087	CCV087	24126	5.0811 -		SFAM01.1	PPB	4/29/2024 10:37	SMPL
CCB087	CCB087	-116	-0.0851 -		SFAM01.1	PPB	4/29/2024 10:40	SMPL