


LB130536

245.1

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/vb
0.0	0.000	-0.000	-0.000	126	0.000	126					1
0.05	0.050	0.058	0.008	361	0.0 %	361	0				2
0.2	0.200	0.205	0.005	954	0.0 %	954					2
2.5	2.500	2.444	-0.056	9952	0.0 %	9952					2
5.0	5.000	4.990	-0.010	20191	0.0 %	20191					0
7.5	7.500	7.651	0.151	30889	0.0 %	30889					2
10.0	10.000	9.902	-0.098	39935	0.0 %	39935					1

LB130536 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Apr 2024 13:01:06

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	126	-	0.0000	245.1	PPB	29 Apr 2024 13:39:04	S	Std
0.05 - 1	50.05	361	-	0.0500	245.1	PPB	29 Apr 2024 13:41:21	S	Std
0.2 - 1	50.2	954	-	0.2000	245.1	PPB	29 Apr 2024 13:43:38	S	Std
2.5 - 1	52.5	9952	-	2.5000	245.1	PPB	29 Apr 2024 13:45:56	S	Std
5.0 - 1	55	20191	-	5.0000	245.1	PPB	29 Apr 2024 13:48:12	S	Std
7.5 - 1	57.5	30889	-	7.5000	245.1	PPB	29 Apr 2024 13:50:29	S	Std
10.0 - 1	510	39935	-	10.0000	245.1	PPB	29 Apr 2024 13:58:06	S	Std
ICV19 - 1	ICV19	16120	3.9778	-245.1	PPB	PPB	29 Apr 2024 14:00:59	U	SMPL
ICB19 - 1	ICB19	1	-0.0316	-245.1	PPB	PPB	29 Apr 2024 14:03:15	U	SMPL
CCV88 - 1	CCV88	20653	5.1054	-245.1	PPB	PPB	29 Apr 2024 14:05:33	U	SMPL
CCB88 - 1	CCB88	-19	-0.0366	-245.1	PPB	PPB	29 Apr 2024 14:07:48	U	SMPL
CRA - 1	CRA	290	0.0403	-245.1	PPB	PPB	29 Apr 2024 14:10:07	U	SMPL
HighStd - 1	HighStd	40378	10.0117	-245.1	PPB	PPB	29 Apr 2024 14:12:22	U	SMPL
ChkStd - 1	ChkStd	29427	7.2878	-245.1	PPB	PPB	29 Apr 2024 14:14:37	U	SMPL
PB160601BL - 1	PBW	125	-0.0007	-245.1	PPB	PPB	29 Apr 2024 14:19:35	U	SMPL
PB160601BS - 1	LCSW	16273	4.0159	-245.1	PPB	PPB	29 Apr 2024 14:24:35	U	SMPL
P2226-01 - 1	GROUND-WATER	133607	33.2013	-245.1	PPB	PPB	29 Apr 2024 14:26:51	U	SMPL
P2226-01DUP - 1	GROUND-WATERDUP	124051	30.8244	-245.1	PPB	PPB	29 Apr 2024 14:29:10	U	SMPL
P2226-01MS - 1	GROUND-WATERMS	134145	33.3352	-245.1	PPB	PPB	29 Apr 2024 14:31:35	U	SMPL
P2226-01MSD - 1	GROUND-WATERMSD	141481	35.1599	-245.1	PPB	PPB	29 Apr 2024 14:33:59	U	SMPL
CCV89 - 1	CCV89	20651	5.1049	-245.1	PPB	PPB	29 Apr 2024 14:36:21	U	SMPL
CCB89 - 1	CCB89	40	-0.0219	-245.1	PPB	PPB	29 Apr 2024 14:38:44	U	SMPL
P2226-01LX5 - 1		27074	6.7025	-245.1	PPB	PPB	29 Apr 2024 14:41:01	U	SMPL
P2226-01A - 1		139799	34.7415	-245.1	PPB	PPB	29 Apr 2024 14:43:17	U	SMPL
P2226-01DLX5 - 1		25487	6.3078	-245.1	PPB	PPB	29 Apr 2024 14:45:34	U	SMPL
P2226-01DUPDLX5 - 1		28480	7.0522	-245.1	PPB	PPB	29 Apr 2024 14:47:53	U	SMPL
P2226-01MSDLX5 - 1		29725	7.3619	-245.1	PPB	PPB	29 Apr 2024 14:50:11	U	SMPL
P2226-01MSDDLX5 - 1		31788	7.8751	-245.1	PPB	PPB	29 Apr 2024 14:52:29	U	SMPL
P2226-01LDLX25 - 1		4842	1.1726	-245.1	PPB	PPB	29 Apr 2024 14:54:48	U	SMPL
P2226-01ADLX5 - 1		33085	8.1977	-245.1	PPB	PPB	29 Apr 2024 14:57:07	U	SMPL
CCV90 - 1	CCV90	21667	5.3576	-245.1	PPB	PPB	29 Apr 2024 14:59:24	U	SMPL
CCB90 - 1	CCB90	76	-0.0129	-245.1	PPB	PPB	29 Apr 2024 15:01:44	U	SMPL