

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/30/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 04/29/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 04/30/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB130541

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2297-01	MDCOT6	1	1.15	8.43	9.58	8.15	83.0	
P2297-02	MDCOT7	2	1.15	8.80	9.95	8.16	79.7	
P2297-03	MDCOT8	3	1.19	8.54	9.73	7.93	78.9	
P2297-04	MDCOT9	4	1.18	8.62	9.8	7.92	78.2	
P2297-05	MDCOU0	5	1.16	8.54	9.7	7.92	79.2	
P2297-06	MDCOU1	6	1.14	8.36	9.5	7.77	79.3	
P2297-07	MDCOU2	7	1.18	8.51	9.69	7.98	79.9	
P2297-08	MDCOU3	8	1.19	8.66	9.85	7.81	76.4	
P2297-09	MDCOU4	9	1.12	8.82	9.94	8.03	78.3	
P2297-10	MDCOU5	10	1.12	8.70	9.82	7.72	75.9	
P2297-11	MDCOU6	11	1.15	8.40	9.55	7.61	76.9	
P2297-12	MDCOU7	12	1.12	8.75	9.87	7.97	78.3	
P2297-13	MDCOU8	13	1.11	8.62	9.73	7.93	79.1	
P2297-14	MDCOV1	14	1.18	8.79	9.97	7.64	73.5	
P2297-15	MDCOV2	15	1.15	8.41	9.56	7.75	78.5	
P2297-16	MDCOV4	16	1.17	8.69	9.86	7.81	76.4	
P2297-17	MDCOV5	17	1.18	8.70	9.88	7.62	74.0	
P2297-18	MDCOV6	18	1.19	8.77	9.96	8.04	78.1	
P2297-19	MDCOV7	19	1.18	8.48	9.66	7.82	78.3	
P2297-20	MDCOW8	20	1.18	8.59	9.77	8.23	82.1	
P2297-21	MDCOW8D	21	1.18	8.59	9.77	8.23	82.1	
P2297-22	MDCOW8S	22	1.18	8.59	9.77	8.23	82.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$