

7470A

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.4378e-004

C= 1.4893e-002

Rho= 0.9992212

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	0.054	0.054	160	0.000	160	0				0/01)
0.2	0.200	0.230	0.030	881	0.0 %	881					151
2.5	2.500	2.528	0.028	10308	0.0 %	10308					36-151
5.0	5.000	4.716	-0.284	19286	0.0 %	19286					0
7.5	7.500	7.703	0.203	31538	0.0 %	31538					
10.0	10.000	9.969	-0.031	40834	0.0 %	40834					

LB130565 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Apr 2024 13:43:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	160	-	0.0000	7470A	PPB	30 Apr 2024 13:43:30	S	Std
0.2 - 1	SP.2	881	-	0.2000	7470A	PPB	30 Apr 2024 13:45:47	S	Std
2.5 - 1	SP.2	10308	-	2.5000	7470A	PPB	30 Apr 2024 13:48:04	S	Std
5.0 - 1	SP.5	19286	-	5.0000	7470A	PPB	30 Apr 2024 13:50:21	S	Std
7.5 - 1	SP.75	31538	-	7.5000	7470A	PPB	30 Apr 2024 13:52:38	S	Std
10.0 - 1	SP.10	40834	-	10.0000	7470A	PPB	30 Apr 2024 13:54:56	S	Std
ICV21 - 1	ICV21	16699	4.0857	-	7470A	PPB	30 Apr 2024 13:57:54	U	SMPL
ICB21 - 1	ICB21	-175	-0.0278	-	7470A	PPB	30 Apr 2024 14:00:09	U	SMPL
CCV94 - 1	CCV94	23086	5.6427	-	7470A	PPB	30 Apr 2024 14:02:26	U	SMPL
CCB94 - 1	CCB94	0	0.0149	-	7470A	PPB	30 Apr 2024 14:04:40	U	SMPL
CCV94 - 1	CCV94	20481	5.0077	-	7470A	PPB	30 Apr 2024 14:05:50	U	SMPL
CCB94 - 1	CCB94	-174	-0.0275	-	7470A	PPB	30 Apr 2024 14:08:05	U	SMPL
CRA - 1	CRA	858	0.2241	-	7470A	PPB	30 Apr 2024 14:10:22	U	SMPL
HighStd - 1	HighStd	43369	10.5872	-	7470A	PPB	30 Apr 2024 14:12:37	U	SMPL
ChkStd - 1	ChkStd	31960	7.8060	-	7470A	PPB	30 Apr 2024 14:14:53	U	SMPL
PB160617BL - 1	PBW	-372	-0.0758	-	7470A	PPB	30 Apr 2024 14:17:11	U	SMPL
PB160617BS - 1	LCSW	14977	3.6659	-	7470A	PPB	30 Apr 2024 14:19:30	U	SMPL
P2270-02 - 1	72-11936	-136	-0.0183	-	7470A	PPB	30 Apr 2024 14:21:47	U	SMPL
P2270-04 - 1	VNJ-240	108	0.0412	-	7470A	PPB	30 Apr 2024 14:24:05	U	SMPL
P2271-02 - 1	RT-3860	119	0.0439	-	7470A	PPB	30 Apr 2024 14:26:22	U	SMPL
P2283-02 - 1	EXCAV-SOIL-1	171	0.0566	-	7470A	PPB	30 Apr 2024 14:28:39	U	SMPL
P2283-04 - 1	EXCAV-SOIL-2	184	0.0597	-	7470A	PPB	30 Apr 2024 14:30:55	U	SMPL
CCV95 - 1	CCV95	20708	5.0630	-	7470A	PPB	30 Apr 2024 14:33:10	U	SMPL
CCB95 - 1	CCB95	-141	-0.0195	-	7470A	PPB	30 Apr 2024 14:35:25	U	SMPL
P2283-04DUP - 1	EXCAV-SOIL-2DUP	95	0.0381	-	7470A	PPB	30 Apr 2024 14:37:43	U	SMPL
P2283-04MS - 1	EXCAV-SOIL-2MS	15886	3.8875	-	7470A	PPB	30 Apr 2024 14:39:58	U	SMPL
P2283-04MSD - 1	EXCAV-SOIL-2MSD	14516	3.5535	-	7470A	PPB	30 Apr 2024 14:42:13	U	SMPL
PB160626BL - 1	PBW	-161	-0.0244	-	7470A	PPB	30 Apr 2024 14:44:31	U	SMPL
PB160626BS - 1	LCSW	15373	3.7625	-	7470A	PPB	30 Apr 2024 14:46:49	U	SMPL
P2230-01 - 1	SOIL	-117	-0.0136	-	7470A	PPB	30 Apr 2024 14:49:05	U	SMPL
P2261-02 - 1	RO-RBR200059	73	0.0327	-	7470A	PPB	30 Apr 2024 14:51:24	U	SMPL
P2261-02DUP - 1	RO-RBR200059DUP	157	0.0532	-	7470A	PPB	30 Apr 2024 14:53:41	U	SMPL
P2261-02MS - 1	RO-RBR200059MS	15022	3.6769	-	7470A	PPB	30 Apr 2024 14:55:58	U	SMPL
P2261-02MSD - 1	RO-RBR200059MSD	13760	3.3693	-	7470A	PPB	30 Apr 2024 14:58:13	U	SMPL
CCV96 - 1	CCV96	21745	5.3158	-	7470A	PPB	30 Apr 2024 15:00:30	U	SMPL
CCB96 - 1	CCB96	-234	-0.0422	-	7470A	PPB	30 Apr 2024 15:02:48	U	SMPL
PB160627BL - 1	PBW	46	0.0261	-	7470A	PPB	30 Apr 2024 15:05:06	U	SMPL
PB160627BS - 1	LCSW	15193	3.7186	-	7470A	PPB	30 Apr 2024 15:07:22	U	SMPL
P2257-13 - 1	4524	-17	0.0107	-	7470A	PPB	30 Apr 2024 15:09:37	U	SMPL
P2260-04 - 1	60240	3146	0.7818	-	7470A	PPB	30 Apr 2024 15:11:55	U	SMPL
P2261-04 - 1	A069	6304	1.5517	-	7470A	PPB	30 Apr 2024 15:19:22	U	SMPL
P2261-04DUP - 1	A069DUP	5811	1.4315	-	7470A	PPB	30 Apr 2024 15:21:38	U	SMPL
P2261-04MS - 1	A069MS	18941	4.6323	-	7470A	PPB	30 Apr 2024 15:23:56	U	SMPL
P2261-04MSD - 1	A069MSD	18700	4.5735	-	7470A	PPB	30 Apr 2024 15:26:14	U	SMPL
P2266-01 - 1	B-101-GW01	152	0.0519	-	7470A	PPB	30 Apr 2024 15:28:33	U	SMPL
P2266-02 - 1	EB-042324	133	0.0473	-	7470A	PPB	30 Apr 2024 15:30:50	U	SMPL
CCV97 - 1	CCV97	18403	4.5011	-	7470A	PPB	30 Apr 2024 16:10:45	U	SMPL
CCB97 - 1	CCB97	-23	0.0093	-	7470A	PPB	30 Apr 2024 16:13:01	U	SMPL
P2283-05 - 1	CONTAIN-WATER	116	0.0432	-	7470A	PPB	30 Apr 2024 16:15:18	U	SMPL
PB160479TB - 1		125	0.0454	-	7470A	PPB	30 Apr 2024 16:17:34	U	SMPL
PB160553TB - 1		87	0.0361	-	7470A	PPB	30 Apr 2024 16:19:49	U	SMPL
P2283-04LX5 - 1		75	0.0332	-	7470A	PPB	30 Apr 2024 16:22:05	U	SMPL
P2283-04A - 1		16935	4.1432	-	7470A	PPB	30 Apr 2024 16:24:20	U	SMPL
P2261-02LX5 - 1		-35	0.0064	-	7470A	PPB	30 Apr 2024 16:26:36	U	SMPL
P2261-02A - 1		17306	4.2337	-	7470A	PPB	30 Apr 2024 16:28:54	U	SMPL
P2261-04LX5 - 1		1285	0.3281	-	7470A	PPB	30 Apr 2024 16:34:25	U	SMPL
P2261-04A - 1		24426	5.9694	-	7470A	PPB	30 Apr 2024 16:36:41	U	SMPL
CCV98 - 1	CCV98	18989	4.6440	-	7470A	PPB	30 Apr 2024 16:46:44	U	SMPL
CCB98 - 1	CCB98	-163	-0.0248	-	7470A	PPB	30 Apr 2024 16:48:59	U	SMPL