

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130571

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/30/2024	16:35
2	CAL2	1	Water	NA	NA	20.3	7.00	04/30/2024	16:36
3	CAL3	1	Water	NA	NA	20.3	10.02	04/30/2024	16:37
4	ICV	1	Water	NA	NA	20.4	7.02	04/30/2024	16:39
5	CCV1	1	Water	NA	NA	20.3	2.01	04/30/2024	16:41
6	P2323-03	1	Solid	20.02	20	22.7	8.35	04/30/2024	16:45
7	P2323-03DUP	1	Solid	20.03	20	22.8	8.36	04/30/2024	16:46
8	P2323-06	1	Solid	20.02	20	23.0	8.00	04/30/2024	16:50
9	P2323-09	1	Solid	20.03	20	22.9	7.52	04/30/2024	16:55
10	CCV2	1	Water	NA	NA	20.4	12.02	04/30/2024	17:00

WORKLIST(Hardcopy Internal Chain)

VB 130571

WorkList Name : corrosivity p2323 WorkList ID : 179877 Department : Wet-Chemistry Date : 04-30-2024 16:22:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2323-03	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	I31	04/30/2024	9045D
P2323-06	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	I31	04/30/2024	9045D
P2323-09	TP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	I31	04/30/2024	9045D

Date/Time 04/30/24 16:30
Raw Sample Received by: JDCDC1
Raw Sample Relinquished by: RJ LATA VAB

Date/Time 04/30/24 18:00
Raw Sample Received by: RJ LATA VAB
Raw Sample Relinquished by: JDCDC1