

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130572

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/30/2024	16:35
2	CAL2	1	Water	NA	NA	20.3	7.00	04/30/2024	16:36
3	CAL3	1	Water	NA	NA	20.3	10.02	04/30/2024	16:37
4	ICV	1	Water	NA	NA	20.4	7.02	04/30/2024	16:39
5	CCV1	1	Water	NA	NA	20.4	2.02	04/30/2024	17:05
6	P2308-01	1	Solid	20.02	20	23.7	7.99	04/30/2024	17:10
7	P2308-01DUP	1	Solid	20.03	20	25.6	8.00	04/30/2024	17:11
8	CCV2	1	Water	NA	NA	20.3	12.02	04/30/2024	17:15

WORKLIST(Hardcopy Internal Chain)

130572

WorkList Name : ph p2308

WorkList ID : 179876

Department : Wet-Chemistry

Date : 04-30-2024 16:21:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2308-01	346	Solid	pH	Cool 4 deg C	PSEG03	I41	04/29/2024	9045D

Date/Time 04/30/24 16:27
Raw Sample Received by: AD WOC
Raw Sample Relinquished by: (ccom)

Date/Time 04/30/24 18:10
Raw Sample Received by: (ccom)
Raw Sample Relinquished by: (ccom)