


LB 130579

SFAM01.1

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/O
0.00	0.000	0.022	0.022	135	0.000	135					0/0
0.05	0.050					442	0				=
0.20	0.200	0.234	0.034	997	0.0 %	997					17
2.50	2.500	2.446	-0.054	9990	0.0 %	9990					-2
5.00	5.000	5.080	0.080	20705	0.0 %	20705					2
7.50	7.500	7.304	-0.196	29748	0.0 %	29748					-3
10.0	10.000	10.114	0.114	41177	0.0 %	41177					1

LB130579

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	135	-		0 SFAM01.1	PPB	5/1/2024 11:26	Std
	0.2 S01	997	-		0.2 SFAM01.1	PPB	5/1/2024 11:29	Std
	2.5 S02	9990	-		2.5 SFAM01.1	PPB	5/1/2024 11:31	Std
	5 S03	20705	-		5 SFAM01.1	PPB	5/1/2024 11:33	Std
	7.5 S04	29748	-		7.5 SFAM01.1	PPB	5/1/2024 11:35	Std
	10 S05	41177	-		10 SFAM01.1	PPB	5/1/2024 11:38	Std
ICV023	ICV023	16006	3.9248 -		SFAM01.1	PPB	5/1/2024 11:41	SMPL
ICB023	ICB023	127	0.0204 -		SFAM01.1	PPB	5/1/2024 11:45	SMPL
CCV001	CCV001	21651	5.3128 -		SFAM01.1	PPB	5/1/2024 11:48	SMPL
CCB001	CCB001	-317	-0.0887 -		SFAM01.1	PPB	5/1/2024 11:50	SMPL
PB160654BL	PBW654	449	0.0996 -		SFAM01.1	PPB	5/1/2024 11:52	SMPL
PB160521TB	LEB521	399	0.0873 -		SFAM01.1	PPB	5/1/2024 11:57	SMPL
P2206-01	E1CX1	719	0.166 -		SFAM01.1	PPB	5/1/2024 11:59	SMPL
P2206-02	E1CW8	747	0.1729 -		SFAM01.1	PPB	5/1/2024 12:01	SMPL
P2206-03	E1CX4	788	0.183 -		SFAM01.1	PPB	5/1/2024 12:08	SMPL
P2206-04	E1CX4D	788	0.183 -		SFAM01.1	PPB	5/1/2024 12:11	SMPL
P2206-05	E1CX4S	3911	0.9509 -		SFAM01.1	PPB	5/1/2024 12:13	SMPL
PB160655BL	PBW655	243	0.049 -		SFAM01.1	PPB	5/1/2024 12:15	SMPL
PB160590TB	LEB590	411	0.0903 -		SFAM01.1	PPB	5/1/2024 12:18	SMPL
P2237-01	E0808	895	0.2093 -		SFAM01.1	PPB	5/1/2024 12:20	SMPL
P2237-02	E0813	942	0.2208 -		SFAM01.1	PPB	5/1/2024 12:22	SMPL
P2237-03	E0813D	933	0.2186 -		SFAM01.1	PPB	5/1/2024 12:24	SMPL
P2237-04	E0813S	3848	0.9354 -		SFAM01.1	PPB	5/1/2024 12:29	SMPL
CCV002	CCV002	21148	5.1891 -		SFAM01.1	PPB	5/1/2024 12:31	SMPL
CCB002	CCB002	168	0.0305 -		SFAM01.1	PPB	5/1/2024 12:34	SMPL