

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130580

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/01/2024	12:35
2	CAL2	1	Water	NA	NA	20.3	7.00	05/01/2024	12:36
3	CAL3	1	Water	NA	NA	20.3	10.02	05/01/2024	12:38
4	ICV	1	Water	NA	NA	20.3	7.00	05/01/2024	12:40
5	CCV1	1	Water	NA	NA	20.2	2.01	05/01/2024	12:44
6	P2330-01	1	Solid	20.02	20	22.1	6.13	05/01/2024	12:45
7	P2330-01DUP	1	Solid	20.03	20	22.2	6.14	05/01/2024	12:46
8	P2330-03	1	Solid	20.04	20	22.2	6.61	05/01/2024	12:50
9	P2330-05	1	Solid	20.02	20	22.2	7.73	05/01/2024	12:55
10	P2330-07	1	Solid	20.03	20	22.1	7.80	05/01/2024	12:59
11	P2330-09	1	Solid	20.02	20	22.2	8.93	05/01/2024	13:00
12	P2330-11	1	Solid	20.03	20	22.2	7.96	05/01/2024	13:10
13	P2330-13	1	Solid	20.04	20	22.2	7.99	05/01/2024	13:15
14	P2330-15	1	Solid	20.02	20	22.1	7.85	05/01/2024	13:20
15	P2330-17	1	Solid	20.03	20	21.9	7.71	05/01/2024	13:22
16	CCV2	1	Water	NA	NA	20.3	12.02	05/01/2024	13:25
17	P2330-19	1	Solid	20.04	20	21.8	7.28	05/01/2024	13:30
18	P2330-21	1	Solid	20.02	20	22.2	7.34	05/01/2024	13:37
19	CCV3	1	Water	NA	NA	20.3	2.02	05/01/2024	13:40

WORKLIST(Hardcopy Internal Chain)

MS130580

WorkList Name : corrosion p2330 WorkList ID : 179906 Department : Wet-Chemistry Date : 05-01-2024 12:21:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2330-01	B-1(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/29/2024	9045D
P2330-03	D-1(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/29/2024	9045D
P2330-05	F-1(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/29/2024	9045D
P2330-07	H-1(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/29/2024	9045D
P2330-09	C-2(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/30/2024	9045D
P2330-11	E-2(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/30/2024	9045D
P2330-13	G-2(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/30/2024	9045D
P2330-15	C-3(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/29/2024	9045D
P2330-17	F-3(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/30/2024	9045D
P2330-19	E-4(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/30/2024	9045D
P2330-21	H-4(0.5-1.5)	Solid	Corrosivity	Cool 4 deg C	FIRS02	I31	04/29/2024	9045D

Date/Time 05/01/24 12:30
 Raw Sample Received by: JH GORI
 Raw Sample Relinquished by: RM SON

Date/Time 05/01/24 16:00
 Raw Sample Received by: RM SON
 Raw Sample Relinquished by: WP WEE