


LB130581

SFAM01.1

INSTRUMENT IN: CV1

 μ Abs.:

44107



Linear

A= 0.0000e+000

B= 2.2796e-004 slope

C= 9.1113e-003 y-intercept

Rho= 0.9995025

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.048	0.048	172	0.000	172					=
0.05	0.050					442	0				22
0.20	0.200	0.243	0.043	1026	0.0 %	1026					-7
2.50	2.500	2.335	-0.165	10202	0.0 %	10202					3
5.00	5.000	5.156	0.156	22579	0.0 %	22579					-2
7.50	7.500	7.354	-0.146	32220	0.0 %	32220					1
10.0	10.000	10.064	0.064	44107	0.0 %	44107					

LB130581

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	172	-		0 SFAM01.1	PPB	5/1/2024 13:56	Std
	0.2 S01	1026	-		0.2 SFAM01.1	PPB	5/1/2024 13:59	Std
	2.5 S02	10202	-		2.5 SFAM01.1	PPB	5/1/2024 14:01	Std
	5 S03	22579	-		5 SFAM01.1	PPB	5/1/2024 14:03	Std
	7.5 S04	32220	-		7.5 SFAM01.1	PPB	5/1/2024 14:05	Std
	10 S05	44107	-		10 SFAM01.1	PPB	5/1/2024 14:08	Std
ICV024	ICV024	17162	3.9214 -		SFAM01.1	PPB	5/1/2024 14:11	SMPL
ICB024	ICB024	-315	-0.0627 -		SFAM01.1	PPB	5/1/2024 14:13	SMPL
CCV003	CCV003	22176	5.0643 -		SFAM01.1	PPB	5/1/2024 14:15	SMPL
CCB003	CCB003	-252	-0.0483 -		SFAM01.1	PPB	5/1/2024 14:18	SMPL
PB160659BL	PBW659	237	0.0631 -		SFAM01.1	PPB	5/1/2024 14:20	SMPL
P2107-01	MCOAG6	494	0.1217 -		SFAM01.1	PPB	5/1/2024 14:22	SMPL
P2107-02	MCOAG7	864	0.2061 -		SFAM01.1	PPB	5/1/2024 14:29	SMPL
P2107-03	MCOAG7D	846	0.202 -		SFAM01.1	PPB	5/1/2024 14:31	SMPL
P2107-04	MCOAG7S	4010	0.9232 -		SFAM01.1	PPB	5/1/2024 14:34	SMPL
P2107-05	MCOAG8	721	0.1735 -		SFAM01.1	PPB	5/1/2024 14:36	SMPL
P2107-06	MCOAG9	363	0.0919 -		SFAM01.1	PPB	5/1/2024 14:38	SMPL
P2107-07	MCOAH0	636	0.1541 -		SFAM01.1	PPB	5/1/2024 14:40	SMPL
P2107-08	MCOAH1	400	0.1003 -		SFAM01.1	PPB	5/1/2024 14:43	SMPL
P2107-09	MCOAH2	2333	0.5409 -		SFAM01.1	PPB	5/1/2024 14:45	SMPL
P2107-10	MCOAK8	208	0.0565 -		SFAM01.1	PPB	5/1/2024 14:47	SMPL
P2133-01	YCTK2	1336	0.3137 -		SFAM01.1	PPB	5/1/2024 14:57	SMPL
P2133-02	YCTK2D	1358	0.3187 -		SFAM01.1	PPB	5/1/2024 15:00	SMPL
P2133-03	YCTK2S	8387	1.921 -		SFAM01.1	PPB	5/1/2024 15:02	SMPL
P2133-04	YCTL4	1350	0.3169 -		SFAM01.1	PPB	5/1/2024 15:04	SMPL
P2133-05	YCTNO	419	0.1046 -		SFAM01.1	PPB	5/1/2024 15:06	SMPL
CCV004	CCV004	21637	4.9415 -		SFAM01.1	PPB	5/1/2024 15:09	SMPL
CCB004	CCB004	-180	-0.0319 -		SFAM01.1	PPB	5/1/2024 15:11	SMPL