

LB130584


SFAM01.1

INSTRUMENT ID : CV1

Linear

 $\mu$  Abs.:

48232



A= 0.0000e+000

B= 2.0911e-004 slope

C= -9.8126e-002  $y = mx + c$ 

Rho= 0.9999771

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | W/O 1) |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|--------|
| 0.00   | 0.000  | 0.035 | 0.035  | 636   | 0.000      | 636   |       |       |       |       |        |
| 0.05   | 0.050  |       |        |       |            | 442   | 0     |       |       |       |        |
| 0.20   | 0.200  | 0.186 | -0.014 | 1358  | 0.0 %      | 1358  |       |       |       |       |        |
| 2.50   | 2.500  | 2.465 | -0.035 | 12255 | 0.0 %      | 12255 |       |       |       |       |        |
| 5.00   | 5.000  | 4.997 | -0.003 | 24365 | 0.0 %      | 24365 |       |       |       |       |        |
| 7.50   | 7.500  | 7.530 | 0.030  | 36479 | 0.0 %      | 36479 |       |       |       |       |        |
| 10.0   | 10.000 | 9.988 | -0.012 | 48232 | 0.0 %      | 48232 |       |       |       |       |        |

W/O 1)  
= 7  
- 1  
0  
0  
0

LB130584

INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.    | Std Conc | Method       | Units | Date           | Type |
|------------|-------------|------------|----------|----------|--------------|-------|----------------|------|
|            | 0 S0        | 636        | -        |          | 0 SFAM01.1   | PPB   | 5/1/2024 16:04 | Std  |
|            | 0.2 S01     | 1358       | -        |          | 0.2 SFAM01.1 | PPB   | 5/1/2024 16:06 | Std  |
|            | 2.5 S02     | 12255      | -        |          | 2.5 SFAM01.1 | PPB   | 5/1/2024 16:09 | Std  |
|            | 5 S03       | 24365      | -        |          | 5 SFAM01.1   | PPB   | 5/1/2024 16:11 | Std  |
|            | 7.5 S04     | 36479      | -        |          | 7.5 SFAM01.1 | PPB   | 5/1/2024 16:13 | Std  |
|            | 10 S05      | 48232      | -        |          | 10 SFAM01.1  | PPB   | 5/1/2024 16:16 | Std  |
| ICV025     | ICV025      | 19081      | 3.8919 - |          | SFAM01.1     | PPB   | 5/1/2024 16:18 | SMPL |
| ICB025     | ICB025      | 667        | 0.0414 - |          | SFAM01.1     | PPB   | 5/1/2024 16:21 | SMPL |
| CCV005     | CCV005      | 23820      | 4.8829 - |          | SFAM01.1     | PPB   | 5/1/2024 16:23 | SMPL |
| CCB005     | CCB005      | 671        | 0.0422 - |          | SFAM01.1     | PPB   | 5/1/2024 16:25 | SMPL |
| PB160660BL | PBS660      | 901        | 0.0903 - |          | SFAM01.1     | PPB   | 5/1/2024 16:28 | SMPL |
| P2120-01   | MCOAE6      | 2065       | 0.3337 - |          | SFAM01.1     | PPB   | 5/1/2024 16:30 | SMPL |
| P2120-02   | MCOAE8      | 1329       | 0.1798 - |          | SFAM01.1     | PPB   | 5/1/2024 16:50 | SMPL |
| P2120-03   | MCOAE8D     | 1403       | 0.1953 - |          | SFAM01.1     | PPB   | 5/1/2024 16:52 | SMPL |
| P2120-04   | MCOAE8S     | 10584      | 2.1151 - |          | SFAM01.1     | PPB   | 5/1/2024 16:56 | SMPL |
| P2120-05   | MCOAF1      | 622        | 0.0319 - |          | SFAM01.1     | PPB   | 5/1/2024 16:59 | SMPL |
| P2120-06   | MCOAF2      | 1368       | 0.1879 - |          | SFAM01.1     | PPB   | 5/1/2024 17:01 | SMPL |
| P2120-07   | MCOAL1      | 1143       | 0.1409 - |          | SFAM01.1     | PPB   | 5/1/2024 17:03 | SMPL |
| CCV006     | CCV006      | 24615      | 5.0492 - |          | SFAM01.1     | PPB   | 5/1/2024 17:09 | SMPL |
| CCB006     | CCB006      | 667        | 0.0414 - |          | SFAM01.1     | PPB   | 5/1/2024 17:11 | SMPL |