

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130590

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/02/2024	08:05
2	CAL2	1	Water	NA	NA	20.3	7.00	05/02/2024	08:06
3	CAL3	1	Water	NA	NA	20.3	10.02	05/02/2024	08:10
4	ICV	1	Water	NA	NA	20.2	7.01	05/02/2024	08:11
5	CCV1	1	Water	NA	NA	20.2	2.01	05/02/2024	08:15
6	P2339-04	1	Water	NA	NA	23.0	7.55	05/02/2024	08:22
7	P2339-04DUP	1	Water	NA	NA	23.1	7.56	05/02/2024	08:23
8	P2339-05	1	Water	NA	NA	23.1	6.95	05/02/2024	08:28
9	P2339-06	1	Water	NA	NA	23.1	6.27	05/02/2024	08:33
10	P2339-07	1	Water	NA	NA	23.0	6.21	05/02/2024	08:37
11	P2339-08	1	Water	NA	NA	23.0	6.90	05/02/2024	08:44
12	P2341-01	1	Water	NA	NA	23.0	8.02	05/02/2024	08:49
13	P2341-02	1	Water	NA	NA	23.1	6.95	05/02/2024	08:55
14	P2347-01	1	Water	NA	NA	23.0	8.60	05/02/2024	09:00
15	CCV2	1	Water	NA	NA	23.0	12.02	05/02/2024	09:05

WORKLIST(Hardcopy Internal Chain)

87130590

WorkList Name : ph p2339 WorkList ID : 179909 Department : Wet-Chemistry Date : 05-02-2024 07:45:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2339-04	A D3513	Water	pH	Cool 4 deg C	PSEG03	K13	05/01/2024	9040C
P2339-05	A D3507	Water	pH	Cool 4 deg C	PSEG03	K13	05/01/2024	9040C
P2339-06	A T3438	Water	pH	Cool 4 deg C	PSEG03	K13	05/01/2024	9040C
P2339-07	A T3484	Water	pH	Cool 4 deg C	PSEG03	K13	05/01/2024	9040C
P2339-08	A T3514-15	Water	pH	Cool 4 deg C	PSEG03	K13	05/01/2024	9040C
P2341-01	A D237	Water	pH	Cool 4 deg C	PSEG03	K12	05/01/2024	9040C
P2341-02	A D239	Water	pH	Cool 4 deg C	PSEG03	K12	05/01/2024	9040C
P2347-01	A 392	Water	pH	Cool 4 deg C	PSEG03	K12	05/01/2024	9040C

Date/Time 05/02/24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 05/02/24 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]