

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130591

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/02/2024	08:05
2	CAL2	1	Water	NA	NA	20.3	7.00	05/02/2024	08:06
3	CAL3	1	Water	NA	NA	20.3	10.02	05/02/2024	08:10
4	ICV	1	Water	NA	NA	20.2	7.01	05/02/2024	08:11
5	CCV1	1	Water	NA	NA	20.3	2.02	05/02/2024	09:15
6	P2339-01	1	Solid	20.01	20	21.6	7.25	05/02/2024	09:25
7	P2339-01DUP	1	Solid	20.02	20	21.7	7.26	05/02/2024	09:26
8	P2339-02	1	Solid	20.03	20	21.9	7.28	05/02/2024	09:33
9	P2341-05	1	Solid	20.01	20	20.2	4.57	05/02/2024	09:40
10	CCV2	1	Water	NA	NA	20.3	12.02	05/02/2024	09:44

WORKLIST(Hardcopy Internal Chain)

87130591

WorkList Name : corrosion p2341 WorkList ID : 179910 Department : Wet-Chemistry Date : 05-02-2024 07:46:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2339-01	F D3479	Solid	Corrosivity	Cool 4 deg C	PSEG03	K13	05/01/2024	9045D
P2339-02	P D3522	Solid	Corrosivity	Cool 4 deg C	PSEG03	K13	05/01/2024	9045D
P2341-05	C T240-241	Solid	Corrosivity	Cool 4 deg C	PSEG03		05/01/2024	9045D

Date/Time 05/02/24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 05/02/24 13:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]