

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130601

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/02/2024	11:40
2	P2339-04		1	11	100.00	>212.0	>212.0	05/02/2024	12:10
3	P2339-05		1	13	100.00	>212.0	>212.0	05/02/2024	12:40
4	P2339-06		1	14	100.00	>212.0	>212.0	05/02/2024	13:10
5	P2339-07		1	11	100.00	>212.0	>212.0	05/02/2024	13:40
6	P2339-08		1	12	100.00	>212.0	>212.0	05/02/2024	14:10
7	P2341-01		1	14	100.00	>212.0	>212.0	05/02/2024	14:40
8	P2341-02		1	14	100.00	>212.0	>212.0	05/02/2024	15:10
9	P2347-01		1	11	100.00	>212.0	>212.0	05/02/2024	15:40
10	P2347-01DUP		1	11	100.00	>212.0	>212.0	05/02/2024	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB130601

WorkList Name : FP-5-2 WorkList ID : 179923 Department : Wet-Chemistry Date : 05-02-2024 08:39:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2339-04	D3513	Water	Flash Point	Cool 4 deg C	PSEG03	K13	05/01/2024	1010B
P2339-05	D3507	Water	Flash Point	Cool 4 deg C	PSEG03	K13	05/01/2024	1010B
P2339-06	T3438	Water	Flash Point	Cool 4 deg C	PSEG03	K13	05/01/2024	1010B
P2339-07	T3484	Water	Flash Point	Cool 4 deg C	PSEG03	K13	05/01/2024	1010B
P2339-08	T3514-15	Water	Flash Point	Cool 4 deg C	PSEG03	K13	05/01/2024	1010B
P2341-01	D237	Water	Flash Point	Cool 4 deg C	PSEG03	K12	05/01/2024	1010B
P2341-02	D239	Water	Flash Point	Cool 4 deg C	PSEG03	K12	05/01/2024	1010B
P2347-01	392	Water	Flash Point	Cool 4 deg C	PSEG03	K12	05/01/2024	1010B

Date/Time 05/02/2024 11:30
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: DP (CWC)

Date/Time 05/02/2024 16:45
 Raw Sample Received by: B (CWC)
 Raw Sample Relinquished by: RM CWL