

LB130602


SFAM01.1

INSTRUMENT 11: CVT

Linear

 μ Abs.:

4429E



A= 0.0000e+000

B= 2.2721e-004 *slope*C= -1.8596e-002 *y=intercept*

Rho= 0.9998368

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>9/10</i>
0.00	0.000	0.060	0.060	345	0.000	345					<i>5</i>
0.05	0.050					442	0				<i>11</i>
0.20	0.200	0.221	0.021	1056	0.0 %	1056					<i>-5</i>
2.50	2.500	2.381	-0.119	10561	0.0 %	10561					<i>-1</i>
5.00	5.000	5.053	0.053	22320	0.0 %	22320					<i>-1</i>
7.50	7.500	7.440	-0.060	32827	0.0 %	32827					<i>-1</i>
10.0	10.000	10.045	0.045	44295	0.0 %	44295					<i>0</i>

LB130602

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	345	-		0 SFAM01.1	PPB	5/2/2024 16:07	Std
	0.2 S01	1056	-		0.2 SFAM01.1	PPB	5/2/2024 16:09	Std
	2.5 S02	10561	-		2.5 SFAM01.1	PPB	5/2/2024 16:11	Std
	5 S03	22320	-		5 SFAM01.1	PPB	5/2/2024 16:14	Std
	7.5 S04	32827	-		7.5 SFAM01.1	PPB	5/2/2024 16:16	Std
	10 S05	44295	-		10 SFAM01.1	PPB	5/2/2024 16:18	Std
ICV027	ICV027	17037	3.8523 -		SFAM01.1	PPB	5/2/2024 16:21	SMPL
ICB027	ICB027	23	-0.0134 -		SFAM01.1	PPB	5/2/2024 16:23	SMPL
CCV012	CCV012	22574	5.1103 -		SFAM01.1	PPB	5/2/2024 16:26	SMPL
CCB012	CCB012	5	-0.0175 -		SFAM01.1	PPB	5/2/2024 16:28	SMPL
PB160691BL	PBW691	93	0.0025 -		SFAM01.1	PPB	5/2/2024 16:30	SMPL
P2132-01	YBRS9	150	0.0155 -		SFAM01.1	PPB	5/2/2024 16:33	SMPL
P2132-02	YBRT1	203	0.0275 -		SFAM01.1	PPB	5/2/2024 16:35	SMPL
P2132-03	YBRT3	147	0.0148 -		SFAM01.1	PPB	5/2/2024 16:37	SMPL
P2132-04	YCTJ3	147	0.0148 -		SFAM01.1	PPB	5/2/2024 16:39	SMPL
P2132-05	YCTJ6	139	0.013 -		SFAM01.1	PPB	5/2/2024 16:42	SMPL
P2132-06	YCTJ8	154	0.0164 -		SFAM01.1	PPB	5/2/2024 16:44	SMPL
P2132-07	YCTL8	145	0.0143 -		SFAM01.1	PPB	5/2/2024 16:46	SMPL
P2132-08	YCTK8	197	0.0262 -		SFAM01.1	PPB	5/2/2024 16:48	SMPL
P2132-09	YCTL2	113	0.0071 -		SFAM01.1	PPB	5/2/2024 16:51	SMPL
P2132-10	YCTM0	154	0.0164 -		SFAM01.1	PPB	5/2/2024 16:53	SMPL
P2132-11	YCTM2	149	0.0153 -		SFAM01.1	PPB	5/2/2024 16:55	SMPL
P2132-12	YCTM4	180	0.0223 -		SFAM01.1	PPB	5/2/2024 16:57	SMPL
P2132-13	YCTM6	177	0.0216 -		SFAM01.1	PPB	5/2/2024 17:00	SMPL
P2132-14	YBRS7	180	0.0223 -		SFAM01.1	PPB	5/2/2024 17:02	SMPL
P2132-15	YCTJ4	192	0.025 -		SFAM01.1	PPB	5/2/2024 17:04	SMPL
P2132-16	YCTJ4D	194	0.0255 -		SFAM01.1	PPB	5/2/2024 17:07	SMPL
P2132-17	YCTJ4S	4608	1.0284 -		SFAM01.1	PPB	5/2/2024 17:11	SMPL
P2132-18	YCTK0	206	0.0282 -		SFAM01.1	PPB	5/2/2024 17:14	SMPL
P2132-19	YCTL0	211	0.0293 -		SFAM01.1	PPB	5/2/2024 17:16	SMPL
CCV013	CCV013	20823	4.7125 -		SFAM01.1	PPB	5/2/2024 17:18	SMPL
CCB013	CCB013	52	-0.0068 -		SFAM01.1	PPB	5/2/2024 17:21	SMPL
P2132-20	YCTM8	143	0.0139 -		SFAM01.1	PPB	5/2/2024 17:23	SMPL
P2132-21	YCTK4	142	0.0137 -		SFAM01.1	PPB	5/2/2024 17:25	SMPL
P2132-22	YCTK6	166	0.0191 -		SFAM01.1	PPB	5/2/2024 17:27	SMPL
PB160692BL	PBW692	278	0.0446 -		SFAM01.1	PPB	5/2/2024 17:30	SMPL
P2134-01	YBRT0	167	0.0193 -		SFAM01.1	PPB	5/2/2024 17:32	SMPL
P2134-02	YBRT2	294	0.0482 -		SFAM01.1	PPB	5/2/2024 17:34	SMPL
P2134-03	YBRT4	191	0.0248 -		SFAM01.1	PPB	5/2/2024 17:36	SMPL
P2134-04	YCTJ7	164	0.0187 -		SFAM01.1	PPB	5/2/2024 17:39	SMPL
P2134-05	YCTJ9	151	0.0157 -		SFAM01.1	PPB	5/2/2024 17:41	SMPL
P2134-06	YCTL9	146	0.0146 -		SFAM01.1	PPB	5/2/2024 17:43	SMPL
P2134-07	YCTJ2	136	0.0123 -		SFAM01.1	PPB	5/2/2024 17:45	SMPL
P2134-08	YCTK9	1579	0.3402 -		SFAM01.1	PPB	5/2/2024 17:48	SMPL
P2134-09	YCTL3	-15	-0.022 -		SFAM01.1	PPB	5/2/2024 17:50	SMPL
P2134-10	YCTM1	1302	0.2772 -		SFAM01.1	PPB	5/2/2024 17:52	SMPL

LB130602

INSTRUMENT ID : CV1

P2134-11	YCTM3	41	-0.0093 -	SFAM01.1	PPB	5/2/2024 17:55 SMPL
P2134-12	YCTM5	197	0.0262 -	SFAM01.1	PPB	5/2/2024 17:57 SMPL
P2134-13	YCTM7	195	0.0257 -	SFAM01.1	PPB	5/2/2024 17:59 SMPL
P2134-14	YBRS8	164	0.0187 -	SFAM01.1	PPB	5/2/2024 18:01 SMPL
P2134-15	YCTJ5	449	0.0834 -	SFAM01.1	PPB	5/2/2024 18:04 SMPL
P2134-16	YCTJ5D	493	0.0934 -	SFAM01.1	PPB	5/2/2024 18:06 SMPL
CCV014	CCV014	22350	5.0594 -	SFAM01.1	PPB	5/2/2024 18:08 SMPL
CCB014	CCB014	85	0.0007 -	SFAM01.1	PPB	5/2/2024 18:10 SMPL
P2134-17	YCTJ5S	2483	0.5456 -	SFAM01.1	PPB	5/2/2024 18:13 SMPL
P2134-18	YCTK1	423	0.0775 -	SFAM01.1	PPB	5/2/2024 18:15 SMPL
P2134-19	YCTL1	2384	0.5231 -	SFAM01.1	PPB	5/2/2024 18:17 SMPL
P2134-20	YCTM9	98	0.0037 -	SFAM01.1	PPB	5/2/2024 18:20 SMPL
P2134-21	YCTK5	125	0.0098 -	SFAM01.1	PPB	5/2/2024 18:22 SMPL
P2134-22	YCTK7	292	0.0477 -	SFAM01.1	PPB	5/2/2024 18:24 SMPL
PB160693BL	PBW693	262	0.0409 -	SFAM01.1	PPB	5/2/2024 18:26 SMPL
P2135-01	YCTK3	218	0.0309 -	SFAM01.1	PPB	5/2/2024 18:29 SMPL
P2135-02	YCTK3D	244	0.0368 -	SFAM01.1	PPB	5/2/2024 18:31 SMPL
P2135-03	YCTK3S	2250	0.4926 -	SFAM01.1	PPB	5/2/2024 18:33 SMPL
P2135-04	YCTL5	130	0.0109 -	SFAM01.1	PPB	5/2/2024 18:35 SMPL
P2135-05	YCTN1	156	0.0168 -	SFAM01.1	PPB	5/2/2024 18:38 SMPL
CCV015	CCV015	20871	4.7234 -	SFAM01.1	PPB	5/2/2024 18:40 SMPL
CCB015	CCB015	91	0.0021 -	SFAM01.1	PPB	5/2/2024 18:42 SMPL