

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130607

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/03/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.01	05/03/2024	08:01
3	CAL3	1	Water	NA	NA	20.3	10.02	05/03/2024	08:05
4	ICV	1	Water	NA	NA	20.2	7.01	05/03/2024	08:11
5	CCV1	1	Water	NA	NA	20.3	2.01	05/03/2024	08:15
6	P2354-02	1	Solid	20.02	20	22.0	10.66	05/03/2024	08:22
7	P2355-02	1	Solid	20.03	20	22.5	7.61	05/03/2024	08:27
8	P2355-02DUP	1	Solid	20.04	20	22.6	7.62	05/03/2024	08:29
9	CCV2	1	Water	NA	NA	20.2	12.02	05/03/2024	08:33

WORKLIST(Hardcopy Internal Chain)

130607

WorkList Name : corrosivity p2354 WorkList ID : 179954 Department : Wet-Chemistry Date : 05-03-2024 07:46:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2354-02	V359	Solid	Corrosivity	Cool 4 deg C	PSEG03	K21	05/02/2024	9045D
P2355-02	VB16173	Solid	Corrosivity	Cool 4 deg C	PSEG03	K21	05/02/2024	9045D

Date/Time 05/03/24 07:50
Raw Sample Received by: JH (WR)
Raw Sample Relinquished by: CUL

Date/Time 05/03/24 11:00
Raw Sample Received by: QP
Raw Sample Relinquished by: JH (WR)