

LB130609


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41792



A= 0.0000e+000

B= 2.4023e-004

C= 8.2545e-003

Rho= 0.9999563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	112	0.000	112	0			
0.2	0.200	0.183	-0.017	728	0.0 %	728				
2.5	2.500	2.507	0.007	10402	0.0 %	10402				
5.0	5.000	4.980	-0.020	20697	0.0 %	20697				
7.5	7.500	7.446	-0.054	30962	0.0 %	30962				
10.0	10.000	10.048	0.048	41792	0.0 %	41792				

0.61
1.9
1.0
1.0
0

LB130609 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 May 2024 12:16:36

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	112	-	0.0000	7470A	PPB	03 May 2024 12:20:35	S	Std
0.2 - 1	50.2	728	-	0.2000	7470A	PPB	03 May 2024 12:22:52	S	Std
2.5 - 1	52.5	10402	-	2.5000	7470A	PPB	03 May 2024 12:25:09	S	Std
5.0 - 1	55	20697	-	5.0000	7470A	PPB	03 May 2024 12:27:26	S	Std
7.5 - 1	57.5	30962	-	7.5000	7470A	PPB	03 May 2024 12:29:44	S	Std
10.0 - 1	50	41792	-	10.0000	7470A	PPB	03 May 2024 12:32:02	S	Std
ICV28 - 1	ICV28	17156	4.1297	-	7470A	PPB	03 May 2024 12:35:41	U	SMPL
ICB28 - 1	ICB28	-123	-0.0213	-	7470A	PPB	03 May 2024 12:37:55	U	SMPL
CCV16 - 1	CCV16	20631	4.9645	-	7470A	PPB	03 May 2024 12:40:12	U	SMPL
CCB16 - 1	CCB16	-158	-0.0297	-	7470A	PPB	03 May 2024 12:42:27	U	SMPL
CRA - 1	CRA	700	0.1764	-	7470A	PPB	03 May 2024 12:44:44	U	SMPL
HighStd - 1	HighStd	40967	9.8498	-	7470A	PPB	03 May 2024 12:46:59	U	SMPL
ChkStd - 1	ChkStd	27690	6.6602	-	7470A	PPB	03 May 2024 12:49:14	U	SMPL
PB160695BL - 1	PBW	-358	-0.0777	-	7470A	PPB	03 May 2024 12:51:32	U	SMPL
PB160695BS - 1	LCSW	14595	3.5144	-	7470A	PPB	03 May 2024 12:57:12	U	SMPL
P2307-03 - 1	WC-1	-191	-0.0376	-	7470A	PPB	03 May 2024 12:59:28	U	SMPL
P2307-06 - 1	WC-2	62	0.0231	-	7470A	PPB	03 May 2024 13:01:47	U	SMPL
P2307-09 - 1	WC-3	91	0.0301	-	7470A	PPB	03 May 2024 13:04:04	U	SMPL
P2323-03 - 1	TP-1	109	0.0344	-	7470A	PPB	03 May 2024 13:06:21	U	SMPL
P2323-06 - 1	TP-2	118	0.0366	-	7470A	PPB	03 May 2024 13:08:36	U	SMPL
CCV17 - 1	CCV17	21266	5.1170	-	7470A	PPB	03 May 2024 13:15:58	U	SMPL
CCB17 - 1	CCB17	-362	-0.0787	-	7470A	PPB	03 May 2024 13:18:14	U	SMPL
P2323-09 - 1	TP-3	21	0.0133	-	7470A	PPB	03 May 2024 13:20:32	U	SMPL
P2323-09DUP - 1	TP-3DUP	70	0.0251	-	7470A	PPB	03 May 2024 13:22:48	U	SMPL
P2323-09MS - 1	TP-3MS	13492	3.2494	-	7470A	PPB	03 May 2024 13:25:04	U	SMPL
P2323-09MSD - 1	TP-3MSD	13581	3.2708	-	7470A	PPB	03 May 2024 13:27:19	U	SMPL
PB160630TB - 1	PB160630TB	-255	-0.0530	-	7470A	PPB	03 May 2024 13:29:37	U	SMPL
P2323-09LX5 - 1		-47	-0.0030	-	7470A	PPB	03 May 2024 13:31:56	U	SMPL
P2323-09A - 1		12412	2.9900	-	7470A	PPB	03 May 2024 13:34:12	U	SMPL
CCV18 - 1	CCV18	21948	5.2808	-	7470A	PPB	03 May 2024 13:50:29	U	SMPL
CCB18 - 1	CCB18	-296	-0.0629	-	7470A	PPB	03 May 2024 13:52:47	U	SMPL