

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130612

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3086

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	05/03/2024	08:30
2	P2341-05		1	4	7.00	44.6	43.9	05/03/2024	09:00
3	P2341-05DUP		1	4	7.00	44.6	43.9	05/03/2024	09:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16130612

WorkList Name : fp5-3

WorkList ID : 179978

Department : Wet-Chemistry

Date : 05-03-2024 08:14:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2341-05	T240-241	Solid	Flash Point	Cool 4 deg C	PSEG03	K12	05/01/2024	1010B

Date/Time 05/03/2024 08:20
Raw Sample Received by: RM CWS
Raw Sample Relinquished by: JH GRC

Date/Time 05/03/2024 10:15
Raw Sample Received by: JH GRC
Raw Sample Relinquished by: RM CWS