

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/6/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 05/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:44
Out Date: 05/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB130613

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2368-01	A43K6	1	1.18	8.68	9.86	8.59	85.4	
P2368-02	A43K7	2	1.16	8.82	9.98	8.47	82.9	
P2368-03	A43K8	3	1.16	8.49	9.65	7.98	80.3	
P2368-04	A43K9	4	1.16	8.40	9.56	7.95	80.8	
P2368-05	A43K9MS	5	1.16	8.40	9.56	7.95	80.8	
P2368-06	A43K9MSD	6	1.16	8.40	9.56	7.95	80.8	
P2368-07	A43L0	7	1.17	8.42	9.59	8.31	84.8	
P2368-08	A43L1	8	1.18	8.50	9.68	8.16	82.1	
P2368-09	A43L2	9	1.18	8.60	9.78	8.23	82.0	
P2368-10	A43L3	10	1.15	8.65	9.8	8.1	80.3	
P2368-11	A43L4	11	1.16	8.54	9.7	8.46	85.5	
P2368-12	A43L5	12	1.16	8.71	9.87	8.56	85.0	
P2368-13	A43L6	13	1.17	8.51	9.68	8.02	80.5	
P2368-14	A43L7	14	1.17	8.56	9.73	7.67	75.9	
P2368-15	A43L8	15	1.18	8.52	9.7	8.51	86.0	
P2368-16	A43L9	16	1.17	8.62	9.79	8.22	81.8	
P2368-17	A43M0	17	1.17	8.50	9.67	8.03	80.7	
P2368-18	A43M1	18	1.17	8.38	9.55	8.19	83.8	
P2368-19	A43M2	19	1.17	8.63	9.8	8.37	83.4	
P2368-20	A43M3	20	1.19	8.50	9.69	7.68	76.4	
P2368-21	A43M9	21	1.17	8.70	9.87	7.74	75.5	
P2368-22	A43X6	22	1.14	8.49	9.63	7.59	76.0	
P2368-23	A43Y4	23	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$