

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/6/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:30
In Date: 05/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 05/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB130616

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2371-01	A43Q5	1	1.16	8.82	9.98	8.88	87.5	
P2371-02	A43Q5MS	2	1.16	8.82	9.98	8.88	87.5	
P2371-03	A43Q5MSD	3	1.16	8.82	9.98	8.88	87.5	
P2371-04	A43R5	4	1.17	8.80	9.97	8.28	80.8	
P2371-05	A43R7	5	1.16	8.42	9.58	8.47	86.8	
P2371-06	A43R8	6	1.16	8.60	9.76	8.44	84.7	
P2371-07	A43R9	7	1.16	8.82	9.98	8.53	83.6	
P2371-08	A43S0	8	1.14	8.50	9.64	8.38	85.2	
P2371-09	A43S1	9	1.16	8.79	9.95	8.34	81.7	
P2371-10	A43S2	10	1.14	8.42	9.56	7.8	79.1	
P2371-11	A43S3	11	1.16	8.53	9.69	7.76	77.4	
P2371-12	A43S4	12	1.17	8.59	9.76	8.88	89.8	
P2371-13	A43S5	13	1.15	8.83	9.98	8.6	84.4	
P2371-14	A43S6	14	1.15	8.83	9.98	8.32	81.2	
P2371-15	A43S7	15	1.16	8.55	9.71	8.32	83.7	
P2371-16	A43S8	16	1.17	8.69	9.86	9.22	92.6	
P2371-17	A43S9	17	1.16	8.52	9.68	7.97	79.9	
P2371-18	A43T0	18	1.17	8.52	9.69	8.25	83.1	
P2371-19	A43T1	19	1.16	8.58	9.74	8.1	80.9	
P2371-20	A43T2	20	1.16	8.81	9.97	8.46	82.9	
P2371-21	A43T3	21	1.17	8.56	9.73	8.06	80.5	
P2371-22	A43X8	22	1.16	8.82	9.98	8.43	82.4	
P2371-23	A43Y7	23	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$