

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/6/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:40
In Date: 05/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 05/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB130619

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2372-01	A43R1	1	1.19	8.73	9.92	8.14	79.6	
P2372-02	A43R1MS	2	1.19	8.73	9.92	8.14	79.6	
P2372-03	A43R1MSD	3	1.19	8.73	9.92	8.14	79.6	
P2372-04	A43T4	4	1.16	8.50	9.66	8.00	80.5	
P2372-05	A43T5	5	1.19	8.72	9.91	8.26	81.1	
P2372-06	A43T6	6	1.18	8.60	9.78	8.48	84.9	
P2372-07	A43T7	7	1.12	8.79	9.91	8.44	83.3	
P2372-08	A43T8	8	1.19	8.62	9.81	8.21	81.4	
P2372-09	A43T9	9	1.19	8.52	9.71	7.96	79.5	
P2372-10	A43W0	10	1.16	8.73	9.89	9.14	91.4	
P2372-11	A43W1	11	1.15	8.65	9.8	8.72	87.5	
P2372-12	A43W2	12	1.17	8.50	9.67	8.04	80.8	
P2372-13	A43W3	13	1.17	8.60	9.77	8.27	82.6	
P2372-14	A43W4	14	1.16	8.43	9.59	8.21	83.6	
P2372-15	A43W5	15	1.18	8.40	9.58	7.98	81.0	
P2372-16	A43W6	16	1.16	8.50	9.66	8.08	81.4	
P2372-17	A43W7	17	1.19	8.58	9.77	8.00	79.4	
P2372-18	A43W8	18	1.15	8.73	9.88	8.39	82.9	
P2372-19	A43W9	19	1.18	8.81	9.99	8.5	83.1	
P2372-20	A43X0	20	1.14	8.82	9.96	8.39	82.2	
P2372-21	A43X1	21	1.16	8.82	9.98	8.68	85.3	
P2372-22	A43Y0	22	1.12	8.61	9.73	8.16	81.8	
P2372-23	A43Y3	23	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$