

LB130621


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

41841



A= 0.0000e+000

B= 2.4161e-004

C= -7.0262e-002

Rho= 0.9999616

Accept=Accepted

5 slope
y=mx+cept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/dl
0.00	0.000	-0.005	-0.005	269	0.000	269					11
0.05	0.050					442	0				24
0.20	0.200	0.248	0.048	1317	0.0 %	1317					11
2.50	2.500	2.478	-0.022	10547	0.0 %	10547					11
5.00	5.000	4.961	-0.039	20825	0.0 %	20825					11
7.50	7.500	7.479	-0.021	31245	0.0 %	31245					11
10.0	10.000	10.039	0.039	41841	0.0 %	41841					0

LB130621

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	269	-		0 SFAM01.1	PPB	5/3/2024 17:21	Std
	0.2 S01	1317	-		0.2 SFAM01.1	PPB	5/3/2024 17:23	Std
	2.5 S02	10547	-		2.5 SFAM01.1	PPB	5/3/2024 17:25	Std
	5 S03	20825	-		5 SFAM01.1	PPB	5/3/2024 17:27	Std
	7.5 S04	31245	-		7.5 SFAM01.1	PPB	5/3/2024 17:33	Std
	10 S05	41841	-		10 SFAM01.1	PPB	5/3/2024 17:43	Std
ICV029	ICV029	16507	3.918 -		SFAM01.1	PPB	5/3/2024 17:46	SMPL
ICB029	ICB029	21	-0.0652 -		SFAM01.1	PPB	5/3/2024 17:48	SMPL
CCV019	CCV019	19960	4.7523 -		SFAM01.1	PPB	5/3/2024 17:51	SMPL
CCB019	CCB019	-28	-0.077 -		SFAM01.1	PPB	5/3/2024 17:53	SMPL
PB160713BL	PBS713	143	-0.0357 -		SFAM01.1	PPB	5/3/2024 17:55	SMPL
P2168-01	YCM06	9697	2.2727 -		SFAM01.1	PPB	5/3/2024 17:57	SMPL
P2168-02	YCMN3	2815	0.6099 -		SFAM01.1	PPB	5/3/2024 18:00	SMPL
P2168-03	YCMPO	3181	0.6983 -		SFAM01.1	PPB	5/3/2024 18:03	SMPL
P2168-04	YCMP2	1928	0.3956 -		SFAM01.1	PPB	5/3/2024 18:05	SMPL
P2168-05	YCMP6	2516	0.5376 -		SFAM01.1	PPB	5/3/2024 18:07	SMPL
P2168-06	YCMP7	1983	0.4089 -		SFAM01.1	PPB	5/3/2024 18:10	SMPL
P2168-07	YCTC3	3000	0.6546 -		SFAM01.1	PPB	5/3/2024 18:12	SMPL
P2168-08	YCTC3D	2975	0.6485 -		SFAM01.1	PPB	5/3/2024 18:14	SMPL
P2168-09	YCTC3S	15315	3.63 -		SFAM01.1	PPB	5/3/2024 18:16	SMPL
P2168-10	YCTC4	2926	0.6367 -		SFAM01.1	PPB	5/3/2024 18:19	SMPL
P2168-11	YCTC5	3985	0.8926 -		SFAM01.1	PPB	5/3/2024 18:21	SMPL
P2168-12	YCTC6	7117	1.6493 -		SFAM01.1	PPB	5/3/2024 18:23	SMPL
P2168-13	YCTC7	3359	0.7413 -		SFAM01.1	PPB	5/3/2024 18:26	SMPL
P2168-14	YCTC8	2358	0.4995 -		SFAM01.1	PPB	5/3/2024 18:28	SMPL
P2168-15	YCTC9	3031	0.6621 -		SFAM01.1	PPB	5/3/2024 18:30	SMPL
P2168-16	YCTD0	4113	0.9235 -		SFAM01.1	PPB	5/3/2024 18:32	SMPL
P2168-17	YCTD1	2307	0.4871 -		SFAM01.1	PPB	5/3/2024 18:35	SMPL
P2168-18	YCTD2	3474	0.7691 -		SFAM01.1	PPB	5/3/2024 18:37	SMPL
P2168-19	YCTD3	2603	0.5587 -		SFAM01.1	PPB	5/3/2024 18:39	SMPL
CCV020	CCV020	22565	5.3817 -		SFAM01.1	PPB	5/3/2024 18:41	SMPL
CCB020	CCB020	83	-0.0502 -		SFAM01.1	PPB	5/3/2024 18:44	SMPL
P2168-20	YCTD4	4160	0.9348 -		SFAM01.1	PPB	5/3/2024 18:46	SMPL
P2168-21	YCTD5	15520	3.6796 -		SFAM01.1	PPB	5/3/2024 18:48	SMPL
P2168-22	YCTD6	5649	1.2946 -		SFAM01.1	PPB	5/3/2024 18:51	SMPL
P2169-01	YCTD7	4604	1.0421 -		SFAM01.1	PPB	5/3/2024 18:55	SMPL
P2169-02	YCTD8	4890	1.1112 -		SFAM01.1	PPB	5/3/2024 18:57	SMPL
P2169-03	YCTD9	3464	0.7667 -		SFAM01.1	PPB	5/3/2024 19:00	SMPL
P2169-04	YCTE0	1325	0.2499 -		SFAM01.1	PPB	5/3/2024 19:02	SMPL
P2169-05	YCTE1	1714	0.3439 -		SFAM01.1	PPB	5/3/2024 19:04	SMPL
P2169-06	YCTE2	4631	1.0486 -		SFAM01.1	PPB	5/3/2024 19:06	SMPL
P2169-07	YCTE3	3434	0.7594 -		SFAM01.1	PPB	5/3/2024 19:09	SMPL
P2169-08	YCTE3D	3204	0.7039 -		SFAM01.1	PPB	5/3/2024 19:11	SMPL
P2169-09	YCTE3S	15363	3.6416 -		SFAM01.1	PPB	5/3/2024 19:13	SMPL
P2169-10	YCTE4	1235	0.2281 -		SFAM01.1	PPB	5/3/2024 19:16	SMPL
P2169-11	YCMR6	1881	0.3842 -		SFAM01.1	PPB	5/3/2024 19:18	SMPL

LB130621

INSTRUMENT ID : CV1

P2169-12	YCMR8	2356	0.499 -	SFAM01.1	PPB	5/3/2024 19:20 SMPL
P2169-13	YCMR9	2634	0.5661 -	SFAM01.1	PPB	5/3/2024 19:22 SMPL
P2169-14	YCMS0	15017	3.558 -	SFAM01.1	PPB	5/3/2024 19:25 SMPL
P2169-15	YCTE5	1262	0.2347 -	SFAM01.1	PPB	5/3/2024 19:27 SMPL
P2169-16	YCTE6	3111	0.6814 -	SFAM01.1	PPB	5/3/2024 19:29 SMPL
CCV021	CCV021	23204	5.5361 -	SFAM01.1	PPB	5/3/2024 19:31 SMPL
CCB021	CCB021	87	-0.0492 -	SFAM01.1	PPB	5/3/2024 19:34 SMPL
P2169-17	YCTE7	2827	0.6128 -	SFAM01.1	PPB	5/3/2024 19:36 SMPL
P2169-18	YCTE8	3289	0.7244 -	SFAM01.1	PPB	5/3/2024 19:38 SMPL
P2169-19	YCTF5	10318	2.4227 -	SFAM01.1	PPB	5/3/2024 19:41 SMPL
PB160719BL	PBS719	-24	-0.0761 -	SFAM01.1	PPB	5/3/2024 19:43 SMPL
P2351-01	E08P9	882	0.1428 -	SFAM01.1	PPB	5/3/2024 19:45 SMPL
P2351-02	E08Q0	832	0.1308 -	SFAM01.1	PPB	5/3/2024 19:47 SMPL
P2351-03	E08Q1	967	0.1634 -	SFAM01.1	PPB	5/3/2024 19:50 SMPL
P2351-04	E08Q2	1188	0.2168 -	SFAM01.1	PPB	5/3/2024 19:52 SMPL
P2351-05	E08Q3	1465	0.2837 -	SFAM01.1	PPB	5/3/2024 19:54 SMPL
P2351-06	E08Q3D	1641	0.3262 -	SFAM01.1	PPB	5/3/2024 19:59 SMPL
P2351-07	E08Q3S	13238	3.1282 -	SFAM01.1	PPB	5/3/2024 20:01 SMPL
PB160714BL	PBS714	119	-0.0415 -	SFAM01.1	PPB	5/3/2024 20:04 SMPL
CCV022	CCV022	22070	5.2621 -	SFAM01.1	PPB	5/3/2024 20:06 SMPL
CCB022	CCB022	107	-0.0444 -	SFAM01.1	PPB	5/3/2024 20:08 SMPL