

**PERCENT SOLID**

**Supervisor:** Iwona  
**Analyst:** jignesh  
**Date:** 5/7/2024

**OVENTEMP IN Celsius(°C):** 107  
**Time IN:** 17:00  
**In Date:** 05/06/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**OvenID:** M OVEN#1

**OVENTEMP OUT Celsius(°C):** 103  
**Time OUT:** 08:11  
**Out Date:** 05/07/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**BalanceID:** M SC-4  
**Thermometer ID:** % SOLID- OVEN

QC:LB130632

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments        |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|-----------------|
| P2312-01 | B-125-SB-01     | 33     | 1.11            | 8.77          | 9.88                     | 8.02                       | 78.8    |                 |
| P2312-02 | B-125-SB-02     | 34     | 1.16            | 8.48          | 9.64                     | 7.91                       | 79.6    |                 |
| P2375-01 | WC-A-COMP       | 1      | 1.17            | 8.57          | 9.74                     | 9.02                       | 91.6    |                 |
| P2375-03 | WC-A-GRAB       | 2      | 1.19            | 8.72          | 9.91                     | 9.19                       | 91.7    |                 |
| P2388-01 | 1405            | 19     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | wipe sample     |
| P2388-02 | 1406            | 20     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | wipe sample     |
| P2388-03 | 1407            | 21     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | wipe sample     |
| P2388-04 | 1407-CONCRETE   | 22     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | CONCRETE sample |
| P2388-05 | 1404            | 23     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | wipe sample     |
| P2390-01 | T-1             | 3      | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | oil sample      |
| P2390-02 | BG-1            | 4      | 1.14            | 8.59          | 9.73                     | 7.74                       | 76.8    |                 |
| P2390-03 | BG-2            | 5      | 1.15            | 8.55          | 9.7                      | 8.11                       | 81.4    |                 |
| P2390-04 | BG-3            | 6      | 1.16            | 8.49          | 9.65                     | 7.57                       | 75.5    |                 |
| P2390-05 | PE-1            | 7      | 1.18            | 8.54          | 9.72                     | 8.4                        | 84.5    |                 |
| P2390-06 | PE-2            | 8      | 1.19            | 8.43          | 9.62                     | 8.3                        | 84.3    |                 |
| P2390-07 | PE-3            | 9      | 1.18            | 8.46          | 9.64                     | 7.98                       | 80.4    |                 |
| P2390-08 | PE-4            | 10     | 1.19            | 8.67          | 9.86                     | 8.28                       | 81.8    |                 |
| P2390-09 | PE-5            | 11     | 1.19            | 8.63          | 9.82                     | 8.41                       | 83.7    |                 |
| P2390-10 | D5224           | 12     | 1.17            | 8.73          | 9.9                      | 8.6                        | 85.1    |                 |
| P2391-04 | PE-1            | 13     | 1.19            | 8.62          | 9.81                     | 8.73                       | 87.5    |                 |
| P2391-05 | PE-2            | 14     | 1.15            | 8.40          | 9.55                     | 8.49                       | 87.4    |                 |
| P2391-06 | PE-3            | 15     | 1.11            | 8.56          | 9.67                     | 8.83                       | 90.2    |                 |
| P2391-07 | PE-4            | 16     | 1.12            | 8.67          | 9.79                     | 9.1                        | 92.0    |                 |
| P2391-08 | PE-5            | 17     | 1.16            | 8.50          | 9.66                     | 8.45                       | 85.8    |                 |
| P2391-09 | D827AP          | 18     | 1.19            | 8.50          | 9.69                     | 8.82                       | 89.8    |                 |
| P2395-01 | SP-DRIVEWAY     | 24     | 1.15            | 8.48          | 9.63                     | 9.00                       | 92.6    |                 |
| P2395-02 | SP-DRIVEWAY     | 25     | 1.19            | 8.56          | 9.75                     | 9.03                       | 91.6    |                 |
| P2395-03 | SP-DRIVEWAY-E2  | 26     | 1.14            | 8.63          | 9.77                     | 9.1                        | 92.2    |                 |

**PERCENT SOLID**

**Supervisor:** Iwona  
**Analyst:** jignesh  
**Date:** 5/7/2024

**OVENTEMP IN Celsius(°C):** 107  
**Time IN:** 17:00  
**In Date:** 05/06/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**OvenID:** M OVEN#1

**OVENTEMP OUT Celsius(°C):** 103  
**Time OUT:** 08:11  
**Out Date:** 05/07/2024  
**Weight Check 1.0g:** 1.00  
**Weight Check 10g:** 10.00  
**BalanceID:** M SC-4  
**Thermometer ID:** % SOLID- OVEN

QC:LB130632

| Lab ID   | Client SampleID   | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments       |
|----------|-------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------------|
| P2395-04 | SP-INSIDE-GATE    | 27     | 1.15            | 8.38          | 9.53                     | 8.19                       | 84.0    |                |
| P2395-05 | SP-INSIDE-GATE    | 28     | 1.19            | 8.67          | 9.86                     | 8.19                       | 80.7    |                |
| P2395-06 | SP-INSIDE-GATE-E2 | 29     | 1.19            | 8.41          | 9.6                      | 8.06                       | 81.7    |                |
| P2396-01 | 149TH-ST-21-PPE   | 35     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | plastic sample |
| P2396-02 | 149TH-ST-22-POLY  | 36     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | plastic sample |
| P2398-01 | WC-KK             | 30     | 1.13            | 8.58          | 9.71                     | 8.55                       | 86.5    |                |
| P2398-02 | WC-KK-EPH         | 31     | 1.2             | 8.31          | 9.51                     | 8.37                       | 86.3    |                |
| P2398-03 | WC-KK-VOC         | 32     | 1.16            | 8.81          | 9.97                     | 9.06                       | 89.7    |                |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

130632

WorkList Name : %1-050624

WorkList ID : 179997

Department : Wet-Chemistry

Date : 05-06-2024 08:21:08

| Sample   | Customer Sample | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-----------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| P2312-01 | B-125-SB-01     | Solid  | Percent Solids | Cool 4 deg C | PORT06   | I31                         | 05/06/2024   | Chemtech -SO |
| P2312-02 | B-125-SB-02     | Solid  | Percent Solids | Cool 4 deg C | PORT06   | I31                         | 05/06/2024   | Chemtech -SO |
| P2375-01 | WC-A-COMP       | Solid  | Percent Solids | Cool 4 deg C | EART12   | K22                         | 05/03/2024   | Chemtech -SO |
| P2375-03 | WC-A-GRAB       | Solid  | Percent Solids | Cool 4 deg C | EART12   | K22                         | 05/03/2024   | Chemtech -SO |
| P2388-01 | 1405            | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |
| P2388-02 | 1406            | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |
| P2388-03 | 1407            | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |
| P2388-04 | 1407-CONCRETE   | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |
| P2388-05 | 1404            | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |
| P2390-01 | T-1             | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-02 | BG-1            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-03 | BG-2            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-04 | BG-3            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-05 | PE-1            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-06 | PE-2            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-07 | PE-3            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-08 | PE-4            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-09 | PE-5            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2390-10 | D5224           | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K31                         | 05/02/2024   | Chemtech -SO |
| P2391-04 | PE-1            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K21                         | 05/02/2024   | Chemtech -SO |
| P2391-05 | PE-2            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K21                         | 05/02/2024   | Chemtech -SO |

Date/Time 05/06/24 15:30  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]

Date/Time 05/06/24 17:10  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]

# WORKLIST(Hardcopy Internal Chain)

130632

WorkList Name : %1-050624 WorkList ID : 17997 Department : Wet-Chemistry Date : 05-06-2024 08:21:08

| Sample   | Customer Sample   | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-------------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| P2391-06 | PE-3              | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K21                         | 05/02/2024   | Chemtech -SO |
| P2391-07 | PE-4              | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K21                         | 05/02/2024   | Chemtech -SO |
| P2391-08 | PE-5              | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K21                         | 05/02/2024   | Chemtech -SO |
| P2391-09 | D827AP            | Solid  | Percent Solids | Cool 4 deg C | DEWB01   | K21                         | 05/02/2024   | Chemtech -SO |
| P2395-01 | SP-DRIVEWAY       | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K11                         | 05/06/2024   | Chemtech -SO |
| P2395-02 | SP-DRIVEWAY       | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K11                         | 05/06/2024   | Chemtech -SO |
| P2395-03 | SP-DRIVEWAY-E2    | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K11                         | 05/06/2024   | Chemtech -SO |
| P2395-04 | SP-INSIDE-GATE    | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K11                         | 05/06/2024   | Chemtech -SO |
| P2395-05 | SP-INSIDE-GATE    | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K11                         | 05/06/2024   | Chemtech -SO |
| P2395-06 | SP-INSIDE-GATE-E2 | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K11                         | 05/06/2024   | Chemtech -SO |
| P2396-01 | 149TH-ST-21-PPE   | Solid  | Percent Solids | Cool 4 deg C | TULL02   | K31                         | 05/03/2024   | Chemtech -SO |
| P2396-02 | 149TH-ST-22-POLY  | Solid  | Percent Solids | Cool 4 deg C | TULL02   | K31                         | 05/03/2024   | Chemtech -SO |
| P2398-01 | WC-KK             | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |
| P2398-02 | WC-KK-EPH         | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |
| P2398-03 | WC-KK-VOC         | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | K31                         | 05/06/2024   | Chemtech -SO |

Date/Time 05/06/24 15:30  
Raw Sample Received by: [Signature]  
Raw Sample Relinquished by: [Signature]

Date/Time 05/06/24 15:10  
Raw Sample Received by: [Signature]  
Raw Sample Relinquished by: [Signature]