

LB130695


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

31559



A= 0.0000e+000

B= 3.1447e-004 slope

C= -8.4716e-002 y-intercept

Rho= 0.9995121

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Adj
0.00	0.000	-0.052	-0.052	103	0.000	103					1
0.05	0.050					442	0				1
0.20	0.200	0.155	-0.045	761	0.0 %	761					-23
2.50	2.500	2.477	-0.023	8145	0.0 %	8145					1
5.00	5.000	5.208	0.208	16830	0.0 %	16830					4
7.50	7.500	7.574	0.074	24354	0.0 %	24354					1
10.0	10.000	9.840	-0.160	31559	0.0 %	31559					-2

LB130635

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	103	-		0 SFAM01.1	PPB	5/6/2024 12:20	Std
	0.2 S01	761	-		0.2 SFAM01.1	PPB	5/6/2024 12:22	Std
	2.5 S02	8145	-		2.5 SFAM01.1	PPB	5/6/2024 12:25	Std
	5 S03	16830	-		5 SFAM01.1	PPB	5/6/2024 12:27	Std
	7.5 S04	24354	-		7.5 SFAM01.1	PPB	5/6/2024 12:29	Std
	10 S05	31559	-		10 SFAM01.1	PPB	5/6/2024 12:37	Std
ICV030	ICV030	12120	3.7266 -		SFAM01.1	PPB	5/6/2024 12:40	SMPL
ICB030	ICB030	35	-0.0737 -		SFAM01.1	PPB	5/6/2024 12:45	SMPL
CCV023	CCV023	15433	4.7685 -		SFAM01.1	PPB	5/6/2024 13:11	SMPL
CCB023	CCB023	4	-0.0835 -		SFAM01.1	PPB	5/6/2024 13:13	SMPL
PB160743BL	PBW743	319	0.0156 -		SFAM01.1	PPB	5/6/2024 13:15	SMPL
P2175-08	MDCOZ5	104	-0.052 -		SFAM01.1	PPB	5/6/2024 13:17	SMPL
P2175-09	MDCOZ6	130	-0.0438 -		SFAM01.1	PPB	5/6/2024 13:20	SMPL
P2351-08	E08R3	149	-0.0379 -		SFAM01.1	PPB	5/6/2024 13:22	SMPL
CCV024	CCV024	15919	4.9213 -		SFAM01.1	PPB	5/6/2024 13:24	SMPL
CCB024	CCB024	6	-0.0828 -		SFAM01.1	PPB	5/6/2024 13:26	SMPL