

LB130637


7471B

INSTRUMENT ID: EV1

Linear

μ Abs.:

56428



A= 0.0000e+000

B= 1.7615e-004

C= -4.2897e-002

Rho= 0.9991715

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	254	0.000	254				
0.2	0.200	0.192	-0.008	1331	0.0 %	1331				
2.5	2.500	2.602	0.102	15018	0.0 %	15018				
5.0	5.000	4.765	-0.235	27295	0.0 %	27295				
7.5	7.500	7.743	0.243	44199	0.0 %	44199				
10.0	10.000	9.897	-0.103	56428	0.0 %	56428				

9.10
-4
4
-5
3
1

LB130637 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 May 2024 14:23:34

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	254	-	0.0000	7471B	PPB	06 May 2024 14:34:18	S	Std
0.2 - 1	50.2	1331	-	0.2000	7471B	PPB	06 May 2024 14:36:35	S	Std
2.5 - 1	52.5	15018	-	2.5000	7471B	PPB	06 May 2024 14:38:52	S	Std
5.0 - 1	55	27295	-	5.0000	7471B	PPB	06 May 2024 14:41:09	S	Std
7.5 - 1	57.5	44199	-	7.5000	7471B	PPB	06 May 2024 14:43:28	S	Std
10.0 - 1	510	56428	-	10.0000	7471B	PPB	06 May 2024 14:45:47	S	Std
ICV31 - 1	ICV31	23157	4.0361	-	7471B	PPB	06 May 2024 14:48:57	U	SMPL
ICB31 - 1	ICB31	-116	-0.0633	-	7471B	PPB	06 May 2024 14:51:13	U	SMPL
CCV25 - 1	CCV25	27369	4.7780	-	7471B	PPB	06 May 2024 14:53:32	U	SMPL
CCB25 - 1	CCB25	-134	-0.0665	-	7471B	PPB	06 May 2024 14:55:47	U	SMPL
CRA - 1	CRA	1259	0.1789	-	7471B	PPB	06 May 2024 14:58:07	U	SMPL
HighStd - 1	HighStd	56123	9.8429	-	7471B	PPB	06 May 2024 15:00:23	U	SMPL
ChkStd - 1	ChkStd	41517	7.2701	-	7471B	PPB	06 May 2024 15:02:38	U	SMPL
PB160732BL - 1	PBS	-157	-0.0706	-	7471B	PPB	06 May 2024 15:05:04	U	SMPL
PB160732BS - 1	LCSS	25868	4.5136	-	7471B	PPB	06 May 2024 15:07:28	U	SMPL
P2331-01 - 1	TP-1	9399	1.6127	-	7471B	PPB	06 May 2024 15:09:44	U	SMPL
P2331-05 - 1	TP-2	4313	0.7168	-	7471B	PPB	06 May 2024 15:12:04	U	SMPL
P2331-09 - 1	TP-3	4893	0.8190	-	7471B	PPB	06 May 2024 15:14:25	U	SMPL
P2331-16 - 1	TP-4	1033	0.1391	-	7471B	PPB	06 May 2024 15:16:44	U	SMPL
P2332-01 - 1	PAT-DIS-02-05012024	4040	0.6687	-	7471B	PPB	06 May 2024 15:19:01	U	SMPL
CCV26 - 1	CCV26	28451	4.9686	-	7471B	PPB	06 May 2024 15:21:16	U	SMPL
CCB26 - 1	CCB26	131	-0.0198	-	7471B	PPB	06 May 2024 15:23:33	U	SMPL
P2354-01 - 1	V359	7913	1.3509	-	7471B	PPB	06 May 2024 15:25:51	U	SMPL
P2355-01 - 1	VB16173	15641	2.7122	-	7471B	PPB	06 May 2024 15:28:07	U	SMPL
P2357-01 - 1	HD-01-05022024	2443	0.3874	-	7471B	PPB	06 May 2024 15:30:25	U	SMPL
P2362-01 - 1	INSIDE	2734	0.4387	-	7471B	PPB	06 May 2024 15:32:43	U	SMPL
P2362-02 - 1	OUTSIDE-FOOTINGS	2466	0.3915	-	7471B	PPB	06 May 2024 15:34:59	U	SMPL
P2378-01 - 1	OR-02-05032024	4062	0.6726	-	7471B	PPB	06 May 2024 15:37:16	U	SMPL
P2386-01 - 1	EO-01-050324-A	1783	0.2712	-	7471B	PPB	06 May 2024 15:39:34	U	SMPL
P2386-03 - 1	EO-01-050324-B	24421	4.2587	-	7471B	PPB	06 May 2024 15:41:52	U	SMPL
P2386-03DUP - 1	EO-01-050324-BDUP	20046	3.4881	-	7471B	PPB	06 May 2024 15:44:09	U	SMPL
P2386-03MS - 1	EO-01-050324-BMS	49218	8.6266	-	7471B	PPB	06 May 2024 15:46:27	U	SMPL
CCV27 - 1	CCV27	29464	5.1470	-	7471B	PPB	06 May 2024 15:48:45	U	SMPL
CCB27 - 1	CCB27	129	-0.0202	-	7471B	PPB	06 May 2024 15:51:06	U	SMPL
P2386-03MSD - 1	EO-01-050324-BMSD	52217	9.1549	-	7471B	PPB	06 May 2024 15:53:25	U	SMPL
P2386-03LX5 - 1		5801	0.9789	-	7471B	PPB	06 May 2024 15:55:41	U	SMPL
P2386-03A - 1		46438	8.1369	-	7471B	PPB	06 May 2024 15:58:03	U	SMPL
CCV28 - 1	CCV28	28743	5.0200	-	7471B	PPB	06 May 2024 16:00:21	U	SMPL
CCB28 - 1	CCB28	95	-0.0262	-	7471B	PPB	06 May 2024 16:02:42	U	SMPL