

LB130639


7470A

INSTRUMENT IN: CV1

Linear

 μ Abs.:

46575



A= 0.0000e+000

B= 2.1554e-004

C= 7.5095e-003

Rho= 0.9998415

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.0	0.000	0.057	0.057	228	0.000	228	0				0/0
0.2	0.200	0.243	0.043	1092	0.0 %	1092					22
2.5	2.500	2.370	-0.130	10961	0.0 %	10961					5
5.0	5.000	5.021	0.021	23262	0.0 %	23262					0
7.5	7.500	7.463	-0.037	34590	0.0 %	34590					0
10.0	10.000	10.046	0.046	46575	0.0 %	46575					0

LB130639 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 May 2024 16:34:45

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	228	-	0.0000	7470A	PPB	06 May 2024 16:38:26	S	Std
0.2 - 1	50.2	1092	-	0.2000	7470A	PPB	06 May 2024 16:40:43	S	Std
2.5 - 1	52.5	10961	-	2.5000	7470A	PPB	06 May 2024 16:43:00	S	Std
5.0 - 1	55	23262	-	5.0000	7470A	PPB	06 May 2024 16:45:17	S	Std
7.5 - 1	57.5	34590	-	7.5000	7470A	PPB	06 May 2024 16:47:34	S	Std
10.0 - 1	50	46575	-	10.0000	7470A	PPB	06 May 2024 16:52:32	S	Std
ICV34 - 1	ICV34	18252	3.9415	-	7470A	PPB	06 May 2024 16:55:24	U	SMPL
ICB34 - 1	ICB34	-212	-0.0382	-	7470A	PPB	06 May 2024 16:57:40	U	SMPL
CCV36 - 1	CCV36	24590	5.3076	-	7470A	PPB	06 May 2024 16:59:57	U	SMPL
CCB36 - 1	CCB36	-258	-0.0481	-	7470A	PPB	06 May 2024 17:02:12	U	SMPL
CRA - 1	CRA	954	0.2131	-	7470A	PPB	06 May 2024 17:04:30	U	SMPL
HighStd - 1	HighStd	46551	10.0410	-	7470A	PPB	06 May 2024 17:06:45	U	SMPL
ChkStd - 1	ChkStd	32576	7.0289	-	7470A	PPB	06 May 2024 17:09:00	U	SMPL
PB160720BL - 1	PBW	-409	-0.0806	-	7470A	PPB	06 May 2024 17:11:19	U	SMPL
PB160720BS - 1	LC5W	21308	4.6002	-	7470A	PPB	06 May 2024 17:13:37	U	SMPL
P2339-01 - 1	D3479	-39	-0.0009	-	7470A	PPB	06 May 2024 17:15:54	U	SMPL
P2339-02 - 1	D3522	167	0.0435	-	7470A	PPB	06 May 2024 17:18:12	U	SMPL
P2354-02 - 1	V359	-228	-0.0416	-	7470A	PPB	06 May 2024 17:22:46	U	SMPL
P2355-02 - 1	VB16173	-44	-0.0020	-	7470A	PPB	06 May 2024 17:25:02	U	SMPL
CCV37 - 1	CCV37	23072	4.9804	-	7470A	PPB	06 May 2024 17:27:18	U	SMPL
CCB37 - 1	CCB37	-262	-0.0490	-	7470A	PPB	06 May 2024 17:29:33	U	SMPL
P2355-02DUP - 1	VB16173DUP	30	0.0140	-	7470A	PPB	06 May 2024 17:31:51	U	SMPL
P2355-02MS - 1	VB16173MS	17242	3.7238	-	7470A	PPB	06 May 2024 17:34:06	U	SMPL
P2355-02MSD - 1	VB16173MSD	17839	3.8525	-	7470A	PPB	06 May 2024 17:36:22	U	SMPL
PB160721BL - 1	PBW	-181	-0.0315	-	7470A	PPB	06 May 2024 17:38:39	U	SMPL
PB160721BS - 1	LC5W	20234	4.3687	-	7470A	PPB	06 May 2024 17:40:57	U	SMPL
P2339-04 - 1	D3513	-243	-0.0449	-	7470A	PPB	06 May 2024 17:43:13	U	SMPL
P2339-04DUP - 1	D3513DUP	-73	-0.0082	-	7470A	PPB	06 May 2024 17:45:32	U	SMPL
P2339-04MS - 1	D3513MS	9970	2.1564	-	7470A	PPB	06 May 2024 17:47:48	U	SMPL
P2339-04MSD - 1	D3513MSD	9611	2.0790	-	7470A	PPB	06 May 2024 17:50:06	U	SMPL
P2339-05 - 1	D3507	-328	-0.0632	-	7470A	PPB	06 May 2024 17:52:23	U	SMPL
CCV38 - 1	CCV38	21845	4.7159	-	7470A	PPB	06 May 2024 17:54:40	U	SMPL
CCB38 - 1	CCB38	-292	-0.0554	-	7470A	PPB	06 May 2024 17:56:56	U	SMPL
P2339-06 - 1	T3438	-2	0.0071	-	7470A	PPB	06 May 2024 17:59:14	U	SMPL
P2339-07 - 1	T3484	90	0.0269	-	7470A	PPB	06 May 2024 18:01:29	U	SMPL
P2339-08 - 1	T3514-15	146	0.0390	-	7470A	PPB	06 May 2024 18:03:45	U	SMPL
P2341-01 - 1	D237	831	0.1866	-	7470A	PPB	06 May 2024 18:06:00	U	SMPL
P2341-02 - 1	D239	30	0.0140	-	7470A	PPB	06 May 2024 18:08:17	U	SMPL
P2347-01 - 1	392	451	0.1047	-	7470A	PPB	06 May 2024 18:10:33	U	SMPL
PB160673TB - 1	PB160673TB	145	0.0388	-	7470A	PPB	06 May 2024 18:12:50	U	SMPL
PB160674TB - 1	PB160674TB	141	0.0379	-	7470A	PPB	06 May 2024 18:15:06	U	SMPL
P2341-05DLX10 - 1	T240-241	1121	0.2491	-	7470A	PPB	06 May 2024 18:17:24	U	SMPL
CCV39 - 1	CCV39	22314	4.8170	-	7470A	PPB	06 May 2024 18:19:40	U	SMPL
CCB39 - 1	CCB39	-109	-0.0160	-	7470A	PPB	06 May 2024 18:21:55	U	SMPL
P2355-02LX5 - 1		148	0.0394	-	7470A	PPB	06 May 2024 18:24:14	U	SMPL
P2355-02A - 1		18164	3.9225	-	7470A	PPB	06 May 2024 18:26:29	U	SMPL
P2339-04LX5 - 1		-23	0.0026	-	7470A	PPB	06 May 2024 18:28:45	U	SMPL
P2339-04A - 1		12131	2.6222	-	7470A	PPB	06 May 2024 18:31:02	U	SMPL
CCV40 - 1	CCV40	22587	4.8759	-	7470A	PPB	06 May 2024 18:33:18	U	SMPL
CCB40 - 1	CCB40	-227	-0.0414	-	7470A	PPB	06 May 2024 18:35:35	U	SMPL