

LB730644


SFAM01.1

INSTRUMENT 191 CY1

Linear

 μ Abs.:

49969



A= 0.0000e+000

B= 2.0253e-004 slope

C= -7.0081e-002 intercept

Rho= 0.9998315

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.060	0.060	643	0.000	643					11
0.05	0.050					442	0				14
0.20	0.200	0.192	-0.008	1295	0.0 %	1295					1
2.50	2.500	2.523	0.023	12805	0.0 %	12805					1
5.00	5.000	4.857	-0.143	24327	0.0 %	24327					3
7.50	7.500	7.517	0.017	37463	0.0 %	37463					0
10.0	10.000	10.050	0.050	49969	0.0 %	49969					1

INSTREUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	643	-		0 SFAM01.1	PPB	5/6/2024 19:06	Std
	0.2 S01	1295	-		0.2 SFAM01.1	PPB	5/6/2024 19:08	Std
	2.5 S02	12805	-		2.5 SFAM01.1	PPB	5/6/2024 19:10	Std
	5 S03	24327	-		5 SFAM01.1	PPB	5/6/2024 19:13	Std
	7.5 S04	37463	-		7.5 SFAM01.1	PPB	5/6/2024 19:15	Std
	10 S05	49969	-		10 SFAM01.1	PPB	5/6/2024 19:20	Std
ICV035	ICV035	20447	4.0711 -		SFAM01.1	PPB	5/6/2024 19:23	SMPL
ICB035	ICB035	541	0.0395 -		SFAM01.1	PPB	5/6/2024 19:25	SMPL
CCV041	CCV041	24795	4.9517 -		SFAM01.1	PPB	5/6/2024 19:27	SMPL
CCB041	CCB041	495	0.0302 -		SFAM01.1	PPB	5/6/2024 19:29	SMPL
PB160746BL	PBS746	280	-0.0134 -		SFAM01.1	PPB	5/6/2024 19:32	SMPL
P2171-01	MDCOS9	3185	0.575 -		SFAM01.1	PPB	5/6/2024 19:39	SMPL
P2171-02	MDCOS9D	2698	0.4763 -		SFAM01.1	PPB	5/6/2024 19:41	SMPL
P2171-03	MDCOS9S	16678	3.3077 -		SFAM01.1	PPB	5/6/2024 19:43	SMPL
P2171-04	MDCOT3	1864	0.3074 -		SFAM01.1	PPB	5/6/2024 19:45	SMPL
P2171-05	MDCOT4	1323	0.1979 -		SFAM01.1	PPB	5/6/2024 19:48	SMPL
P2171-06	MDCOT5	1320	0.1973 -		SFAM01.1	PPB	5/6/2024 19:50	SMPL
P2171-07	MDCOT6	1166	0.1661 -		SFAM01.1	PPB	5/6/2024 19:52	SMPL
P2171-08	MDCOT7	1603	0.2546 -		SFAM01.1	PPB	5/6/2024 19:54	SMPL
P2171-09	MDCOT8	2232	0.382 -		SFAM01.1	PPB	5/6/2024 19:57	SMPL
P2171-10	MDCOT9	1524	0.2386 -		SFAM01.1	PPB	5/6/2024 19:59	SMPL
P2171-11	MDCOU0	1671	0.2683 -		SFAM01.1	PPB	5/6/2024 20:01	SMPL
P2171-12	MDCOY5	3186	0.5752 -		SFAM01.1	PPB	5/6/2024 20:04	SMPL
P2171-13	MDCOY6	1866	0.3078 -		SFAM01.1	PPB	5/6/2024 20:06	SMPL
P2171-14	MDCOZ1	2600	0.4565 -		SFAM01.1	PPB	5/6/2024 20:08	SMPL
P2171-15	MDCOZ2	1983	0.3315 -		SFAM01.1	PPB	5/6/2024 20:10	SMPL
P2171-16	MDCOR5	2408	0.4176 -		SFAM01.1	PPB	5/6/2024 20:13	SMPL
P2171-17	MDCOR6	4163	0.7731 -		SFAM01.1	PPB	5/6/2024 20:15	SMPL
P2171-18	MDCOR7	2219	0.3793 -		SFAM01.1	PPB	5/6/2024 20:17	SMPL
P2171-19	MDCOR8	2505	0.4373 -		SFAM01.1	PPB	5/6/2024 20:19	SMPL
CCV042	CCV042	23505	4.6904 -		SFAM01.1	PPB	5/6/2024 20:22	SMPL
CCB042	CCB042	358	0.0024 -		SFAM01.1	PPB	5/6/2024 20:24	SMPL
P2171-20	MDCOU3	2341	0.404 -		SFAM01.1	PPB	5/6/2024 20:26	SMPL
P2171-21	MDCOU4	2460	0.4281 -		SFAM01.1	PPB	5/6/2024 20:29	SMPL
P2171-22	MDCOU5	2105	0.3562 -		SFAM01.1	PPB	5/6/2024 20:31	SMPL
PB160747BL	PBS747	22	-0.0656 -		SFAM01.1	PPB	5/6/2024 20:33	SMPL
P2172-01	MDCOT1	2254	0.3864 -		SFAM01.1	PPB	5/6/2024 20:35	SMPL
P2172-02	MDCOT1D	2139	0.3631 -		SFAM01.1	PPB	5/6/2024 20:38	SMPL
P2172-03	MDCOT1S	16371	3.2455 -		SFAM01.1	PPB	5/6/2024 20:40	SMPL
P2172-04	MDCOU6	2413	0.4186 -		SFAM01.1	PPB	5/6/2024 20:42	SMPL
P2172-05	MDCOU7	1614	0.2568 -		SFAM01.1	PPB	5/6/2024 20:44	SMPL
P2172-06	MDCOU8	1347	0.2027 -		SFAM01.1	PPB	5/6/2024 20:47	SMPL
P2172-07	MDCOV1	1652	0.2645 -		SFAM01.1	PPB	5/6/2024 20:49	SMPL
P2172-08	MDCOV2	1773	0.289 -		SFAM01.1	PPB	5/6/2024 20:51	SMPL
P2172-09	MDCOR9	1447	0.223 -		SFAM01.1	PPB	5/6/2024 20:54	SMPL
P2172-10	MDCOS0	2058	0.3467 -		SFAM01.1	PPB	5/6/2024 20:56	SMPL

LB130644

INSTREUMENT ID : CV1

P2172-11	MDCOS3	5487	1.0412 -	SFAM01.1	PPB	5/6/2024 20:58 SMPL
P2172-12	MDCOS4	6885	1.3243 -	SFAM01.1	PPB	5/6/2024 21:00 SMPL
P2172-13	MDCOV5	2234	0.3824 -	SFAM01.1	PPB	5/6/2024 21:03 SMPL
P2172-14	MDCOV6	1681	0.2704 -	SFAM01.1	PPB	5/6/2024 21:05 SMPL
P2172-15	MDCOV7	2003	0.3356 -	SFAM01.1	PPB	5/6/2024 21:07 SMPL
P2172-16	MDCOV8	1779	0.2902 -	SFAM01.1	PPB	5/6/2024 21:09 SMPL
CCV043	CCV043	22814	4.5504 -	SFAM01.1	PPB	5/6/2024 21:12 SMPL
CCB043	CCB043	342	-0.0008 -	SFAM01.1	PPB	5/6/2024 21:14 SMPL
P2172-17	MDCOV9	2436	0.4233 -	SFAM01.1	PPB	5/6/2024 21:16 SMPL
P2172-18	MDCOW0	1932	0.3212 -	SFAM01.1	PPB	5/6/2024 21:19 SMPL
P2172-19	MDCOS5	2092	0.3536 -	SFAM01.1	PPB	5/6/2024 21:21 SMPL
P2172-20	MDCOS6	1562	0.2463 -	SFAM01.1	PPB	5/6/2024 21:23 SMPL
P2172-21	MDCOW1	4907	0.9237 -	SFAM01.1	PPB	5/6/2024 21:25 SMPL
P2172-22	MDCOW2	1981	0.3311 -	SFAM01.1	PPB	5/6/2024 21:28 SMPL
CCV044	CCV044	23713	4.7325 -	SFAM01.1	PPB	5/6/2024 21:32 SMPL
CCB044	CCB044	352	0.0012 -	SFAM01.1	PPB	5/6/2024 21:35 SMPL