

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130649

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/07/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.01	05/07/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	05/07/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.00	05/07/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	05/07/2024	08:20
6	P2375-02	1	Solid	20.02	20	22.4	9.21	05/07/2024	08:30
7	P2375-02DUP	1	Solid	20.03	20	22.5	9.22	05/07/2024	08:31
8	P2398-04	1	Solid	20.02	20	22.6	5.49	05/07/2024	08:40
9	CCV2	1	Water	NA	NA	20.3	12.02	05/07/2024	08:45

WORKLIST(Hardcopy Internal Chain)

12130649

WorkList Name : corrosion p2375 WorkList ID : 180031 Department : Wet-Chemistry Date : 05-07-2024 07:52:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2375-02	WC-A-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	K22	05/03/2024	9045D
P2398-04	WC-KK	Solid	Corrosivity	Cool 4 deg C	PSEG03	K31	05/06/2024	9045D

Date/Time: 05/07/24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time: 05/07/24 11:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]