

LB130650


SFAM01.1

INSTRUMENT ID: EVI

Linear

 μ Abs.:

45929



A= 0.0000e+000

B= 2.1758e-004 slope

C= -5.7407e-002 $\rightarrow$ intercept

Rho= 0.9998256

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	obj
0.00	0.000	0.048	0.048	484	0.000	484					11
0.05	0.050					442	0				13
0.20	0.200	0.195	-0.005	1162	0.0 %	1162					13
2.50	2.500	2.419	-0.081	11382	0.0 %	11382					20
5.00	5.000	4.979	-0.021	23146	0.0 %	23146					1
7.50	7.500	7.623	0.123	35298	0.0 %	35298					
10.0	10.000	9.936	-0.064	45929	0.0 %	45929					

LB130650

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	484	-		0 SFAM01.1	PPB	5/7/2024 8:58	Std
	0.2 S01	1162	-		0.2 SFAM01.1	PPB	5/7/2024 9:00	Std
	2.5 S02	11382	-		2.5 SFAM01.1	PPB	5/7/2024 9:02	Std
	5 S03	23146	-		5 SFAM01.1	PPB	5/7/2024 9:04	Std
	7.5 S04	35298	-		7.5 SFAM01.1	PPB	5/7/2024 9:07	Std
	10 S05	45929	-		10 SFAM01.1	PPB	5/7/2024 9:15	Std
ICV036	ICV036	19697	4.2283 -		SFAM01.1	PPB	5/7/2024 9:20	SMPL
ICB036	ICB036	593	0.0716 -		SFAM01.1	PPB	5/7/2024 9:28	SMPL
CCV045	CCV045	23793	5.1195 -		SFAM01.1	PPB	5/7/2024 9:31	SMPL
CCB045	CCB045	343	0.0172 -		SFAM01.1	PPB	5/7/2024 9:33	SMPL
PB160750BL	PBS750	215	-0.0106 -		SFAM01.1	PPB	5/7/2024 9:35	SMPL
P2173-01	MDCOY3	2538	0.4948 -		SFAM01.1	PPB	5/7/2024 9:37	SMPL
P2173-02	MDCOY4	2301	0.4433 -		SFAM01.1	PPB	5/7/2024 9:40	SMPL
P2173-03	MDCOY7	2994	0.594 -		SFAM01.1	PPB	5/7/2024 9:42	SMPL
P2173-04	MDCOY8	2467	0.4794 -		SFAM01.1	PPB	5/7/2024 9:44	SMPL
P2173-05	MDCOY9	2427	0.4707 -		SFAM01.1	PPB	5/7/2024 9:47	SMPL
P2173-06	MDCOZ0	2208	0.423 -		SFAM01.1	PPB	5/7/2024 9:49	SMPL
P2173-07	MDCOZ3	2485	0.4833 -		SFAM01.1	PPB	5/7/2024 9:51	SMPL
P2173-08	MDCOZ4	2373	0.4589 -		SFAM01.1	PPB	5/7/2024 9:53	SMPL
P2173-09	MDCOR4	933	0.1456 -		SFAM01.1	PPB	5/7/2024 9:56	SMPL
P2173-10	MDCOS7	1490	0.2668 -		SFAM01.1	PPB	5/7/2024 9:58	SMPL
P2173-11	MDCOS8	1202	0.2041 -		SFAM01.1	PPB	5/7/2024 10:00	SMPL
P2173-12	MDCOU1	1821	0.3388 -		SFAM01.1	PPB	5/7/2024 10:02	SMPL
P2173-13	MDCOU2	2080	0.3952 -		SFAM01.1	PPB	5/7/2024 10:05	SMPL
P2173-14	MDCOW3	18640	3.9983 -		SFAM01.1	PPB	5/7/2024 10:07	SMPL
P2173-15	MDCOW4	4498	0.9213 -		SFAM01.1	PPB	5/7/2024 10:09	SMPL
P2173-16	MDCOT0	1563	0.2827 -		SFAM01.1	PPB	5/7/2024 10:12	SMPL
P2173-17	MDCOT2	2102	0.4 -		SFAM01.1	PPB	5/7/2024 10:14	SMPL
P2173-18	MDCOT2D	2162	0.413 -		SFAM01.1	PPB	5/7/2024 10:16	SMPL
P2173-19	MDCOT2S	15147	3.2383 -		SFAM01.1	PPB	5/7/2024 10:18	SMPL
CCV046	CCV046	24255	5.2201 -		SFAM01.1	PPB	5/7/2024 10:21	SMPL
CCB046	CCB046	80	-0.04 -		SFAM01.1	PPB	5/7/2024 10:27	SMPL
P2173-20	MDCOV4	3644	0.7355 -		SFAM01.1	PPB	5/7/2024 10:29	SMPL
P2173-21	MDCOX1	5666	1.1754 -		SFAM01.1	PPB	5/7/2024 10:31	SMPL
P2173-22	MDCOX2	5380	1.1132 -		SFAM01.1	PPB	5/7/2024 10:33	SMPL
PB160751BL	PBS751	91	-0.0376 -		SFAM01.1	PPB	5/7/2024 10:36	SMPL
P2174-01	MDCOW9	11767	2.5029 -		SFAM01.1	PPB	5/7/2024 10:38	SMPL
P2174-02	MDCOX0	5873	1.2205 -		SFAM01.1	PPB	5/7/2024 10:40	SMPL
P2174-03	MDCOX5	2931	0.5803 -		SFAM01.1	PPB	5/7/2024 10:42	SMPL
P2174-04	MDCOX6	2140	0.4082 -		SFAM01.1	PPB	5/7/2024 10:45	SMPL
P2174-05	MDCOX7	3215	0.6421 -		SFAM01.1	PPB	5/7/2024 10:47	SMPL
P2174-06	MDCOX8	3032	0.6023 -		SFAM01.1	PPB	5/7/2024 10:49	SMPL
P2174-07	MDCPA3	2859	0.5647 -		SFAM01.1	PPB	5/7/2024 10:52	SMPL
P2174-08	MDCPA4	5056	1.0427 -		SFAM01.1	PPB	5/7/2024 10:54	SMPL
P2174-09	MDCOX4	2214	0.4243 -		SFAM01.1	PPB	5/7/2024 10:56	SMPL
P2174-10	MDCOX9	2673	0.5242 -		SFAM01.1	PPB	5/7/2024 10:58	SMPL

LB130650

INSTRUMENT ID : CV1

P2174-11	MDCOY0	1259	0.2165 -	SFAM01.1	PPB	5/7/2024 11:01 SMPL
P2174-12	MDCOY1	1350	0.2363 -	SFAM01.1	PPB	5/7/2024 11:03 SMPL
P2174-13	MDCOY2	1310	0.2276 -	SFAM01.1	PPB	5/7/2024 11:05 SMPL
P2174-14	MDCOZ9	3022	0.6001 -	SFAM01.1	PPB	5/7/2024 11:07 SMPL
P2174-15	MDCPA9	1560	0.282 -	SFAM01.1	PPB	5/7/2024 11:10 SMPL
P2174-16	MDCOW5	2340	0.4517 -	SFAM01.1	PPB	5/7/2024 11:12 SMPL
CCV047	CCV047	23526	5.0614 -	SFAM01.1	PPB	5/7/2024 11:14 SMPL
CCB047	CCB047	-105	-0.0803 -	SFAM01.1	PPB	5/7/2024 11:16 SMPL
P2174-17	MDCOW6	2644	0.5179 -	SFAM01.1	PPB	5/7/2024 11:19 SMPL
P2174-18	MDCOW8	3014	0.5984 -	SFAM01.1	PPB	5/7/2024 11:21 SMPL
P2174-19	MDCOZ7	1928	0.3621 -	SFAM01.1	PPB	5/7/2024 11:30 SMPL
P2174-20	MDCOZ7D	1971	0.3714 -	SFAM01.1	PPB	5/7/2024 11:32 SMPL
P2174-21	MDCOZ7S	14833	3.17 -	SFAM01.1	PPB	5/7/2024 11:34 SMPL
P2174-22	MDCOZ8	3870	0.7846 -	SFAM01.1	PPB	5/7/2024 11:36 SMPL
CCV048	CCV048	23783	5.1174 -	SFAM01.1	PPB	5/7/2024 11:39 SMPL
CCB048	CCB048	198	-0.0143 -	SFAM01.1	PPB	5/7/2024 11:41 SMPL