


L3130654

SFAM01.1

INSTRUMENT IN CV1



Linear

A= 0.0000e+000

B= 2.0091e-004 slope

C= -1.2414e-001 y-intercept

Rho= 0.9999031

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.00	0.000	0.003	0.003	632	0.000	632					1
0.05	0.050					442					1
0.20	0.200	0.202	0.002	1625	0.0 %	1625	0				2
2.50	2.500	2.443	-0.057	12778	0.0 %	12778					1
5.00	5.000	5.038	0.038	25692	0.0 %	25692					1
7.50	7.500	7.581	0.081	38354	0.0 %	38354					1
10.0	10.000	9.933	-0.067	50057	0.0 %	50057					1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	632	-		0 SFAM01.1	PPB	5/7/2024 15:47	Std
	0.2 S01	1625	-		0.2 SFAM01.1	PPB	5/7/2024 15:49	Std
	2.5 S02	12778	-		2.5 SFAM01.1	PPB	5/7/2024 15:51	Std
	5 S03	25692	-		5 SFAM01.1	PPB	5/7/2024 15:53	Std
	7.5 S04	38354	-		7.5 SFAM01.1	PPB	5/7/2024 15:56	Std
	10 S05	50057	-		10 SFAM01.1	PPB	5/7/2024 16:01	Std
ICV037	ICV037	19957	3.8854 -		SFAM01.1	PPB	5/7/2024 16:04	SMPL
ICB037	ICB037	458	-0.0321 -		SFAM01.1	PPB	5/7/2024 16:06	SMPL
CCV049	CCV049	25669	5.033 -		SFAM01.1	PPB	5/7/2024 16:11	SMPL
CCB049	CCB049	429	-0.0379 -		SFAM01.1	PPB	5/7/2024 16:13	SMPL
PB160763BL	PBS763	801	0.0368 -		SFAM01.1	PPB	5/7/2024 16:15	SMPL
P2175-01	MDCPA0	2747	0.4278 -		SFAM01.1	PPB	5/7/2024 16:18	SMPL
P2175-02	MDCPA1	4026	0.6847 -		SFAM01.1	PPB	5/7/2024 16:20	SMPL
P2175-03	MDCPA2	2586	0.3954 -		SFAM01.1	PPB	5/7/2024 16:22	SMPL
P2175-04	MDCPA5	2884	0.4553 -		SFAM01.1	PPB	5/7/2024 16:24	SMPL
P2175-05	MDCPA6	2941	0.4667 -		SFAM01.1	PPB	5/7/2024 16:27	SMPL
P2175-06	MDCPA7	2350	0.348 -		SFAM01.1	PPB	5/7/2024 16:29	SMPL
P2175-07	MDCPA8	2378	0.3536 -		SFAM01.1	PPB	5/7/2024 16:31	SMPL
P2175-10	MDCPG8	2083	0.2944 -		SFAM01.1	PPB	5/7/2024 16:34	SMPL
P2175-11	MDCPG9	2534	0.385 -		SFAM01.1	PPB	5/7/2024 16:36	SMPL
P2175-12	MDCPH0	2538	0.3858 -		SFAM01.1	PPB	5/7/2024 16:38	SMPL
P2175-13	MDCPH0D	2652	0.4087 -		SFAM01.1	PPB	5/7/2024 16:40	SMPL
P2175-14	MDCPH0S	15663	3.0227 -		SFAM01.1	PPB	5/7/2024 16:43	SMPL
P2175-15	MDCPH1	2296	0.3371 -		SFAM01.1	PPB	5/7/2024 16:45	SMPL
P2175-16	MDCPH2	2519	0.3819 -		SFAM01.1	PPB	5/7/2024 16:47	SMPL
P2175-17	MDCPH3	2248	0.3275 -		SFAM01.1	PPB	5/7/2024 16:49	SMPL
P2175-18	MDCPH4	2862	0.4509 -		SFAM01.1	PPB	5/7/2024 16:52	SMPL
P2175-19	MDCPH5	2730	0.4243 -		SFAM01.1	PPB	5/7/2024 16:54	SMPL
P2175-20	MDCPH6	3196	0.518 -		SFAM01.1	PPB	5/7/2024 16:56	SMPL
P2175-21	MDCPH7	4747	0.8296 -		SFAM01.1	PPB	5/7/2024 16:59	SMPL
CCV050	CCV050	25635	5.0261 -		SFAM01.1	PPB	5/7/2024 17:01	SMPL
CCB050	CCB050	326	-0.0586 -		SFAM01.1	PPB	5/7/2024 17:03	SMPL
P2175-22	MDCPH8	2638	0.4059 -		SFAM01.1	PPB	5/7/2024 17:05	SMPL
PB160764BL	PBS764	600	-0.0036 -		SFAM01.1	PPB	5/7/2024 17:08	SMPL
P2179-01	DCOR0	334497	HIGH -		SFAM01.1	PPB	5/7/2024 17:10	
P2179-02	MDCPH9	2634	0.4051 -		SFAM01.1	PPB	5/7/2024 17:12	SMPL
P2179-03	MDCPJ0	2951	0.4687 -		SFAM01.1	PPB	5/7/2024 17:14	SMPL
P2179-04	MDCPJ1	2700	0.4183 -		SFAM01.1	PPB	5/7/2024 17:16	SMPL
P2179-05	MDCPB0	3485	0.576 -		SFAM01.1	PPB	5/7/2024 17:19	SMPL
P2179-06	MDCPB1	3309	0.5407 -		SFAM01.1	PPB	5/7/2024 17:21	SMPL
P2179-07	MDCPB2	2297	0.3373 -		SFAM01.1	PPB	5/7/2024 17:23	SMPL
P2179-08	MDCPB3	2248	0.3275 -		SFAM01.1	PPB	5/7/2024 17:25	SMPL
P2179-09	MDCPB4	1909	0.2594 -		SFAM01.1	PPB	5/7/2024 17:28	SMPL
P2179-10	MDCPB5	1793	0.2361 -		SFAM01.1	PPB	5/7/2024 17:30	SMPL
P2179-11	MDCPB6	2733	0.4249 -		SFAM01.1	PPB	5/7/2024 17:32	SMPL
P2179-12	MDCPB7	2303	0.3386 -		SFAM01.1	PPB	5/7/2024 17:35	SMPL

LB130654

INSTRUMENT ID : CV1

P2179-13	MDCPB8	2176	0.313 -	SFAM01.1	PPB	5/7/2024 17:37 SMPL
P2179-14	MDCPB9	2520	0.3821 -	SFAM01.1	PPB	5/7/2024 17:39 SMPL
P2179-15	MDCPG0	3400	0.5589 -	SFAM01.1	PPB	5/7/2024 17:41 SMPL
P2179-16	MDCPG1	2463	0.3707 -	SFAM01.1	PPB	5/7/2024 17:44 SMPL
P2179-17	MDCPG1D	2192	0.3163 -	SFAM01.1	PPB	5/7/2024 17:46 SMPL
P2179-18	MDCPG1S	14017	2.692 -	SFAM01.1	PPB	5/7/2024 17:48 SMPL
CCV051	CCV051	24591	4.8164 -	SFAM01.1	PPB	5/7/2024 17:50 SMPL
CCB051	CCB051	424	-0.039 -	SFAM01.1	PPB	5/7/2024 17:53 SMPL
P2179-19	MDCPG2	2861	0.4507 -	SFAM01.1	PPB	5/7/2024 17:55 SMPL
P2179-20	MDCPG3	2263	0.3305 -	SFAM01.1	PPB	5/7/2024 17:57 SMPL
P2179-21	MDCPG4	1835	0.2445 -	SFAM01.1	PPB	5/7/2024 18:00 SMPL
P2179-22	MDCPG5	2114	0.3006 -	SFAM01.1	PPB	5/7/2024 18:02 SMPL
P2179-23	MDCPG6	2216	0.3211 -	SFAM01.1	PPB	5/7/2024 18:04 SMPL
PB160765BL	PBS765	479	-0.0279 -	SFAM01.1	PPB	5/7/2024 18:06 SMPL
P2180-01	MDCPG7	2172	0.3122 -	SFAM01.1	PPB	5/7/2024 18:09 SMPL
P2180-02	MDCPG7D	2255	0.3289 -	SFAM01.1	PPB	5/7/2024 18:11 SMPL
P2180-03	MDCPG7S	15375	2.9648 -	SFAM01.1	PPB	5/7/2024 18:13 SMPL
CCV052	CCV052	24070	4.7117 -	SFAM01.1	PPB	5/7/2024 18:21 SMPL
CCB052	CCB052	384	-0.047 -	SFAM01.1	PPB	5/7/2024 18:23 SMPL