

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130656

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/07/2024	16:06
2	CAL2	1	Water	NA	NA	20.3	7.00	05/07/2024	16:08
3	CAL3	1	Water	NA	NA	20.3	10.02	05/07/2024	16:10
4	ICV	1	Water	NA	NA	20.3	7.01	05/07/2024	16:15
5	CCV1	1	Water	NA	NA	20.3	2.02	05/07/2024	16:20
6	P2410-01	1	Water	NA	NA	22.1	7.30	05/07/2024	16:25
7	P2410-01DUP	1	Water	NA	NA	22.2	7.31	05/07/2024	16:26
8	CCV2	1	Water	NA	NA	20.3	12.02	05/07/2024	16:30

WORKLIST(Hardcopy Internal Chain)

UB130656

WorkList Name : PH P2310

WorkList ID : 180053

Department : Wet-Chemistry

Date : 05-07-2024 15:40:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2410-01	Box A2295	Water	pH	Cool 4 deg C	PSEG03	K41	05/07/2024	9040C

Date/Time 05/07/24 15:50
Raw Sample Received by: JH LWC
Raw Sample Relinquished by: CR SM

Date/Time 05/07/24 17:20
Raw Sample Received by: CR SM
Raw Sample Relinquished by: JH LWC