

LB130664


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

52907



A= 0.0000e+000

B= 1.8949e-004 *skip*C= -1.0104e-001 *-1.0104e-001*

Rho= 0.9999001

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.025	-0.025	401	0.000	401				
0.05	0.050					442				
0.20	0.200	0.210	0.010	1642	0.0 %	1642	0			
2.50	2.500	2.483	-0.017	13638	0.0 %	13638				
5.00	5.000	5.010	0.010	26973	0.0 %	26973				
7.50	7.500	7.597	0.097	40627	0.0 %	40627				
10.0	10.000	9.924	-0.076	52907	0.0 %	52907				

1-00911 %D

LB130664

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S00	401	-		0 SFAM01.1	PPB	5/8/2024 10:30	Std
	0.2 S01	1642	-		0.2 SFAM01.1	PPB	5/8/2024 10:33	Std
	2.5 S02	13638	-		2.5 SFAM01.1	PPB	5/8/2024 10:36	Std
	5 S03	26973	-		5 SFAM01.1	PPB	5/8/2024 10:38	Std
	7.5 S04	40627	-		7.5 SFAM01.1	PPB	5/8/2024 10:40	Std
	10 S05	52907	-		10 SFAM01.1	PPB	5/8/2024 10:42	Std
ICV038	ICV038	19901	3.67 -		SFAM01.1	PPB	5/8/2024 10:48	SMPL
ICB038	ICB038	539	0.0011 -		SFAM01.1	PPB	5/8/2024 10:50	SMPL
CCV083	CCV083	27172	5.0478 -		SFAM01.1	PPB	5/8/2024 10:53	SMPL
CCB053	CCB053	290	-0.0461 -		SFAM01.1	PPB	5/8/2024 10:55	SMPL
P2179-01DLX25	DCOR0	24004	4.4475 -		SFAM01.1	PPB	5/8/2024 11:04	SMPL
CCV054	CCV054	24554	4.5517 -		SFAM01.1	PPB	5/8/2024 11:06	SMPL
CCB054	CCB054	401	-0.0251 -		SFAM01.1	PPB	5/8/2024 11:09	SMPL